

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1822

Client: ENTACT

Analytical Method: 8270E DataFile: BP024340.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	RPD		Low	Limits	
							Qual	Qual		High	RPD
PB167625BSD	Phenol	50	45.8	ug/L	92	1			20 (66)	160 (118)	20 (20)
	1,4-Dichlorobenzene	50	44.8	ug/L	90	3			70 (76)	130 (103)	20 (20)
	1,2,4-Trichlorobenzene	50	45.0	ug/L	90	2			70 (41)	130 (115)	20 (20)
	Naphthalene	50	42.6	ug/L	85	2			70 (64)	130 (107)	20 (20)