

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: ENTA05

Lab Code: CHEM Case No.: Q1822 SAS No.: Q1822 SDG No.: Q1822

Instrument ID: MSVOA_X Calibration Date/Time: 04/16/2025 09:28

Lab File ID: VX045797.D Init. Calib. Date(s): 04/01/2025 04/01/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 17:06 19:02

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Methyl tert-butyl Ether	2.100	2.349		11.86	20
Carbon Tetrachloride	0.518	0.503		-2.9	20
Chloroform	1.306	1.328		1.68	20
1,1,1-Trichloroethane	1.105	1.092		-1.18	20
Benzene	1.463	1.414		-3.35	20
Toluene	0.885	0.862		-2.6	20
Tetrachloroethene	0.353	0.315		-10.77	20
Ethyl Benzene	1.914	1.888		-1.36	20
m/p-Xylenes	0.696	0.691		-0.72	20
o-Xylene	0.686	0.694		1.17	20
1,2-Dichloroethane-d4	0.914	0.994		8.75	20
Dibromofluoromethane	0.355	0.371		4.51	20
Toluene-d8	1.238	1.235		-0.24	20
4-Bromofluorobenzene	0.451	0.505		11.97	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.