

Report of Analysis

Client:	ENTACT		Date Collected:	
Project:	540 Degraw St, Brooklyn, NY - E9309		Date Received:	
Client Sample ID:	VX0416WBS01		SDG No.:	Q1822
Lab Sample ID:	VX0416WBS01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group4
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045800.D	1		04/16/25 12:11	VX041625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
1634-04-4	Methyl tert-butyl Ether	20.4		0.16	1.00	ug/L
56-23-5	Carbon Tetrachloride	19.9		0.25	1.00	ug/L
67-66-3	Chloroform	19.7		0.25	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	20.1		0.20	1.00	ug/L
71-43-2	Benzene	18.9		0.15	1.00	ug/L
108-88-3	Toluene	19.2		0.14	1.00	ug/L
127-18-4	Tetrachloroethene	19.2		0.23	1.00	ug/L
100-41-4	Ethyl Benzene	19.6		0.13	1.00	ug/L
1330-20-7	Total Xylenes	58.9		0.36	3.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	53.5		70 (74) - 130 (125)	107%	SPK: 50
1868-53-7	Dibromofluoromethane	52.4		70 (75) - 130 (124)	105%	SPK: 50
2037-26-5	Toluene-d8	51.6		70 (86) - 130 (113)	103%	SPK: 50
460-00-4	4-Bromofluorobenzene	54.2		70 (77) - 130 (121)	108%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	91300	5.544			
540-36-3	1,4-Difluorobenzene	163000	6.757			
3114-55-4	Chlorobenzene-d5	145000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	66600	12.018			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products