

Cover Page

Order ID : Q1822

Project ID : 540 Degraw St, Brooklyn, NY - E9309

Client : ENTACT

Lab Sample Number

Q1822-01

Client Sample Number

TW-WTS-06

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 4/22/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following "Results Qualifiers" are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. "10 U". This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as "12 B".
E	Indicates the analyte 's concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a "P".
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q1822

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Castody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: KETAN PATEL

Date: 04/22/2025



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

LAB CHRONICLE

OrderID:	Q1822	OrderDate:	4/16/2025 1:17:00 PM
Client:	ENTACT	Project:	540 Degraw St, Brooklyn, NY - E9309
Contact:	Jarod Stanfield	Location:	L21,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1822-01	TW-WTS-06	Water	VOCMS Group4	8260-Low	04/15/25		04/16/25	04/16/25



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q1822

Client: ENTACT

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
-----------	-----------	--------	-----------	---------------	---	-----	-----	-------

Client ID:

0

Total Voc :

Total Concentration:



QC SUMMARY

Surrogate Summary

SDG No.: Q1822

Client: ENTACT

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1822-01	TW-WTS-06	1,2-Dichloroethane-d4	50	54.3	109	70 (74)	130 (125)
		Dibromofluoromethane	50	50.3	101	70 (75)	130 (124)
		Toluene-d8	50	50.3	101	70 (86)	130 (113)
		4-Bromofluorobenzene	50	50.4	101	70 (77)	130 (121)
VX0416WBL01	VX0416WBL01	1,2-Dichloroethane-d4	50	53.0	106	70 (74)	130 (125)
		Dibromofluoromethane	50	50.5	101	70 (75)	130 (124)
		Toluene-d8	50	49.6	99	70 (86)	130 (113)
		4-Bromofluorobenzene	50	51.9	104	70 (77)	130 (121)
VX0416WBS01	VX0416WBS01	1,2-Dichloroethane-d4	50	53.5	107	70 (74)	130 (125)
		Dibromofluoromethane	50	52.4	105	70 (75)	130 (124)
		Toluene-d8	50	51.6	103	70 (86)	130 (113)
		4-Bromofluorobenzene	50	54.3	108	70 (77)	130 (121)
VX0416WBSD0	VX0416WBSD01	1,2-Dichloroethane-d4	50	54.1	108	70 (74)	130 (125)
		Dibromofluoromethane	50	53.5	107	70 (75)	130 (124)
		Toluene-d8	50	51.1	102	70 (86)	130 (113)
		4-Bromofluorobenzene	50	54.4	109	70 (77)	130 (121)

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1822

Client: ENTACT

Analytical Method: SW8260-Low

Datafile : VX045800.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0416WBS01	Methyl tert-butyl Ether	20	20.4	ug/L	102			70 (78)	130 (114)	
	Carbon Tetrachloride	20	19.9	ug/L	100			70 (77)	130 (113)	
	Chloroform	20	19.7	ug/L	99			70 (79)	130 (113)	
	1,1,1-Trichloroethane	20	20.1	ug/L	101			70 (80)	130 (108)	
	Benzene	20	18.9	ug/L	95			70 (82)	130 (109)	
	Toluene	20	19.2	ug/L	96			70 (82)	130 (110)	
	Tetrachloroethene	20	19.2	ug/L	96			70 (67)	130 (123)	
	Ethyl Benzene	20	19.6	ug/L	98			70 (83)	130 (109)	
	m/p-Xylenes	40	39.0	ug/L	98			70 (82)	130 (110)	
	o-Xylene	20	19.9	ug/L	100			70 (83)	130 (109)	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1822
Client: ENTACT
Analytical Method: SW8260-Low **Datafile :** VX045801.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0416WBSD01	Methyl tert-butyl Ether	20	21.0	ug/L	105	3		70 (78)	130 (114)	20 (20)
	Carbon Tetrachloride	20	19.8	ug/L	99	1		70 (77)	130 (113)	20 (20)
	Chloroform	20	20.0	ug/L	100	1		70 (79)	130 (113)	20 (20)
	1,1,1-Trichloroethane	20	19.9	ug/L	100	1		70 (80)	130 (108)	20 (20)
	Benzene	20	19.1	ug/L	96	1		70 (82)	130 (109)	20 (20)
	Toluene	20	19.3	ug/L	97	1		70 (82)	130 (110)	20 (20)
	Tetrachloroethene	20	19.4	ug/L	97	1		70 (67)	130 (123)	20 (20)
	Ethyl Benzene	20	19.8	ug/L	99	1		70 (83)	130 (109)	20 (20)
	m/p-Xylenes	40	40.0	ug/L	100	2		70 (82)	130 (110)	20 (20)
	o-Xylene	20	19.9	ug/L	100	0		70 (83)	130 (109)	20 (20)

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VX0416WBL01

Lab Name: CHEMTECH

Contract: ENTA05

Lab Code: CHEM Case No.: Q1822

SAS No.: Q1822 SDG NO.: Q1822

Lab File ID: VX045799.D

Lab Sample ID: VX0416WBL01

Date Analyzed: 04/16/2025

Time Analyzed: 11:47

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA_X

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VX0416WBS01	VX0416WBS01	VX045800.D	04/16/2025
VX0416WBSD01	VX0416WBSD01	VX045801.D	04/16/2025
TW-WTS-06	Q1822-01	VX045814.D	04/16/2025

COMMENTS: _____



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: ENTA05
Lab Code: CHEM Case No.: Q1822 SAS No.: Q1822 SDG NO.: Q1822
Lab File ID: VX045524.D BFB Injection Date: 04/01/2025
Instrument ID: MSVOA_X BFB Injection Time: 16:15
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	23.6
75	30.0 - 60.0% of mass 95	58.2
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.3
173	Less than 2.0% of mass 174	0.8 (1.2) 1
174	50.0 - 100.0% of mass 95	66.8
175	5.0 - 9.0% of mass 174	5.2 (7.8) 1
176	95.0 - 101.0% of mass 174	65.3 (97.8) 1
177	5.0 - 9.0% of mass 176	4.4 (6.8) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC001	VSTDIC001	VX045525.D	04/01/2025	17:06
VSTDIC005	VSTDIC005	VX045526.D	04/01/2025	17:29
VSTDIC020	VSTDIC020	VX045527.D	04/01/2025	17:52
VSTDIC050	VSTDIC050	VX045528.D	04/01/2025	18:15
VSTDIC100	VSTDIC100	VX045529.D	04/01/2025	18:38
VSTDIC150	VSTDIC150	VX045530.D	04/01/2025	19:02



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: ENTA05
Lab Code: CHEM Case No.: Q1822 SAS No.: Q1822 SDG NO.: Q1822
Lab File ID: VX045796.D BFB Injection Date: 04/16/2025
Instrument ID: MSVOA_X BFB Injection Time: 08:26
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	22.6
75	30.0 - 60.0% of mass 95	56.1
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.3
173	Less than 2.0% of mass 174	0.9 (1.3) 1
174	50.0 - 100.0% of mass 95	67
175	5.0 - 9.0% of mass 174	5 (7.5) 1
176	95.0 - 101.0% of mass 174	64.6 (96.5) 1
177	5.0 - 9.0% of mass 176	4.3 (6.6) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX045797.D	04/16/2025	09:28
VX0416WBL01	VX0416WBL01	VX045799.D	04/16/2025	11:47
VX0416WBS01	VX0416WBS01	VX045800.D	04/16/2025	12:11
VX0416WBSD01	VX0416WBSD01	VX045801.D	04/16/2025	12:39
TW-WTS-06	Q1822-01	VX045814.D	04/16/2025	17:41

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ENTA05
 Lab Code: CHEM Case No.: Q1822 SAS No.: Q1822 SDG NO.: Q1822
 Lab File ID: VX045797.D Date Analyzed: 04/16/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:28
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	85794	5.54	152447	6.75	137129	10.05
UPPER LIMIT	171588	6.038	304894	7.251	274258	10.549
LOWER LIMIT	42897	5.038	76223.5	6.251	68564.5	9.549
EPA SAMPLE NO.						
TW-WTS-06	71191	5.55	141409	6.76	129813	10.05
VX0416WBL01	72471	5.54	141858	6.76	129459	10.05
VX0416WBS01	91256	5.54	163123	6.76	144629	10.05
VX0416WBSD01	87856	5.54	155557	6.76	137506	10.05

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ENTA05
 Lab Code: CHEM Case No.: Q1822 SAS No.: Q1822 SDG NO.: Q1822
 Lab File ID: VX045797.D Date Analyzed: 04/16/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:28
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	66178	12.018				
UPPER LIMIT	132356	12.518				
LOWER LIMIT	33089	11.518				
EPA SAMPLE NO.						
TW-WTS-06	52878	12.02				
VX0416WBL01	56266	12.02				
VX0416WBS01	66591	12.02				
VX0416WBSD01	65216	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.



SAMPLE DATA

Report of Analysis

Client:	ENTACT	Date Collected:	04/15/25
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	04/16/25
Client Sample ID:	TW-WTS-06	SDG No.:	Q1822
Lab Sample ID:	Q1822-01	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group4
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045814.D	1		04/16/25 17:41	VX041625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	1.00	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	1.00	ug/L
67-66-3	Chloroform	0.25	U	0.25	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.20	U	0.20	1.00	ug/L
71-43-2	Benzene	0.15	U	0.15	1.00	ug/L
108-88-3	Toluene	0.14	U	0.14	1.00	ug/L
127-18-4	Tetrachloroethene	0.23	U	0.23	1.00	ug/L
100-41-4	Ethyl Benzene	0.13	U	0.13	1.00	ug/L
1330-20-7	Total Xylenes	0.36	U	0.36	3.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	54.3		70 (74) - 130 (125)	109%	SPK: 50
1868-53-7	Dibromofluoromethane	50.3		70 (75) - 130 (124)	101%	SPK: 50
2037-26-5	Toluene-d8	50.3		70 (86) - 130 (113)	101%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.4		70 (77) - 130 (121)	101%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	71200	5.55			
540-36-3	1,4-Difluorobenzene	141000	6.757			
3114-55-4	Chlorobenzene-d5	130000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	52900	12.018			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041625\
 Data File : VX045814.D
 Acq On : 16 Apr 2025 17:41
 Operator : JC/MD
 Sample : Q1822-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TW-WTS-06

Quant Time: Apr 17 01:38:36 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	5.550	168	71191	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	141409	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	129813	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	52878	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	70707	54.311	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 108.620%	
35) Dibromofluoromethane	5.379	113	50497	50.331	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.660%	
50) Toluene-d8	8.647	98	176295	50.343	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.680%	
62) 4-Bromofluorobenzene	11.079	95	64264	50.381	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 100.760%	
Target Compounds						
16) Acetone	2.380	43	7434	13.906	ug/l	Qvalue 97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041625\
Data File : VX045814.D
Acq On : 16 Apr 2025 17:41
Operator : JC/MD
Sample : Q1822-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TW-WTS-06

Quant Time: Apr 17 01:38:36 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX045814.D
Acq: 16 Apr 2025 17:41




Instrument :
MSVOA_X
ClientSampleId :
TW-WTS-06


Tgt Ion:168 Resp: 71191
Ion Ratio Lower Upper
168 100
99 71.8 52.3 78.5



#16
Acetone
Concen: 13.906 ug/l
RT: 2.380 min Scan# 212
Delta R.T. -0.006 min
Lab File: VX045814.D
Acq: 16 Apr 2025 17:41

Tgt Ion: 43 Resp: 7434
Ion Ratio Lower Upper
43 100
58 26.5 22.3 33.5





#33

1,2-Dichloroethane-d4

Concen: 54.311 ug/l

RT: 5.952 min Scan# 798

Delta R.T. 0.000 min

Lab File: VX045814.D

Acq: 16 Apr 2025 17:41

Instrument :

MSVOA_X

ClientSampleId :





TW-WTS-06

Tgt Ion: 65 Resp: 70707

Ion Ratio Lower Upper

65 100

67 48.5 0.0 99.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. 0.000 min

Lab File: VX045814.D

Acq: 16 Apr 2025 17:41




Tgt Ion: 114 Resp: 141409


Ion Ratio Lower Upper

114 100

63 24.3 0.0 46.8

88 17.5 0.0 35.4





#35

Dibromofluoromethane

Concen: 50.331 ug/l

RT: 5.379 min Scan# 704

Delta R.T. 0.000 min

Lab File: VX045814.D


Acq: 16 Apr 2025 17:41

Instrument :

MSVOA_X

ClientSampleId :

TW-WTS-06






Tgt Ion:113 Resp: 50497

Ion Ratio Lower Upper

113 100

111 102.9 81.8 122.6

192 16.4 13.8 20.6



#50

Toluene-d8

Concen: 50.343 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. 0.000 min

Lab File: VX045814.D




Acq: 16 Apr 2025 17:41


Tgt Ion: 98 Resp: 176295

Ion Ratio Lower Upper

98 100

100 64.7 52.2 78.4









#62
4-Bromofluorobenzene
Concen: 50.381 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX045814.D
Acq: 16 Apr 2025 17:41




Instrument :
MSVOA_X
ClientSampleId :
TW-WTS-06


Tgt Ion: 95 Resp: 64264
Ion Ratio Lower Upper
95 100
174 68.3 0.0 135.8
176 65.3 0.0 131.4



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. 0.000 min
Lab File: VX045814.D
Acq: 16 Apr 2025 17:41

Tgt Ion: 117 Resp: 129813
Ion Ratio Lower Upper
117 100
82 62.5 49.2 73.8
119 30.8 25.1 37.7





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.018 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX045814.D


Acq: 16 Apr 2025 17:41

Instrument :

MSVOA_X

ClientSampleId :

TW-WTS-06





Tgt Ion:152 Resp: 52878

Ion Ratio Lower Upper

152 100

115 66.3 46.9 140.7

150 158.5 0.0 349.4





CALIBRATION SUMMARY



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: ENTA05
 Lab Code: CHEM Case No.: Q1822 SAS No.: Q1822 SDG No.: Q1822
 Instrument ID: MSVOA_X Calibration Date(s): 04/01/2025 04/01/2025
 Heated Purge: (Y/N) N Calibration Time(s): 17:06 19:02
 GC Column: DB-624UI ID: 0.18 (mm)

LAB FILE ID:		RRF001 = VX045525.D		RRF005 = VX045526.D		RRF020 = VX045527.D			
		RRF050 = VX045528.D		RRF100 = VX045529.D		RRF150 = VX045530.D			
COMPOUND	RRF001	RRF005	RRF020	RRF050	RRF100	RRF150	RRF	% RSD	
Methyl tert-butyl Ether	1.915	1.964	2.169	2.118	2.217	2.216	2.100	6.2	
Carbon Tetrachloride	0.450	0.497	0.521	0.536	0.545	0.556	0.518	7.5	
Chloroform	1.244	1.309	1.348	1.315	1.309	1.309	1.306	2.6	
1,1,1-Trichloroethane	1.028	1.052	1.120	1.109	1.153	1.167	1.105	5	
Benzene	1.414	1.416	1.519	1.481	1.483	1.465	1.463	2.8	
Toluene	0.817	0.866	0.939	0.910	0.905	0.875	0.885	4.8	
Tetrachloroethene	0.346	0.373	0.371	0.347	0.333	0.347	0.353	4.5	
Ethyl Benzene	1.608	1.819	2.002	2.007	1.993	2.053	1.914	8.9	
m/p-Xylenes	0.594	0.669	0.732	0.728	0.723	0.729	0.696	7.9	
o-Xylene	0.562	0.676	0.725	0.719	0.716	0.714	0.686	9.2	
1,2-Dichloroethane-d4		0.962	0.900	0.868	0.904	0.937	0.914	3.9	
Dibromofluoromethane		0.372	0.342	0.345	0.353	0.362	0.355	3.5	
Toluene-d8		1.257	1.233	1.214	1.230	1.257	1.238	1.5	
4-Bromofluorobenzene		0.413	0.448	0.453	0.481	0.460	0.451	5.5	

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.

Method Path : Z:\voasrv\HPCHEM1\MSVOA_X\Method\
 Method File : 82X040225W.M
 Title : SW846 8260
 Last Update : Wed Apr 02 03:11:43 2025
 Response Via : Initial Calibration

Calibration Files

1 =VX045525.D 5 =VX045526.D 20 =VX045527.D 50 =VX045528.D 100 =VX045529.D 150 =VX045530.D

	Compound	1	5	20	50	100	150	Avg	%RSD
1) I	Pentafluorobenzene	-----ISTD-----							
2) T	Dichlorodifluo...	0.615	0.695	0.700	0.820	0.837	0.820	0.748	12.13
3) P	Chloromethane	0.764	0.734	0.777	0.784	0.815	0.734	0.768	4.05
4) C	Vinyl Chloride	0.671	0.662	0.701	0.716	0.738	0.730	0.703	4.43#
5) T	Bromomethane		0.342	0.327	0.330	0.341	0.327	0.333	2.20
6) T	Chloroethane	0.378	0.397	0.390	0.398	0.355	0.319	0.373	8.27
7) T	Trichlorofluor...	1.051	0.999	1.075	1.086	1.089	0.989	1.048	4.22
8) T	Diethyl Ether	0.345	0.340	0.356	0.355	0.358	0.357	0.352	2.11
9) T	1,1,2-Trichlor...	0.575	0.599	0.635	0.621	0.621	0.634	0.614	3.74
10) T	Methyl Iodide		0.705	0.789	0.799	0.788	0.750	0.766	5.08
11) T	Tert butyl alc...		0.111	0.124	0.127	0.128	0.124	0.123	5.44
12) CM	1,1-Dichloroet...	0.563	0.588	0.612	0.604	0.618	0.616	0.600	3.51#
13) T	Acrolein		0.176	0.161	0.167	0.170	0.172	0.169	3.41
14) T	Allyl chloride	1.092	1.068	1.142	1.172	1.199	1.160	1.139	4.36
15) T	Acrylonitrile	0.356	0.382	0.413	0.407	0.394	0.376	0.388	5.38
16) T	Acetone	0.400	0.369	0.387	0.378	0.365	0.355	0.375	4.28
17) T	Carbon Disulfide	1.334	1.327	1.434	1.553	1.604	1.642	1.483	9.24
18) T	Methyl Acetate	0.901	0.863	0.864	0.857	0.855	0.846	0.864	2.22
19) T	Methyl tert-bu...	1.915	1.964	2.169	2.118	2.217	2.216	2.100	6.20
20) T	Methylene Chlo...	0.692	0.695	0.726	0.709	0.705	0.690	0.703	1.96
21) T	trans-1,2-Dich...	0.574	0.594	0.636	0.624	0.621	0.630	0.613	3.94
22) T	Diisopropyl ether	2.003	2.120	2.329	2.293	2.359	2.357	2.243	6.59
23) T	Vinyl Acetate	1.542	1.693	2.005	2.087	2.147	2.156	1.938	13.34
24) P	1,1-Dichloroet...	1.211	1.240	1.318	1.269	1.283	1.292	1.269	3.01
25) T	2-Butanone	0.474	0.537	0.582	0.586	0.571	0.548	0.550	7.55
26) T	2,2-Dichloropr...	0.676	0.757	0.802	0.847	0.903	0.933	0.820	11.67
27) T	cis-1,2-Dichlo...	0.762	0.696	0.760	0.751	0.754	0.760	0.747	3.43
28) T	Bromochloromet...	0.671	0.608	0.617	0.596	0.610	0.576	0.613	5.18
29) T	Tetrahydrofuran	0.339	0.341	0.371	0.371	0.364	0.352	0.356	4.05
30) C	Chloroform	1.244	1.309	1.348	1.315	1.309	1.309	1.306	2.60#
31) T	Cyclohexane		1.044	1.148	1.123	1.145	1.149	1.122	4.00
32) T	1,1,1-Trichlor...	1.028	1.052	1.120	1.109	1.153	1.167	1.105	4.96
33) S	1,2-Dichloroet...		0.962	0.900	0.868	0.904	0.937	0.914	3.94
34) I	1,4-Difluorobenzene	-----ISTD-----							
35) S	Dibromofluorom...		0.372	0.342	0.345	0.353	0.362	0.355	3.48
36) T	1,1-Dichloropr...	0.477	0.443	0.474	0.485	0.496	0.499	0.479	4.26
37) T	Ethyl Acetate	0.582	0.565	0.609	0.623	0.629	0.614	0.604	4.15
38) T	Carbon Tetrach...	0.450	0.497	0.521	0.536	0.545	0.556	0.518	7.49
39) T	Methylcyclohexane	0.519	0.527	0.596	0.622	0.628	0.632	0.587	8.76
40) TM	Benzene	1.414	1.416	1.519	1.481	1.483	1.465	1.463	2.81
41) T	Methacrylonitrile	0.252	0.276	0.338	0.351	0.350	0.343	0.318	13.52
42) TM	1,2-Dichloroet...	0.533	0.585	0.649	0.622	0.620	0.617	0.604	6.72
43) T	Isopropyl Acetate	0.735	0.831	0.952	0.978	1.002	0.991	0.915	11.77
44) TM	Trichloroethene	0.349	0.322	0.356	0.351	0.351	0.356	0.348	3.67
45) C	1,2-Dichloropr...	0.309	0.365	0.388	0.376	0.379	0.374	0.365	7.83#
46) T	Dibromomethane	0.237	0.287	0.297	0.292	0.286	0.282	0.280	7.76
47) T	Bromodichlorom...	0.521	0.519	0.576	0.571	0.587	0.583	0.560	5.56
48) T	Methyl methacr...	0.405	0.416	0.486	0.508	0.518	0.502	0.472	10.44
49) T	1,4-Dioxane	0.007	0.010	0.009	0.009	0.009	0.008	0.009	13.93
50) S	Toluene-d8		1.257	1.233	1.214	1.230	1.257	1.238	1.50
51) T	4-Methyl-2-Pen...	0.499	0.578	0.661	0.663	0.647	0.589	0.606	10.57
52) CM	Toluene	0.817	0.866	0.939	0.910	0.905	0.875	0.885	4.79#
53) T	t-1,3-Dichloro...	0.316	0.400	0.465	0.508	0.555	0.558	0.467	20.27
54) T	cis-1,3-Dichlo...	0.371	0.474	0.546	0.568	0.597	0.600	0.526	16.88
55) T	1,1,2-Trichlor...	0.337	0.346	0.376	0.359	0.358	0.340	0.353	4.11
56) T	Ethyl methacry...	0.407	0.482	0.562	0.600	0.629	0.595	0.546	15.54


Method Path : Z:\voasrv\HPCHEM1\MSVOA_X\Method\
 Method File : 82X040225W.M

57)	T	1,3-Dichloropr...	0.555	0.614	0.652	0.633	0.630	0.601	0.614	5.49
58)	T	2-Chloroethyl ...	0.221	0.257	0.286	0.298	0.301	0.293	0.276	11.26
59)	T	2-Hexanone	0.363	0.429	0.488	0.492	0.484	0.439	0.449	11.14
60)	T	Dibromochlorom...	0.313	0.352	0.404	0.412	0.421	0.405	0.385	11.06
61)	T	1,2-Dibromoethane	0.294	0.351	0.380	0.373	0.380	0.364	0.357	9.12
62)	S	4-Bromofluorob...		0.413	0.448	0.453	0.481	0.460	0.451	5.45
63)	I	Chlorobenzene-d5	-----ISTD-----							
64)	T	Tetrachloroethene	0.346	0.373	0.371	0.347	0.333	0.347	0.353	4.47
65)	PM	Chlorobenzene	0.951	1.054	1.123	1.084	1.086	1.112	1.068	5.83
66)	T	1,1,1,2-Tetrac...	0.368	0.344	0.372	0.373	0.379	0.390	0.371	4.18
67)	C	Ethyl Benzene	1.608	1.819	2.002	2.007	1.993	2.053	1.914	8.89#
68)	T	m/p-Xylenes	0.594	0.669	0.732	0.728	0.723	0.729	0.696	7.92
69)	T	o-Xylene	0.562	0.676	0.725	0.719	0.716	0.714	0.686	9.17
70)	T	Styrene	0.897	1.043	1.202	1.216	1.230	1.202	1.132	11.82
71)	P	Bromoform	0.220	0.252	0.272	0.296	0.307	0.315	0.277	13.07
72)	I	1,4-Dichlorobenzen...	-----ISTD-----							
73)	T	Isopropylbenzene	3.581	3.850	4.224	4.181	4.043	4.151	4.005	6.17
74)	T	N-amyl acetate	1.442	1.682	1.960	2.058	2.132	2.221	1.916	15.51
75)	P	1,1,2,2-Tetrac...	1.457	1.428	1.461	1.384	1.354	1.358	1.407	3.43
76)	T	1,2,3-Trichlor...	1.221	1.257	1.251	1.225	1.192	1.171	1.220	2.71
77)	T	Bromobenzene	0.877	0.927	0.953	0.938	0.925	0.932	0.925	2.77
78)	T	n-propylbenzene	3.734	4.346	5.012	4.896	4.819	4.872	4.613	10.59
79)	T	2-Chlorotoluene	2.894	2.873	3.130	2.966	2.915	2.918	2.949	3.19
80)	T	1,3,5-Trimethy...	2.931	3.119	3.571	3.511	3.367	3.371	3.312	7.35
81)	T	trans-1,4-Dich...		0.274	0.323	0.371	0.401	0.428	0.359	17.16
82)	T	4-Chlorotoluene	2.961	3.156	3.470	3.449	3.350	3.343	3.288	5.93
83)	T	tert-Butylbenzene	2.894	3.124	3.437	3.427	3.387	3.405	3.279	6.78
84)	T	1,2,4-Trimethy...	2.906	3.129	3.529	3.523	3.438	3.420	3.324	7.57
85)	T	sec-Butylbenzene	3.195	3.828	4.316	4.300	4.232	4.292	4.027	11.12
86)	T	p-Isopropyltol...	2.734	3.157	3.515	3.524	3.506	3.484	3.320	9.63
87)	T	1,3-Dichlorobe...	1.658	1.605	1.726	1.699	1.714	1.706	1.684	2.70
88)	T	1,4-Dichlorobe...	1.671	1.724	1.768	1.703	1.674	1.706	1.708	2.09
89)	T	n-Butylbenzene	2.117	2.486	2.974	3.160	3.256	3.283	2.879	16.50
90)	T	Hexachloroethane	0.493	0.493	0.556	0.597	0.623	0.661	0.571	12.14
91)	T	1,2-Dichlorobe...	1.644	1.645	1.750	1.678	1.665	1.667	1.675	2.34
92)	T	1,2-Dibromo-3-...	0.196	0.260	0.312	0.316	0.333	0.349	0.294	19.37
93)	T	1,2,4-Trichlor...	0.727	0.844	0.948	0.947	1.045	1.083	0.932	14.07
94)	T	Hexachlorobuta...	0.379	0.389	0.401	0.404	0.415	0.422	0.402	4.03
95)	T	Naphthalene	2.544	3.070	3.630	3.722	3.867	3.982	3.469	15.93
96)	T	1,2,3-Trichlor...	0.782	0.916	0.999	1.000	1.063	1.097	0.976	11.63

(#) = Out of Range

t-1,3-Dichloropropene

Response Ratio



$$\text{Response} = 5.639\text{e-}001 * \text{Amt} - 2.582\text{e-}002$$

Coef of Det (r^2) = 0.999232 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA X\Method\82X040225W.M

Calibration Table Last Updated: Wed Apr 02 03:11:43 2025

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
 Data File : VX045525.D
 Acq On : 01 Apr 2025 17:06
 Operator : JC/MD
 Sample : VSTDICC001
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC001

Manual Integrations
 APPROVED

Reviewed By :Amit Patel 04/02/2025
 Supervised By :Mahesh Dadoda 04/02/2025

Quant Time: Apr 02 02:50:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 02:44:48 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	96737	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	175266	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	153312	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	61114	50.000	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	0.000	65	0d	0.000	ug/l	
Spiked Amount	50.000	Range	74 - 125	Recovery	=	0.000%#
35) Dibromofluoromethane	0.000	113	0d	0.000	ug/l	
Spiked Amount	50.000	Range	75 - 124	Recovery	=	0.000%#
50) Toluene-d8	0.000	98	0d	0.000	ug/l	
Spiked Amount	50.000	Range	86 - 113	Recovery	=	0.000%#
62) 4-Bromofluorobenzene	0.000	95	0d	0.000	ug/l	
Spiked Amount	50.000	Range	77 - 121	Recovery	=	0.000%#

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.167	85	1189	0.822	ug/l	94
3) Chloromethane	1.307	50	1479	0.995	ug/l #	86
4) Vinyl Chloride	1.374	62	1299	0.955	ug/l	98
6) Chloroethane	1.673	64	732	1.014	ug/l #	67
7) Trichlorofluoromethane	1.880	101	2034	1.003	ug/l	93
8) Diethyl Ether	2.130	74	668	0.981	ug/l	83
9) 1,1,2-Trichlorotrifluo...	2.325	101	1113	0.936	ug/l	94
12) 1,1-Dichloroethene	2.313	96	1090	0.939	ug/l	95
14) Allyl chloride	2.660	41	2113	0.959	ug/l	93
15) Acrylonitrile	3.075	53	3446	4.592	ug/l	95
16) Acetone	2.380	43	3867	5.323	ug/l	98
17) Carbon Disulfide	2.508	76	2580	0.899	ug/l #	95
18) Methyl Acetate	2.703	43	1744	1.043	ug/l	91
19) Methyl tert-butyl Ether	3.117	73	3705	0.912	ug/l	92
20) Methylene Chloride	2.782	84	1338	0.984	ug/l	91
21) trans-1,2-Dichloroethene	3.087	96	1111	0.936	ug/l #	93
22) Diisopropyl ether	3.758	45	3875	0.893	ug/l #	59
23) Vinyl Acetate	3.727	43	14921	3.979	ug/l	95
24) 1,1-Dichloroethane	3.611	63	2343	0.954	ug/l #	90
25) 2-Butanone	4.581	43	4590	4.317	ug/l	92
26) 2,2-Dichloropropane	4.471	77	1307	0.824	ug/l	85
27) cis-1,2-Dichloroethene	4.495	96	1475	1.020	ug/l	80
28) Bromochloromethane	4.891	49	1298	1.094	ug/l #	97
29) Tetrahydrofuran	5.032	42	3276	4.752	ug/l	97
30) Chloroform	5.093	83	2406	0.953	ug/l	84
32) 1,1,1-Trichloroethane	5.385	97	1989	0.931	ug/l #	53
36) 1,1-Dichloropropene	5.696	75	1671	0.995	ug/l #	89
37) Ethyl Acetate	4.751	43	2040	0.964	ug/l #	76
38) Carbon Tetrachloride	5.672	117	1578	0.870	ug/l #	91
39) Methylcyclohexane	7.379	83	1818	0.883	ug/l #	89
40) Benzene	6.038	78	4957	0.967	ug/l	97
41) Methacrylonitrile	4.952	41	883m	0.784	ug/l	
42) 1,2-Dichloroethane	6.092	62	1867	0.882	ug/l	86
43) Isopropyl Acetate	6.361	43	2578	0.804	ug/l #	86
44) Trichloroethene	7.123	130	1224	1.004	ug/l	87
45) 1,2-Dichloropropane	7.434	63	1082	0.846	ug/l	90

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
 Data File : VX045525.D
 Acq On : 01 Apr 2025 17:06
 Operator : JC/MD
 Sample : VSTDICC001
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC001

Manual Integrations
 APPROVED

Reviewed By :Amit Patel 04/02/2025
 Supervised By :Mahesh Dadoda 04/02/2025

Quant Time: Apr 02 02:50:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 02:44:48 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) Dibromomethane	7.592	93	831	0.846	ug/l	90
47) Bromodichloromethane	7.824	83	1827	0.931	ug/l #	77
48) Methyl methacrylate	7.720	41	1419	0.857	ug/l	92
49) 1,4-Dioxane	7.671	88	458m	15.288	ug/l	
51) 4-Methyl-2-Pentanone	8.574	43	8746	4.114	ug/l	100
52) Toluene	8.720	92	2863	0.923	ug/l	96
53) t-1,3-Dichloropropene	8.988	75	1109	2.850	ug/l	93
54) cis-1,3-Dichloropropene	8.379	75	1300	0.705	ug/l	90
55) 1,1,2-Trichloroethane	9.159	97	1183	0.957	ug/l #	76
56) Ethyl methacrylate	9.129	69	1425m	0.745	ug/l	
57) 1,3-Dichloropropane	9.311	76	1946	0.904	ug/l	91
58) 2-Chloroethyl Vinyl ether	8.245	63	3880	4.008	ug/l	93
59) 2-Hexanone	9.439	43	6360	4.040	ug/l	85
60) Dibromochloromethane	9.519	129	1097	0.814	ug/l	99
61) 1,2-Dibromoethane	9.610	107	1032	0.825	ug/l	96
64) Tetrachloroethene	9.269	164	1062	0.981	ug/l	92
65) Chlorobenzene	10.080	112	2916	0.890	ug/l	95
66) 1,1,1,2-Tetrachloroethane	10.159	131	1128	0.992	ug/l #	64
67) Ethyl Benzene	10.195	91	4930	0.840	ug/l	90
68) m/p-Xylenes	10.305	106	3645	1.708	ug/l	97
69) o-Xylene	10.640	106	1724	0.820	ug/l	83
70) Styrene	10.659	104	2750	0.793	ug/l	94
71) Bromoform	10.805	173	675	0.795	ug/l #	94
73) Isopropylbenzene	10.964	105	4377	0.894	ug/l	95
74) N-amyl acetate	10.854	43	1762	0.752	ug/l #	93
75) 1,1,2,2-Tetrachloroethane	11.214	83	1781	1.036	ug/l	94
76) 1,2,3-Trichloropropane	11.244	75	1493m	1.002	ug/l	
77) Bromobenzene	11.201	156	1072	0.948	ug/l	95
78) n-propylbenzene	11.305	91	4564	0.809	ug/l	97
79) 2-Chlorotoluene	11.366	91	3537	0.981	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	3582	0.885	ug/l	100
82) 4-Chlorotoluene	11.457	91	3619	0.900	ug/l	96
83) tert-Butylbenzene	11.713	119	3537	0.883	ug/l	94
84) 1,2,4-Trimethylbenzene	11.750	105	3552	0.874	ug/l	99
85) sec-Butylbenzene	11.890	105	3905	0.793	ug/l	98
86) p-Isopropyltoluene	12.012	119	3342	0.824	ug/l	94
87) 1,3-Dichlorobenzene	11.969	146	2026	0.984	ug/l	97
88) 1,4-Dichlorobenzene	12.043	146	2042m	0.978	ug/l	
89) n-Butylbenzene	12.335	91	2587	0.735	ug/l	99
90) Hexachloroethane	12.543	117	602	0.863	ug/l	95
91) 1,2-Dichlorobenzene	12.335	146	2009	0.981	ug/l	93
92) 1,2-Dibromo-3-Chloropr...	12.945	75	239	0.665	ug/l #	73
93) 1,2,4-Trichlorobenzene	13.591	180	888	0.779	ug/l	94
94) Hexachlorobutadiene	13.725	225	463	0.943	ug/l	90
95) Naphthalene	13.780	128	3109	0.733	ug/l	99
96) 1,2,3-Trichlorobenzene	13.969	180	956	0.801	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
Data File : VX045525.D
Acq On : 01 Apr 2025 17:06
Operator : JC/MD
Sample : VSTDIC001
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025

Quant Time: Apr 02 02:50:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 02:44:48 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
 Data File : VX045526.D
 Acq On : 01 Apr 2025 17:29
 Operator : JC/MD
 Sample : VSTDICC005
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC005

Manual Integrations
 APPROVED

Reviewed By :Amit Patel 04/02/2025
 Supervised By :Mahesh Dadoda 04/02/2025

Quant Time: Apr 02 02:51:06 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 02:44:48 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	93234	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	166137	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	143937	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	63909	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	8969	5.260	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	10.520%#		
35) Dibromofluoromethane	5.391	113	6181	5.244	ug/l	0.01
Spiked Amount 50.000	Range 75 - 124		Recovery =	10.480%#		
50) Toluene-d8	8.647	98	20890	5.077	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	10.160%#		
62) 4-Bromofluorobenzene	11.079	95	6867	4.582	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	9.160%#		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.166	85	6482	4.649	ug/l	91
3) Chloromethane	1.307	50	6846	4.779	ug/l	99
4) Vinyl Chloride	1.374	62	6169	4.706	ug/l	100
5) Bromomethane	1.593	94	3185	5.124	ug/l	90
6) Chloroethane	1.666	64	3703	5.324	ug/l	96
7) Trichlorofluoromethane	1.874	101	9316	4.766	ug/l	94
8) Diethyl Ether	2.136	74	3171	4.832	ug/l	96
9) 1,1,2-Trichlorotrifluo...	2.325	101	5589	4.879	ug/l	99
10) Methyl Iodide	2.447	142	6574	4.601	ug/l	97
11) Tert butyl alcohol	2.983	59	5186	22.633	ug/l	99
12) 1,1-Dichloroethene	2.313	96	5485	4.901	ug/l	90
13) Acrolein	2.239	56	8202	25.993	ug/l	97
14) Allyl chloride	2.654	41	9960	4.690	ug/l	95
15) Acrylonitrile	3.069	53	17811	24.624	ug/l	97
16) Acetone	2.386	43	17188	24.550	ug/l	99
17) Carbon Disulfide	2.502	76	12376	4.477	ug/l #	93
18) Methyl Acetate	2.703	43	8042	4.989	ug/l	96
19) Methyl tert-butyl Ether	3.111	73	18314	4.677	ug/l	100
20) Methylene Chloride	2.782	84	6481	4.945	ug/l	96
21) trans-1,2-Dichloroethene	3.087	96	5534	4.840	ug/l	95
22) Diisopropyl ether	3.764	45	19765	4.725	ug/l #	82
23) Vinyl Acetate	3.721	43	78907	21.831	ug/l	98
24) 1,1-Dichloroethane	3.605	63	11562	4.887	ug/l	98
25) 2-Butanone	4.574	43	25026	24.419	ug/l	99
26) 2,2-Dichloropropane	4.477	77	7054	4.615	ug/l	95
27) cis-1,2-Dichloroethene	4.483	96	6485	4.655	ug/l	95
28) Bromochloromethane	4.898	49	5666	4.956	ug/l	95
29) Tetrahydrofuran	5.013	42	15913	23.950	ug/l	100
30) Chloroform	5.087	83	12200	5.012	ug/l	99
31) Cyclohexane	5.465	56	9731	4.653	ug/l	97
32) 1,1,1-Trichloroethane	5.379	97	9811	4.763	ug/l	99
36) 1,1-Dichloropropene	5.690	75	7354	4.621	ug/l	98
37) Ethyl Acetate	4.721	43	9384	4.679	ug/l #	91
38) Carbon Tetrachloride	5.666	117	8264	4.804	ug/l	99
39) Methylcyclohexane	7.379	83	8757	4.489	ug/l	95
40) Benzene	6.038	78	23522	4.839	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
 Data File : VX045526.D
 Acq On : 01 Apr 2025 17:29
 Operator : JC/MD
 Sample : VSTDICC005
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC005

Manual Integrations
 APPROVED

Reviewed By :Amit Patel 04/02/2025
 Supervised By :Mahesh Dadoda 04/02/2025

Quant Time: Apr 02 02:51:06 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 02:44:48 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	4593	4.302	ug/l	92
42) 1,2-Dichloroethane	6.092	62	9713	4.838	ug/l	100
43) Isopropyl Acetate	6.348	43	13803	4.541	ug/l	97
44) Trichloroethene	7.129	130	5355	4.635	ug/l	99
45) 1,2-Dichloropropane	7.434	63	6060	4.997	ug/l	99
46) Dibromomethane	7.580	93	4767	5.120	ug/l	99
47) Bromodichloromethane	7.818	83	8619	4.636	ug/l #	97
48) Methyl methacrylate	7.696	41	6908	4.401	ug/l	98
49) 1,4-Dioxane	7.665	88	3205	112.864	ug/l	95
51) 4-Methyl-2-Pentanone	8.574	43	48050	23.844	ug/l	97
52) Toluene	8.714	92	14391	4.893	ug/l	100
53) t-1,3-Dichloropropene	8.982	75	6643	5.834	ug/l	97
54) cis-1,3-Dichloropropene	8.366	75	7873	4.506	ug/l	95
55) 1,1,2-Trichloroethane	9.153	97	5741	4.898	ug/l	90
56) Ethyl methacrylate	9.116	69	8012	4.417	ug/l	95
57) 1,3-Dichloropropane	9.305	76	10208	5.002	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.244	63	21382	23.302	ug/l	99
59) 2-Hexanone	9.433	43	35628	23.875	ug/l	99
60) Dibromochloromethane	9.519	129	5855	4.582	ug/l	99
61) 1,2-Dibromoethane	9.610	107	5837	4.920	ug/l	98
64) Tetrachloroethene	9.275	164	5373	5.287	ug/l	93
65) Chlorobenzene	10.073	112	15178	4.935	ug/l	96
66) 1,1,1,2-Tetrachloroethane	10.159	131	4945	4.631	ug/l	97
67) Ethyl Benzene	10.195	91	26180	4.753	ug/l	97
68) m/p-Xylenes	10.299	106	19268	9.617	ug/l	99
69) o-Xylene	10.640	106	9735	4.932	ug/l	99
70) Styrene	10.653	104	15017	4.610	ug/l	96
71) Bromoform	10.799	173	3634	4.558	ug/l #	95
73) Isopropylbenzene	10.964	105	24604	4.806	ug/l	100
74) N-amyl acetate	10.848	43	10750	4.390	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.213	83	9127	5.075	ug/l	98
76) 1,2,3-Trichloropropane	11.238	75	8032m	5.152	ug/l	
77) Bromobenzene	11.195	156	5927	5.011	ug/l	97
78) n-propylbenzene	11.305	91	27777	4.711	ug/l	98
79) 2-Chlorotoluene	11.360	91	18358	4.870	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	19936	4.710	ug/l	98
81) trans-1,4-Dichloro-2-b...	11.018	75	1752	3.814	ug/l	85
82) 4-Chlorotoluene	11.457	91	20171	4.799	ug/l	99
83) tert-Butylbenzene	11.713	119	19962	4.763	ug/l	98
84) 1,2,4-Trimethylbenzene	11.750	105	19998	4.707	ug/l	99
85) sec-Butylbenzene	11.890	105	24465	4.753	ug/l	100
86) p-Isopropyltoluene	12.006	119	20178	4.755	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	10255	4.763	ug/l	98
88) 1,4-Dichlorobenzene	12.036	146	11016	5.047	ug/l	93
89) n-Butylbenzene	12.329	91	15887	4.317	ug/l	98
90) Hexachloroethane	12.536	117	3151	4.320	ug/l	98
91) 1,2-Dichlorobenzene	12.335	146	10510	4.910	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	12.945	75	1661	4.417	ug/l	92
93) 1,2,4-Trichlorobenzene	13.591	180	5391	4.524	ug/l	95
94) Hexachlorobutadiene	13.719	225	2483	4.836	ug/l	98
95) Naphthalene	13.774	128	19618	4.424	ug/l	97
96) 1,2,3-Trichlorobenzene	13.957	180	5854	4.692	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
Data File : VX045526.D
Acq On : 01 Apr 2025 17:29
Operator : JC/MD
Sample : VSTDICC005
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025

Quant Time: Apr 02 02:51:06 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 02:44:48 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
Data File : VX045526.D
Acq On : 01 Apr 2025 17:29
Operator : JC/MD
Sample : VSTDICC005
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025

Quant Time: Apr 02 02:51:06 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 02:44:48 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
 Data File : VX045527.D
 Acq On : 01 Apr 2025 17:52
 Operator : JC/MD
 Sample : VSTDICC020
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC020

Manual Integrations
 APPROVED

Reviewed By :Amit Patel 04/02/2025
 Supervised By :Mahesh Dadoda 04/02/2025

Quant Time: Apr 02 02:52:02 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 02:44:48 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	90228	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	160754	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	144087	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	65341	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	32500	19.697	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	39.400%#		
35) Dibromofluoromethane	5.385	113	21963	19.256	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	38.520%#		
50) Toluene-d8	8.647	98	79303	19.921	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	39.840%#		
62) 4-Bromofluorobenzene	11.079	95	28795	19.858	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	39.720%#		
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.166	85	25265	18.724	ug/l	97
3) Chloromethane	1.301	50	28034	20.223	ug/l	98
4) Vinyl Chloride	1.374	62	25285	19.931	ug/l	97
5) Bromomethane	1.593	94	11815	19.642	ug/l	99
6) Chloroethane	1.666	64	14071	20.905	ug/l	98
7) Trichlorofluoromethane	1.874	101	38815	20.518	ug/l	98
8) Diethyl Ether	2.136	74	12852	20.235	ug/l	99
9) 1,1,2-Trichlorotrifluo...	2.319	101	22917	20.673	ug/l	95
10) Methyl Iodide	2.441	142	28478	20.596	ug/l	99
11) Tert butyl alcohol	2.983	59	22453	101.253	ug/l	100
12) 1,1-Dichloroethene	2.306	96	22103	20.406	ug/l	96
13) Acrolein	2.239	56	28991	94.938	ug/l	99
14) Allyl chloride	2.660	41	41233	20.063	ug/l	99
15) Acrylonitrile	3.062	53	74487	106.412	ug/l	98
16) Acetone	2.386	43	69788	103.000	ug/l	99
17) Carbon Disulfide	2.502	76	51771	19.352	ug/l	98
18) Methyl Acetate	2.703	43	31190	19.994	ug/l	99
19) Methyl tert-butyl Ether	3.117	73	78284	20.660	ug/l	96
20) Methylene Chloride	2.782	84	26212	20.664	ug/l	96
21) trans-1,2-Dichloroethene	3.087	96	22968	20.755	ug/l	92
22) Diisopropyl ether	3.757	45	84046	20.760	ug/l #	81
23) Vinyl Acetate	3.721	43	361881	103.458	ug/l	99
24) 1,1-Dichloroethane	3.605	63	47560	20.772	ug/l	99
25) 2-Butanone	4.556	43	105062	105.931	ug/l	99
26) 2,2-Dichloropropane	4.471	77	28956	19.574	ug/l	99
27) cis-1,2-Dichloroethene	4.483	96	27425	20.342	ug/l	98
28) Bromochloromethane	4.897	49	22285	20.141	ug/l	100
29) Tetrahydrofuran	5.007	42	66994	104.187	ug/l	99
30) Chloroform	5.093	83	48639	20.646	ug/l	100
31) Cyclohexane	5.458	56	41450	20.478	ug/l	99
32) 1,1,1-Trichloroethane	5.373	97	40407	20.270	ug/l	99
36) 1,1-Dichloropropene	5.684	75	30504	19.808	ug/l	98
37) Ethyl Acetate	4.715	43	39163	20.180	ug/l	98
38) Carbon Tetrachloride	5.672	117	33533	20.148	ug/l	94
39) Methylcyclohexane	7.373	83	38303	20.291	ug/l	97
40) Benzene	6.031	78	97667	20.766	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
 Data File : VX045527.D
 Acq On : 01 Apr 2025 17:52
 Operator : JC/MD
 Sample : VSTDICC020
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC020

Manual Integrations
 APPROVED

Reviewed By :Amit Patel 04/02/2025
 Supervised By :Mahesh Dadoda 04/02/2025

Quant Time: Apr 02 02:52:02 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 02:44:48 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	21740	21.046	ug/l	97
42) 1,2-Dichloroethane	6.086	62	41733	21.484	ug/l	100
43) Isopropyl Acetate	6.342	43	61222	20.814	ug/l	99
44) Trichloroethene	7.123	130	22918	20.501	ug/l	97
45) 1,2-Dichloropropane	7.428	63	24926	21.241	ug/l	90
46) Dibromomethane	7.580	93	19104	21.207	ug/l	99
47) Bromodichloromethane	7.818	83	37036	20.587	ug/l	99
48) Methyl methacrylate	7.696	41	31247	20.575	ug/l	99
49) 1,4-Dioxane	7.659	88	12104	440.515	ug/l	96
51) 4-Methyl-2-Pentanone	8.574	43	212645	109.055	ug/l	99
52) Toluene	8.714	92	60357	21.208	ug/l	99
53) t-1,3-Dichloropropene	8.976	75	29870	18.764	ug/l	96
54) cis-1,3-Dichloropropene	8.366	75	35099	20.762	ug/l	95
55) 1,1,2-Trichloroethane	9.147	97	24191	21.331	ug/l	96
56) Ethyl methacrylate	9.116	69	36155	20.600	ug/l	99
57) 1,3-Dichloropropane	9.305	76	41900	21.217	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.238	63	91923	103.534	ug/l	100
59) 2-Hexanone	9.427	43	156999	108.730	ug/l	99
60) Dibromochloromethane	9.519	129	25992	21.023	ug/l	98
61) 1,2-Dibromoethane	9.604	107	24456	21.303	ug/l	99
64) Tetrachloroethene	9.269	164	21379	21.015	ug/l	98
65) Chlorobenzene	10.079	112	64727	21.023	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.159	131	21447	20.064	ug/l	98
67) Ethyl Benzene	10.189	91	115368	20.921	ug/l	99
68) m/p-Xylenes	10.299	106	84420	42.092	ug/l	99
69) o-Xylene	10.640	106	41803	21.158	ug/l	100
70) Styrene	10.653	104	69290	21.249	ug/l	99
71) Bromoform	10.799	173	15659	19.619	ug/l #	99
73) Isopropylbenzene	10.957	105	110396	21.093	ug/l	100
74) N-amyl acetate	10.842	43	51238	20.466	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.207	83	38177	20.764	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	32690m	20.510	ug/l	
77) Bromobenzene	11.195	156	24902	20.592	ug/l	100
78) n-propylbenzene	11.299	91	130994	21.729	ug/l	98
79) 2-Chlorotoluene	11.360	91	81817	21.229	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	93328	21.564	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.018	75	8435	17.961	ug/l	88
82) 4-Chlorotoluene	11.451	91	90704	21.108	ug/l	99
83) tert-Butylbenzene	11.713	119	89828	20.964	ug/l	98
84) 1,2,4-Trimethylbenzene	11.750	105	92242	21.235	ug/l	100
85) sec-Butylbenzene	11.890	105	112800	21.433	ug/l	100
86) p-Isopropyltoluene	12.006	119	91866	21.173	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	45105	20.490	ug/l	100
88) 1,4-Dichlorobenzene	12.036	146	46208	20.706	ug/l	99
89) n-Butylbenzene	12.329	91	77739	20.662	ug/l	100
90) Hexachloroethane	12.536	117	14544	19.505	ug/l	100
91) 1,2-Dichlorobenzene	12.335	146	45732	20.896	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	8153	21.203	ug/l	100
93) 1,2,4-Trichlorobenzene	13.585	180	24787	20.347	ug/l	98
94) Hexachlorobutadiene	13.725	225	10488	19.979	ug/l	97
95) Naphthalene	13.774	128	94875	20.928	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	26113	20.471	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
Data File : VX045527.D
Acq On : 01 Apr 2025 17:52
Operator : JC/MD
Sample : VSTDICC020
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025

Quant Time: Apr 02 02:52:02 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 02:44:48 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
Data File : VX045527.D
Acq On : 01 Apr 2025 17:52
Operator : JC/MD
Sample : VSTDIC020
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020

Quant Time: Apr 02 02:52:02 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 02:44:48 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
 Data File : VX045528.D
 Acq On : 01 Apr 2025 18:15
 Operator : JC/MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050

Manual Integrations
 APPROVED

Reviewed By :Amit Patel 04/02/2025
 Supervised By :Mahesh Dadoda 04/02/2025

Quant Time: Apr 02 02:53:01 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 02:44:48 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	95658	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	168287	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	150201	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	68786	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	83054	47.477	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	94.960%		
35) Dibromofluoromethane	5.379	113	58108	48.667	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	97.340%		
50) Toluene-d8	8.647	98	204338	49.031	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	98.060%		
62) 4-Bromofluorobenzene	11.079	95	76190	50.190	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	100.380%		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.166	85	78444	54.834	ug/l	100
3) Chloromethane	1.307	50	75032	51.054	ug/l	100
4) Vinyl Chloride	1.374	62	68516	50.942	ug/l	100
5) Bromomethane	1.593	94	31584	49.526	ug/l	100
6) Chloroethane	1.666	64	38112	53.409	ug/l	100
7) Trichlorofluoromethane	1.874	101	103888	51.798	ug/l	100
8) Diethyl Ether	2.136	74	33953	50.424	ug/l	100
9) 1,1,2-Trichlorotrifluo...	2.325	101	59431	50.569	ug/l	100
10) Methyl Iodide	2.447	142	76404	52.121	ug/l	100
11) Tert butyl alcohol	2.977	59	60610	257.809	ug/l	100
12) 1,1-Dichloroethene	2.313	96	57738	50.278	ug/l	100
13) Acrolein	2.239	56	79989	247.074	ug/l	100
14) Allyl chloride	2.660	41	112085	51.441	ug/l	100
15) Acrylonitrile	3.062	53	194445	262.015	ug/l	100
16) Acetone	2.386	43	180563	251.366	ug/l	100
17) Carbon Disulfide	2.508	76	148601	52.393	ug/l	100
18) Methyl Acetate	2.703	43	82000	49.581	ug/l	100
19) Methyl tert-butyl Ether	3.111	73	202582	50.429	ug/l	100
20) Methylene Chloride	2.782	84	67855	50.457	ug/l	100
21) trans-1,2-Dichloroethene	3.087	96	59696	50.882	ug/l	100
22) Diisopropyl ether	3.764	45	219374	51.111	ug/l	100
23) Vinyl Acetate	3.721	43	998064	269.139	ug/l	100
24) 1,1-Dichloroethane	3.605	63	121348	49.992	ug/l	100
25) 2-Butanone	4.556	43	280076	266.363	ug/l	100
26) 2,2-Dichloropropane	4.471	77	81054	51.683	ug/l	100
27) cis-1,2-Dichloroethene	4.483	96	71871	50.282	ug/l	100
28) Bromochloromethane	4.897	49	57046	48.632	ug/l	100
29) Tetrahydrofuran	5.001	42	177402	260.230	ug/l	100
30) Chloroform	5.086	83	125797	50.366	ug/l	100
31) Cyclohexane	5.464	56	107407	50.052	ug/l	100
32) 1,1,1-Trichloroethane	5.379	97	106062	50.184	ug/l	100
36) 1,1-Dichloropropene	5.684	75	81607	50.620	ug/l	100
37) Ethyl Acetate	4.715	43	104896	51.632	ug/l	100
38) Carbon Tetrachloride	5.672	117	90247	51.797	ug/l	100
39) Methylcyclohexane	7.373	83	104678	52.970	ug/l	100
40) Benzene	6.031	78	249161	50.605	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
 Data File : VX045528.D
 Acq On : 01 Apr 2025 18:15
 Operator : JC/MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050

Manual Integrations
 APPROVED

Reviewed By :Amit Patel 04/02/2025
 Supervised By :Mahesh Dadoda 04/02/2025

Quant Time: Apr 02 02:53:01 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 02:44:48 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	59068	54.623	ug/l	100
42) 1,2-Dichloroethane	6.080	62	104601	51.437	ug/l	100
43) Isopropyl Acetate	6.336	43	164588	53.452	ug/l	100
44) Trichloroethene	7.123	130	59042	50.452	ug/l	100
45) 1,2-Dichloropropane	7.427	63	63316	51.540	ug/l	100
46) Dibromomethane	7.574	93	49125	52.092	ug/l	100
47) Bromodichloromethane	7.818	83	96176	51.068	ug/l	100
48) Methyl methacrylate	7.690	41	85483	53.768	ug/l	100
49) 1,4-Dioxane	7.659	88	31199	1084.635	ug/l	100
51) 4-Methyl-2-Pentanone	8.574	43	558029	273.375	ug/l	100
52) Toluene	8.714	92	153143	51.402	ug/l	100
53) t-1,3-Dichloropropene	8.976	75	85462	47.317	ug/l	100
54) cis-1,3-Dichloropropene	8.366	75	95524	53.975	ug/l	100
55) 1,1,2-Trichloroethane	9.147	97	60452	50.919	ug/l	100
56) Ethyl methacrylate	9.116	69	100998	54.970	ug/l	100
57) 1,3-Dichloropropane	9.305	76	106580	51.554	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.238	63	250428	269.434	ug/l	100
59) 2-Hexanone	9.427	43	413824	273.765	ug/l	100
60) Dibromochloromethane	9.519	129	69258	53.511	ug/l	100
61) 1,2-Dibromoethane	9.604	107	62812	52.264	ug/l	100
64) Tetrachloroethene	9.269	164	52155	49.181	ug/l	100
65) Chlorobenzene	10.073	112	162829	50.733	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.159	131	55974	50.233	ug/l	100
67) Ethyl Benzene	10.189	91	301481	52.447	ug/l	100
68) m/p-Xylenes	10.299	106	218702	104.606	ug/l	100
69) o-Xylene	10.640	106	108028	52.452	ug/l	100
70) Styrene	10.653	104	182620	53.723	ug/l	100
71) Bromoform	10.799	173	44397	53.361	ug/l #	100
73) Isopropylbenzene	10.957	105	287628	52.203	ug/l	100
74) N-amyl acetate	10.842	43	141589	53.721	ug/l	100
75) 1,1,2,2-Tetrachloroethane	11.207	83	95200	49.185	ug/l	100
76) 1,2,3-Trichloropropane	11.238	75	84279m	50.230	ug/l	
77) Bromobenzene	11.195	156	64527	50.686	ug/l	100
78) n-propylbenzene	11.299	91	336796	53.069	ug/l	100
79) 2-Chlorotoluene	11.360	91	203994	50.279	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	241536	53.013	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	25488	51.556	ug/l	100
82) 4-Chlorotoluene	11.451	91	237213	52.439	ug/l	100
83) tert-Butylbenzene	11.713	119	235715	52.256	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	242301	52.986	ug/l	100
85) sec-Butylbenzene	11.890	105	295799	53.391	ug/l	100
86) p-Isopropyltoluene	12.006	119	242416	53.073	ug/l	100
87) 1,3-Dichlorobenzene	11.969	146	116890	50.441	ug/l	100
88) 1,4-Dichlorobenzene	12.036	146	117164	49.871	ug/l	100
89) n-Butylbenzene	12.329	91	217345	54.874	ug/l	100
90) Hexachloroethane	12.536	117	41095	52.352	ug/l	100
91) 1,2-Dichlorobenzene	12.335	146	115426	50.099	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	12.939	75	21712	53.638	ug/l	100
93) 1,2,4-Trichlorobenzene	13.585	180	65145	50.797	ug/l	100
94) Hexachlorobutadiene	13.719	225	27795	50.296	ug/l	100
95) Naphthalene	13.774	128	256009	53.643	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	68781	51.219	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
Data File : VX045528.D
Acq On : 01 Apr 2025 18:15
Operator : JC/MD
Sample : VSTDICCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations
APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025

Quant Time: Apr 02 02:53:01 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 02:44:48 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
 Data File : VX045529.D
 Acq On : 01 Apr 2025 18:38
 Operator : JC/MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC100

Manual Integrations
 APPROVED

Reviewed By :Amit Patel 04/02/2025
 Supervised By :Mahesh Dadoda 04/02/2025

Quant Time: Apr 02 02:54:00 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 02:44:48 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	86107	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	152724	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	136369	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	64612	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	155757	98.914	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 197.820%#			
35) Dibromofluoromethane	5.379	113	107908	99.585	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 199.160%#			
50) Toluene-d8	8.647	98	375599	99.309	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 198.620%#			
62) 4-Bromofluorobenzene	11.079	95	146944	106.664	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 213.320%#			
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.166	85	144108	111.909	ug/l	99
3) Chloromethane	1.307	50	140360	106.098	ug/l	100
4) Vinyl Chloride	1.374	62	127174	105.042	ug/l	99
5) Bromomethane	1.593	94	58706	102.266	ug/l	98
6) Chloroethane	1.660	64	61099	95.120	ug/l	98
7) Trichlorofluoromethane	1.867	101	187555	103.887	ug/l	98
8) Diethyl Ether	2.136	74	61683	101.767	ug/l	99
9) 1,1,2-Trichlorotrifluo...	2.319	101	106989	101.133	ug/l	97
10) Methyl Iodide	2.441	142	135732	102.863	ug/l	100
11) Tert butyl alcohol	2.989	59	110044	519.999	ug/l	99
12) 1,1-Dichloroethene	2.306	96	106404	102.934	ug/l	99
13) Acrolein	2.233	56	146268	501.914	ug/l	99
14) Allyl chloride	2.654	41	206484	105.276	ug/l	99
15) Acrylonitrile	3.062	53	338970	507.426	ug/l	99
16) Acetone	2.386	43	314285	486.054	ug/l	100
17) Carbon Disulfide	2.502	76	276285	108.215	ug/l	100
18) Methyl Acetate	2.703	43	147272	98.925	ug/l	99
19) Methyl tert-butyl Ether	3.111	73	381719	105.561	ug/l	99
20) Methylene Chloride	2.782	84	121435	100.316	ug/l	98
21) trans-1,2-Dichloroethene	3.081	96	106896	101.220	ug/l	98
22) Diisopropyl ether	3.757	45	406185	105.132	ug/l	95
23) Vinyl Acetate	3.721	43	1848817	553.853	ug/l	99
24) 1,1-Dichloroethane	3.605	63	220934	101.114	ug/l	100
25) 2-Butanone	4.556	43	491249	519.018	ug/l	100
26) 2,2-Dichloropropane	4.465	77	155592	110.215	ug/l	100
27) cis-1,2-Dichloroethene	4.483	96	129785	100.871	ug/l	98
28) Bromochloromethane	4.891	49	105093	99.530	ug/l	100
29) Tetrahydrofuran	5.001	42	313121	510.264	ug/l	99
30) Chloroform	5.086	83	225402	100.255	ug/l	99
31) Cyclohexane	5.458	56	197136	102.056	ug/l	98
32) 1,1,1-Trichloroethane	5.379	97	198511	104.346	ug/l	99
36) 1,1-Dichloropropene	5.684	75	151448	103.515	ug/l	99
37) Ethyl Acetate	4.715	43	192130	104.207	ug/l	99
38) Carbon Tetrachloride	5.672	117	166465	105.278	ug/l	99
39) Methylcyclohexane	7.373	83	191752	106.921	ug/l	99
40) Benzene	6.031	78	452946	101.369	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
 Data File : VX045529.D
 Acq On : 01 Apr 2025 18:38
 Operator : JC/MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC100

Manual Integrations
 APPROVED

Reviewed By :Amit Patel 04/02/2025
 Supervised By :Mahesh Dadoda 04/02/2025

Quant Time: Apr 02 02:54:00 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 02:44:48 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	107037	109.068	ug/l	99
42) 1,2-Dichloroethane	6.080	62	189528	102.696	ug/l	100
43) Isopropyl Acetate	6.336	43	305974	109.494	ug/l	100
44) Trichloroethene	7.123	130	107332	101.062	ug/l	97
45) 1,2-Dichloropropane	7.421	63	115731	103.806	ug/l	98
46) Dibromomethane	7.574	93	87381	102.101	ug/l	97
47) Bromodichloromethane	7.818	83	179265	104.886	ug/l	98
48) Methyl methacrylate	7.690	41	158241	109.675	ug/l	99
49) 1,4-Dioxane	7.665	88	52624	2015.906	ug/l	97
51) 4-Methyl-2-Pentanone	8.574	43	988881	533.814	ug/l	99
52) Toluene	8.714	92	276323	102.199	ug/l	99
53) t-1,3-Dichloropropene	8.976	75	169413	100.645	ug/l	100
54) cis-1,3-Dichloropropene	8.366	75	182328	113.521	ug/l	99
55) 1,1,2-Trichloroethane	9.147	97	109208	101.361	ug/l	98
56) Ethyl methacrylate	9.116	69	192034	115.168	ug/l	99
57) 1,3-Dichloropropane	9.305	76	192473	102.588	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.238	63	460297	545.696	ug/l	100
59) 2-Hexanone	9.427	43	738772	538.538	ug/l	99
60) Dibromochloromethane	9.519	129	128700	109.570	ug/l	98
61) 1,2-Dibromoethane	9.604	107	115919	106.282	ug/l	99
64) Tetrachloroethene	9.269	164	90817	94.325	ug/l	99
65) Chlorobenzene	10.073	112	296300	101.682	ug/l	98
66) 1,1,1,2-Tetrachloroethane	10.159	131	103437	102.243	ug/l	99
67) Ethyl Benzene	10.189	91	543568	104.152	ug/l	99
68) m/p-Xylenes	10.299	106	394241	207.694	ug/l	100
69) o-Xylene	10.640	106	195381	104.487	ug/l	99
70) Styrene	10.653	104	335333	108.654	ug/l	100
71) Bromoform	10.799	173	83632	110.714	ug/l #	97
73) Isopropylbenzene	10.957	105	522436	100.944	ug/l	100
74) N-amyl acetate	10.842	43	275462	111.267	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.213	83	174977	96.243	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	154048m	97.744	ug/l	
77) Bromobenzene	11.195	156	119561	99.983	ug/l	99
78) n-propylbenzene	11.299	91	622679	104.454	ug/l	100
79) 2-Chlorotoluene	11.360	91	376682	98.840	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	435137	101.675	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	51873	111.704	ug/l	95
82) 4-Chlorotoluene	11.451	91	432884	101.877	ug/l	99
83) tert-Butylbenzene	11.713	119	437649	103.292	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	444211	103.414	ug/l	99
85) sec-Butylbenzene	11.890	105	546857	105.082	ug/l	100
86) p-Isopropyltoluene	12.006	119	453109	105.609	ug/l	100
87) 1,3-Dichlorobenzene	11.969	146	221481	101.749	ug/l	99
88) 1,4-Dichlorobenzene	12.036	146	216367	98.047	ug/l	98
89) n-Butylbenzene	12.329	91	420694	113.075	ug/l	99
90) Hexachloroethane	12.536	117	80489	109.160	ug/l	100
91) 1,2-Dichlorobenzene	12.335	146	215157	99.419	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	12.939	75	43047	113.214	ug/l	99
93) 1,2,4-Trichlorobenzene	13.585	180	135046	112.106	ug/l	99
94) Hexachlorobutadiene	13.725	225	53636	103.326	ug/l	98
95) Naphthalene	13.774	128	499719	111.473	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	137328	108.869	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
Data File : VX045529.D
Acq On : 01 Apr 2025 18:38
Operator : JC/MD
Sample : VSTDICC100
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025

Quant Time: Apr 02 02:54:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 02:44:48 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
Data File : VX045529.D
Acq On : 01 Apr 2025 18:38
Operator : JC/MD
Sample : VSTDICC100
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025

Quant Time: Apr 02 02:54:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 02:44:48 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
 Data File : VX045530.D
 Acq On : 01 Apr 2025 19:02
 Operator : JC/MD
 Sample : VSTDICC150
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC150

Quant Time: Apr 02 02:54:59 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 02:44:48 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Amit Patel 04/02/2025
 Supervised By :Mahesh Dadoda 04/02/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	90033	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	160037	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	133342	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	61380	50.000	ug/l	# 0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	252998	153.661	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 307.320%#			
35) Dibromofluoromethane	5.379	113	173598	152.887	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 305.780%#			
50) Toluene-d8	8.647	98	603264	152.216	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 304.440%#			
62) 4-Bromofluorobenzene	11.079	95	220921	153.035	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 306.060%#			
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.167	85	221436	164.460	ug/l	100
3) Chloromethane	1.307	50	198321	143.373	ug/l	98
4) Vinyl Chloride	1.374	62	197096	155.697	ug/l	100
5) Bromomethane	1.593	94	88225	146.986	ug/l	97
6) Chloroethane	1.660	64	86251	128.421	ug/l	97
7) Trichlorofluoromethane	1.868	101	267098	141.495	ug/l	100
8) Diethyl Ether	2.136	74	96467	152.215	ug/l	100
9) 1,1,2-Trichlorotrifluo...	2.319	101	171118	154.698	ug/l	97
10) Methyl Iodide	2.441	142	202594	146.838	ug/l	99
11) Tert butyl alcohol	2.989	59	167769	758.202	ug/l	100
12) 1,1-Dichloroethene	2.307	96	166365	153.921	ug/l	97
13) Acrolein	2.239	56	232818	764.070	ug/l	100
14) Allyl chloride	2.654	41	313268	152.755	ug/l	99
15) Acrylonitrile	3.069	53	507965	727.247	ug/l	100
16) Acetone	2.386	43	479564	709.323	ug/l	99
17) Carbon Disulfide	2.502	76	443495	166.133	ug/l	99
18) Methyl Acetate	2.703	43	228558	146.832	ug/l	99
19) Methyl tert-butyl Ether	3.117	73	598532	158.301	ug/l	99
20) Methylene Chloride	2.782	84	186382	147.254	ug/l	99
21) trans-1,2-Dichloroethene	3.081	96	170283	154.210	ug/l	98
22) Diisopropyl ether	3.764	45	636728	157.617	ug/l	94
23) Vinyl Acetate	3.721	43	2911326	834.120	ug/l	100
24) 1,1-Dichloroethane	3.605	63	349040	152.778	ug/l	99
25) 2-Butanone	4.556	43	740096	747.834	ug/l	99
26) 2,2-Dichloropropane	4.465	77	252063	170.765	ug/l	100
27) cis-1,2-Dichloroethene	4.483	96	205254	152.571	ug/l	98
28) Bromochloromethane	4.891	49	155600	140.938	ug/l	97
29) Tetrahydrofuran	5.001	42	475589	741.226	ug/l	100
30) Chloroform	5.087	83	353681	150.451	ug/l	98
31) Cyclohexane	5.458	56	310219	153.595	ug/l	97
32) 1,1,1-Trichloroethane	5.373	97	315136	158.426	ug/l	98
36) 1,1-Dichloropropene	5.684	75	239776	156.399	ug/l	98
37) Ethyl Acetate	4.715	43	294559	152.461	ug/l	99
38) Carbon Tetrachloride	5.672	117	266770	161.005	ug/l	99
39) Methylcyclohexane	7.373	83	303251	161.366	ug/l	96
40) Benzene	6.031	78	703311	150.208	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
 Data File : VX045530.D
 Acq On : 01 Apr 2025 19:02
 Operator : JC/MD
 Sample : VSTDICC150
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC150

Manual Integrations
 APPROVED

Reviewed By :Amit Patel 04/02/2025
 Supervised By :Mahesh Dadoda 04/02/2025

Quant Time: Apr 02 02:54:59 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 02:44:48 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	164699	160.155	ug/l	98
42) 1,2-Dichloroethane	6.080	62	296168	153.146	ug/l	100
43) Isopropyl Acetate	6.336	43	475823	162.494	ug/l	99
44) Trichloroethene	7.117	130	170940	153.599	ug/l	95
45) 1,2-Dichloropropane	7.428	63	179457	153.610	ug/l	98
46) Dibromomethane	7.574	93	135415	150.996	ug/l	98
47) Bromodichloromethane	7.818	83	279881	156.272	ug/l	98
48) Methyl methacrylate	7.690	41	240818	159.281	ug/l	99
49) 1,4-Dioxane	7.659	88	74929	2739.196	ug/l	99
51) 4-Methyl-2-Pentanone	8.574	43	1414771	728.817	ug/l	99
52) Toluene	8.714	92	420015	148.245	ug/l	100
53) t-1,3-Dichloropropene	8.976	75	267671	150.589	ug/l	98
54) cis-1,3-Dichloropropene	8.360	75	287973	171.104	ug/l	98
55) 1,1,2-Trichloroethane	9.147	97	163426	144.752	ug/l	98
56) Ethyl methacrylate	9.116	69	285895	163.624	ug/l	98
57) 1,3-Dichloropropane	9.305	76	288430	146.708	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.238	63	703881	796.340	ug/l	99
59) 2-Hexanone	9.433	43	1053971	733.198	ug/l	100
60) Dibromochloromethane	9.519	129	194342	157.895	ug/l	98
61) 1,2-Dibromoethane	9.610	107	174571	152.745	ug/l	99
64) Tetrachloroethene	9.269	164	138926	147.568	ug/l	98
65) Chlorobenzene	10.080	112	444631	156.049	ug/l	98
66) 1,1,1,2-Tetrachloroethane	10.159	131	156069	157.769	ug/l	99
67) Ethyl Benzene	10.189	91	821130	160.907	ug/l	99
68) m/p-Xylenes	10.299	106	583228	314.231	ug/l	99
69) o-Xylene	10.640	106	285670	156.241	ug/l	100
70) Styrene	10.653	104	480722	159.299	ug/l	99
71) Bromoform	10.799	173	126115	170.743	ug/l #	98
73) Isopropylbenzene	10.957	105	764406	155.475	ug/l	100
74) N-amyl acetate	10.842	43	408934	173.877	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.213	83	249980	144.737	ug/l	100
76) 1,2,3-Trichloropropane	11.238	75	215694m	144.064	ug/l	
77) Bromobenzene	11.195	156	171573	151.033	ug/l	99
78) n-propylbenzene	11.305	91	897084	158.408	ug/l	100
79) 2-Chlorotoluene	11.360	91	537263	148.399	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	620819	152.700	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	78805	178.636	ug/l	93
82) 4-Chlorotoluene	11.451	91	615594	152.505	ug/l	99
83) tert-Butylbenzene	11.713	119	627039	155.783	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	629704	154.317	ug/l	100
85) sec-Butylbenzene	11.890	105	790355	159.869	ug/l	100
86) p-Isopropyltoluene	12.006	119	641521	157.397	ug/l	100
87) 1,3-Dichlorobenzene	11.969	146	314077	151.885	ug/l	100
88) 1,4-Dichlorobenzene	12.037	146	314197	149.876	ug/l	99
89) n-Butylbenzene	12.329	91	604445	171.019	ug/l	100
90) Hexachloroethane	12.536	117	121761	173.829	ug/l	99
91) 1,2-Dichlorobenzene	12.335	146	307031	149.342	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	12.939	75	64317	178.062	ug/l	99
93) 1,2,4-Trichlorobenzene	13.585	180	199365	174.213	ug/l	99
94) Hexachlorobutadiene	13.725	225	77792	157.752	ug/l	99
95) Naphthalene	13.774	128	733300	172.191	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	201989	168.562	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
Data File : VX045530.D
Acq On : 01 Apr 2025 19:02
Operator : JC/MD
Sample : VSTDICC150
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations
APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025

Quant Time: Apr 02 02:54:59 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 02:44:48 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
 Data File : VX045530.D
 Acq On : 01 Apr 2025 19:02
 Operator : JC/MD
 Sample : VSTDIC150
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC150

Manual Integrations APPROVED

Reviewed By :Amit Patel 04/02/2025
 Supervised By :Mahesh Dadoda 04/02/2025

Quant Time: Apr 02 02:54:59 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 02:44:48 2025
 Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
 Data File : VX045532.D
 Acq On : 01 Apr 2025 19:48
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX040225

Quant Time: Apr 02 03:14:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Amit Patel 04/02/2025
 Supervised By :Mahesh Dadoda 04/02/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	96972	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	174855	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	152417	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	67633	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	89251	50.329	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 100.660%			
35) Dibromofluoromethane	5.379	113	61107	49.256	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 98.520%			
50) Toluene-d8	8.647	98	217200	50.160	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 100.320%			
62) 4-Bromofluorobenzene	11.079	95	80376	50.959	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 101.920%			
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.166	85	76070	52.454	ug/l	97
3) Chloromethane	1.301	50	72033	48.349	ug/l	99
4) Vinyl Chloride	1.374	62	66559	48.816	ug/l	99
5) Bromomethane	1.593	94	30379	46.991	ug/l	99
6) Chloroethane	1.672	64	37317	51.586	ug/l	100
7) Trichlorofluoromethane	1.880	101	102379	50.354	ug/l	96
8) Diethyl Ether	2.130	74	33603	49.228	ug/l	99
9) 1,1,2-Trichlorotrifluo...	2.325	101	59007	49.528	ug/l	98
10) Methyl Iodide	2.447	142	72261	48.626	ug/l	99
11) Tert butyl alcohol	2.971	59	58774	246.612	ug/l	99
12) 1,1-Dichloroethene	2.313	96	57888	49.726	ug/l	99
13) Acrolein	2.233	56	78156	238.141	ug/l	99
14) Allyl chloride	2.660	41	111636	50.541	ug/l	99
15) Acrylonitrile	3.062	53	186837	248.351	ug/l	100
16) Acetone	2.380	43	176153	241.904	ug/l	99
17) Carbon Disulfide	2.508	76	149677	52.057	ug/l	100
18) Methyl Acetate	2.703	43	81720	48.742	ug/l	100
19) Methyl tert-butyl Ether	3.111	73	207487	50.950	ug/l	100
20) Methylene Chloride	2.782	84	66875	49.055	ug/l	99
21) trans-1,2-Dichloroethene	3.087	96	59229	49.800	ug/l	98
22) Diisopropyl ether	3.757	45	221449	50.895	ug/l	92
23) Vinyl Acetate	3.715	43	1005423	267.449	ug/l	99
24) 1,1-Dichloroethane	3.605	63	122240	49.677	ug/l	99
25) 2-Butanone	4.550	43	271686	254.882	ug/l	97
26) 2,2-Dichloropropane	4.471	77	82704	52.020	ug/l	100
27) cis-1,2-Dichloroethene	4.483	96	71072	49.049	ug/l	98
28) Bromochloromethane	4.891	49	57557	48.403	ug/l	99
29) Tetrahydrofuran	5.001	42	172405	249.473	ug/l	100
30) Chloroform	5.086	83	125621	49.614	ug/l	98
31) Cyclohexane	5.458	56	107862	49.583	ug/l	97
32) 1,1,1-Trichloroethane	5.379	97	107830	50.330	ug/l	99
36) 1,1-Dichloropropene	5.684	75	83901	50.088	ug/l	99
37) Ethyl Acetate	4.708	43	105095	49.787	ug/l	98
38) Carbon Tetrachloride	5.672	117	91723	50.667	ug/l	100
39) Methylcyclohexane	7.373	83	102919	50.124	ug/l	99
40) Benzene	6.031	78	251758	49.212	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
 Data File : VX045532.D
 Acq On : 01 Apr 2025 19:48
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX040225

Quant Time: Apr 02 03:14:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Amit Patel 04/02/2025
 Supervised By :Mahesh Dadoda 04/02/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.916	41	59140	53.098	ug/l	99
42) 1,2-Dichloroethane	6.080	62	103270	48.875	ug/l	100
43) Isopropyl Acetate	6.336	43	164894	51.539	ug/l	100
44) Trichloroethene	7.123	130	57931	47.643	ug/l	92
45) 1,2-Dichloropropane	7.421	63	63210	49.521	ug/l	95
46) Dibromomethane	7.574	93	47564	48.542	ug/l	97
47) Bromodichloromethane	7.818	83	96418	49.273	ug/l	96
48) Methyl methacrylate	7.690	41	85084	51.507	ug/l	99
49) 1,4-Dioxane	7.659	88	28552	955.327	ug/l	99
51) 4-Methyl-2-Pentanone	8.568	43	540822	254.994	ug/l	100
52) Toluene	8.714	92	151664	48.994	ug/l	100
53) t-1,3-Dichloropropene	8.976	75	87744	46.783	ug/l	98
54) cis-1,3-Dichloropropene	8.360	75	99074	53.878	ug/l	98
55) 1,1,2-Trichloroethane	9.147	97	60452	49.007	ug/l	97
56) Ethyl methacrylate	9.116	69	102525	53.705	ug/l	100
57) 1,3-Dichloropropane	9.305	76	107121	49.869	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.238	63	251285	260.201	ug/l	99
59) 2-Hexanone	9.427	43	395336	251.710	ug/l	100
60) Dibromochloromethane	9.519	129	68562	50.983	ug/l	99
61) 1,2-Dibromoethane	9.604	107	63301	50.693	ug/l	100
64) Tetrachloroethene	9.269	164	50834	47.239	ug/l	99
65) Chlorobenzene	10.073	112	163518	50.207	ug/l	98
66) 1,1,1,2-Tetrachloroethane	10.159	131	57096	50.495	ug/l	99
67) Ethyl Benzene	10.189	91	300747	51.558	ug/l	100
68) m/p-Xylenes	10.299	106	217943	102.728	ug/l	99
69) o-Xylene	10.640	106	107947	51.650	ug/l	100
70) Styrene	10.653	104	181524	52.624	ug/l	100
71) Bromoform	10.799	173	44489	52.694	ug/l #	98
73) Isopropylbenzene	10.957	105	286010	52.794	ug/l	100
74) N-amyl acetate	10.842	43	140941	54.387	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.207	83	92505	48.608	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	82060m	49.741	ug/l	
77) Bromobenzene	11.195	156	63244	50.525	ug/l	98
78) n-propylbenzene	11.299	91	336753	53.967	ug/l	100
79) 2-Chlorotoluene	11.360	91	203677	51.057	ug/l	99
80) 1,3,5-Trimethylbenzene	11.445	105	239822	53.534	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	25511	52.482	ug/l	98
82) 4-Chlorotoluene	11.451	91	233159	52.422	ug/l	100
83) tert-Butylbenzene	11.713	119	236276	53.274	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	238257	52.990	ug/l	100
85) sec-Butylbenzene	11.890	105	293633	53.903	ug/l	100
86) p-Isopropyltoluene	12.006	119	244031	54.337	ug/l	99
87) 1,3-Dichlorobenzene	11.963	146	115113	50.521	ug/l	100
88) 1,4-Dichlorobenzene	12.036	146	113458	49.117	ug/l	97
89) n-Butylbenzene	12.329	91	214931	55.189	ug/l	99
90) Hexachloroethane	12.536	117	41729	54.066	ug/l	98
91) 1,2-Dichlorobenzene	12.329	146	113126	49.938	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	12.939	75	21184	53.226	ug/l	100
93) 1,2,4-Trichlorobenzene	13.585	180	65488	51.935	ug/l	98
94) Hexachlorobutadiene	13.725	225	27147	49.961	ug/l	99
95) Naphthalene	13.774	128	247600	52.765	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	66542	50.396	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
Data File : VX045532.D
Acq On : 01 Apr 2025 19:48
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Manual Integrations
APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025

Quant Time: Apr 02 03:14:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


```
Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\  
Data File : VX045532.D  
Acq On    : 01 Apr 2025   19:48  
Operator  : JC/MD  
Sample    : VSTDICV050  
Misc      : 5.0mL/MSVOA_X/WATER  
ALS Vial  : 10   Sample Multiplier: 1
```

Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Manual Integrations APPROVED

Reviewed By :Amit Patel 04/02/2025
Supervised By :Mahesh Dadoda 04/02/2025

Quant Time: Apr 02 03:14:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
 Data File : VX045532.D
 Acq On : 01 Apr 2025 19:48
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICSVVX040225

Quant Time: Apr 02 03:14:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	101	0.00
2 T	Dichlorodifluoromethane	0.748	0.784	-4.8	97	0.00
3 P	Chloromethane	0.768	0.743	3.3	96	0.00
4 C	Vinyl Chloride	0.703	0.686	2.4#	97	0.00
5 T	Bromomethane	0.333	0.313	6.0	96	0.00
6 T	Chloroethane	0.373	0.385	-3.2	98	0.00
7 T	Trichlorofluoromethane	1.048	1.056	-0.8	99	0.00
8 T	Diethyl Ether	0.352	0.347	1.4	99	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.614	0.608	1.0	99	0.00
10 T	Methyl Iodide	0.766	0.745	2.7	95	0.00
11 T	Tert butyl alcohol	0.123	0.121	1.6	97	0.00
12 CM	1,1-Dichloroethene	0.600	0.597	0.5#	100	0.00
13 T	Acrolein	0.169	0.161	4.7	98	0.00
14 T	Allyl chloride	1.139	1.151	-1.1	100	0.00
15 T	Acrylonitrile	0.388	0.385	0.8	96	0.00
16 T	Acetone	0.375	0.363	3.2	98	0.00
17 T	Carbon Disulfide	1.483	1.544	-4.1	101	0.00
18 T	Methyl Acetate	0.864	0.843	2.4	100	0.00
19 T	Methyl tert-butyl Ether	2.100	2.140	-1.9	102	0.00
20 T	Methylene Chloride	0.703	0.690	1.8	99	0.00
21 T	trans-1,2-Dichloroethene	0.613	0.611	0.3	99	0.00
22 T	Diisopropyl ether	2.243	2.284	-1.8	101	0.00
23 T	Vinyl Acetate	1.938	2.074	-7.0	101	0.00
24 P	1,1-Dichloroethane	1.269	1.261	0.6	101	0.00
25 T	2-Butanone	0.550	0.560	-1.8	97	0.00
26 T	2,2-Dichloropropane	0.820	0.853	-4.0	102	0.00
27 T	cis-1,2-Dichloroethene	0.747	0.733	1.9	99	0.00
28 T	Bromochloromethane	0.613	0.594	3.1	101	0.00
29 T	Tetrahydrofuran	0.356	0.356	0.0	97	0.00
30 C	Chloroform	1.306	1.295	0.8#	100	0.00
31 T	Cyclohexane	1.122	1.112	0.9	100	0.00
32 T	1,1,1-Trichloroethane	1.105	1.112	-0.6	102	0.00
33 S	1,2-Dichloroethane-d4	0.914	0.920	-0.7	107	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	104	0.00
35 S	Dibromofluoromethane	0.355	0.349	1.7	105	0.00
36 T	1,1-Dichloropropene	0.479	0.480	-0.2	103	0.00
37 T	Ethyl Acetate	0.604	0.601	0.5	100	0.00
38 T	Carbon Tetrachloride	0.518	0.525	-1.4	102	0.00
39 T	Methylcyclohexane	0.587	0.589	-0.3	98	0.00
40 TM	Benzene	1.463	1.440	1.6	101	0.00
41 T	Methacrylonitrile	0.318	0.338	-6.3	100	0.00
42 TM	1,2-Dichloroethane	0.604	0.591	2.2	99	0.00
43 T	Isopropyl Acetate	0.915	0.943	-3.1	100	0.00
44 TM	Trichloroethene	0.348	0.331	4.9	98	0.00
45 C	1,2-Dichloropropane	0.365	0.361	1.1#	100	0.00
46 T	Dibromomethane	0.280	0.272	2.9	97	0.00
47 T	Bromodichloromethane	0.560	0.551	1.6	100	0.00
48 T	Methyl methacrylate	0.472	0.487	-3.2	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
 Data File : VX045532.D
 Acq On : 01 Apr 2025 19:48
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICSVVX040225

Quant Time: Apr 02 03:14:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.008	11.1	92	0.00
50 S	Toluene-d8	1.238	1.242	-0.3	106	0.00
51 T	4-Methyl-2-Pentanone	0.606	0.619	-2.1	97	0.00
52 CM	Toluene	0.885	0.867	2.0#	99	0.00
53 T	t-1,3-Dichloropropene	0.467	0.502	-7.5	103	0.00
54 T	cis-1,3-Dichloropropene	0.526	0.567	-7.8	104	0.00
55 T	1,1,2-Trichloroethane	0.353	0.346	2.0	100	0.00
56 T	Ethyl methacrylate	0.546	0.586	-7.3	102	0.00
57 T	1,3-Dichloropropane	0.614	0.613	0.2	101	0.00
58 T	2-Chloroethyl Vinyl ether	0.276	0.287	-4.0	100	0.00
59 T	2-Hexanone	0.449	0.452	-0.7	96	0.00
60 T	Dibromochloromethane	0.385	0.392	-1.8	99	0.00
61 T	1,2-Dibromoethane	0.357	0.362	-1.4	101	0.00
62 S	4-Bromofluorobenzene	0.451	0.460	-2.0	105	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	101	0.00
64 T	Tetrachloroethene	0.353	0.334	5.4	97	0.00
65 PM	Chlorobenzene	1.068	1.073	-0.5	100	0.00
66 T	1,1,1,2-Tetrachloroethane	0.371	0.375	-1.1	102	0.00
67 C	Ethyl Benzene	1.914	1.973	-3.1#	100	0.00
68 T	m/p-Xylenes	0.696	0.715	-2.7	100	0.00
69 T	o-Xylene	0.686	0.708	-3.2	100	0.00
70 T	Styrene	1.132	1.191	-5.2	99	0.00
71 P	Bromoform	0.277	0.292	-5.4	100	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	0.00
73 T	Isopropylbenzene	4.005	4.229	-5.6	99	0.00
74 T	N-amyl acetate	1.916	2.084	-8.8	100	0.00
75 P	1,1,2,2-Tetrachloroethane	1.407	1.368	2.8	97	0.00
76 T	1,2,3-Trichloropropane	1.220	1.213	0.6	97	0.00
77 T	Bromobenzene	0.925	0.935	-1.1	98	0.00
78 T	n-propylbenzene	4.613	4.979	-7.9	100	0.00
79 T	2-Chlorotoluene	2.949	3.012	-2.1	100	0.00
80 T	1,3,5-Trimethylbenzene	3.312	3.546	-7.1	99	0.00
81 T	trans-1,4-Dichloro-2-butene	0.359	0.377	-5.0	100	0.00
82 T	4-Chlorotoluene	3.288	3.447	-4.8	98	0.00
83 T	tert-Butylbenzene	3.279	3.494	-6.6	100	0.00
84 T	1,2,4-Trimethylbenzene	3.324	3.523	-6.0	98	0.00
85 T	sec-Butylbenzene	4.027	4.342	-7.8	99	0.00
86 T	p-Isopropyltoluene	3.320	3.608	-8.7	101	0.00
87 T	1,3-Dichlorobenzene	1.684	1.702	-1.1	98	0.00
88 T	1,4-Dichlorobenzene	1.708	1.678	1.8	97	0.00
89 T	n-Butylbenzene	2.879	3.178	-10.4	99	0.00
90 T	Hexachloroethane	0.571	0.617	-8.1	102	0.00
91 T	1,2-Dichlorobenzene	1.675	1.673	0.1	98	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.294	0.313	-6.5	98	0.00
93 T	1,2,4-Trichlorobenzene	0.932	0.968	-3.9	101	0.00
94 T	Hexachlorobutadiene	0.402	0.401	0.2	98	0.00
95 T	Naphthalene	3.469	3.661	-5.5	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
Data File : VX045532.D
Acq On : 01 Apr 2025 19:48
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Quant Time: Apr 02 03:14:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.976	0.984	-0.8	97	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX040225\
 Data File : VX045532.D
 Acq On : 01 Apr 2025 19:48
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX040225

Quant Time: Apr 02 03:14:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	101	0.00
2 T	Dichlorodifluoromethane	50.000	52.454	-4.9	97	0.00
3 P	Chloromethane	50.000	48.349	3.3	96	0.00
4 C	Vinyl Chloride	50.000	48.816	2.4#	97	0.00
5 T	Bromomethane	50.000	46.991	6.0	96	0.00
6 T	Chloroethane	50.000	51.586	-3.2	98	0.00
7 T	Trichlorofluoromethane	50.000	50.354	-0.7	99	0.00
8 T	Diethyl Ether	50.000	49.228	1.5	99	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	49.528	0.9	99	0.00
10 T	Methyl Iodide	50.000	48.626	2.7	95	0.00
11 T	Tert butyl alcohol	250.000	246.612	1.4	97	0.00
12 CM	1,1-Dichloroethene	50.000	49.726	0.5#	100	0.00
13 T	Acrolein	250.000	238.141	4.7	98	0.00
14 T	Allyl chloride	50.000	50.541	-1.1	100	0.00
15 T	Acrylonitrile	250.000	248.351	0.7	96	0.00
16 T	Acetone	250.000	241.904	3.2	98	0.00
17 T	Carbon Disulfide	50.000	52.057	-4.1	101	0.00
18 T	Methyl Acetate	50.000	48.742	2.5	100	0.00
19 T	Methyl tert-butyl Ether	50.000	50.950	-1.9	102	0.00
20 T	Methylene Chloride	50.000	49.055	1.9	99	0.00
21 T	trans-1,2-Dichloroethene	50.000	49.800	0.4	99	0.00
22 T	Diisopropyl ether	50.000	50.895	-1.8	101	0.00
23 T	Vinyl Acetate	250.000	267.449	-7.0	101	0.00
24 P	1,1-Dichloroethane	50.000	49.677	0.6	101	0.00
25 T	2-Butanone	250.000	254.882	-2.0	97	0.00
26 T	2,2-Dichloropropane	50.000	52.020	-4.0	102	0.00
27 T	cis-1,2-Dichloroethene	50.000	49.049	1.9	99	0.00
28 T	Bromochloromethane	50.000	48.403	3.2	101	0.00
29 T	Tetrahydrofuran	250.000	249.473	0.2	97	0.00
30 C	Chloroform	50.000	49.614	0.8#	100	0.00
31 T	Cyclohexane	50.000	49.583	0.8	100	0.00
32 T	1,1,1-Trichloroethane	50.000	50.330	-0.7	102	0.00
33 S	1,2-Dichloroethane-d4	50.000	50.329	-0.7	107	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	104	0.00
35 S	Dibromofluoromethane	50.000	49.256	1.5	105	0.00
36 T	1,1-Dichloropropene	50.000	50.088	-0.2	103	0.00
37 T	Ethyl Acetate	50.000	49.787	0.4	100	0.00
38 T	Carbon Tetrachloride	50.000	50.667	-1.3	102	0.00
39 T	Methylcyclohexane	50.000	50.124	-0.2	98	0.00
40 TM	Benzene	50.000	49.212	1.6	101	0.00
41 T	Methacrylonitrile	50.000	53.098	-6.2	100	0.00
42 TM	1,2-Dichloroethane	50.000	48.875	2.3	99	0.00
43 T	Isopropyl Acetate	50.000	51.539	-3.1	100	0.00
44 TM	Trichloroethene	50.000	47.643	4.7	98	0.00
45 C	1,2-Dichloropropane	50.000	49.521	1.0#	100	0.00
46 T	Dibromomethane	50.000	48.542	2.9	97	0.00
47 T	Bromodichloromethane	50.000	49.273	1.5	100	0.00
48 T	Methyl methacrylate	50.000	51.507	-3.0	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX040225\
 Data File : VX045532.D
 Acq On : 01 Apr 2025 19:48
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICSVVX040225

Quant Time: Apr 02 03:14:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	955.327	4.5	92	0.00
50 S	Toluene-d8	50.000	50.160	-0.3	106	0.00
51 T	4-Methyl-2-Pentanone	250.000	254.994	-2.0	97	0.00
52 CM	Toluene	50.000	48.994	2.0#	99	0.00
53 T	t-1,3-Dichloropropene	50.000	46.783	6.4	103	0.00
54 T	cis-1,3-Dichloropropene	50.000	53.878	-7.8	104	0.00
55 T	1,1,2-Trichloroethane	50.000	49.007	2.0	100	0.00
56 T	Ethyl methacrylate	50.000	53.705	-7.4	102	0.00
57 T	1,3-Dichloropropane	50.000	49.869	0.3	101	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	260.201	-4.1	100	0.00
59 T	2-Hexanone	250.000	251.710	-0.7	96	0.00
60 T	Dibromochloromethane	50.000	50.983	-2.0	99	0.00
61 T	1,2-Dibromoethane	50.000	50.693	-1.4	101	0.00
62 S	4-Bromofluorobenzene	50.000	50.959	-1.9	105	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	101	0.00
64 T	Tetrachloroethene	50.000	47.239	5.5	97	0.00
65 PM	Chlorobenzene	50.000	50.207	-0.4	100	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	50.495	-1.0	102	0.00
67 C	Ethyl Benzene	50.000	51.558	-3.1#	100	0.00
68 T	m/p-Xylenes	100.000	102.728	-2.7	100	0.00
69 T	o-Xylene	50.000	51.650	-3.3	100	0.00
70 T	Styrene	50.000	52.624	-5.2	99	0.00
71 P	Bromoform	50.000	52.694	-5.4	100	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	98	0.00
73 T	Isopropylbenzene	50.000	52.794	-5.6	99	0.00
74 T	N-ethyl acetate	50.000	54.387	-8.8	100	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	48.608	2.8	97	0.00
76 T	1,2,3-Trichloropropane	50.000	49.741	0.5	97	0.00
77 T	Bromobenzene	50.000	50.525	-1.0	98	0.00
78 T	n-propylbenzene	50.000	53.967	-7.9	100	0.00
79 T	2-Chlorotoluene	50.000	51.057	-2.1	100	0.00
80 T	1,3,5-Trimethylbenzene	50.000	53.534	-7.1	99	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	52.482	-5.0	100	0.00
82 T	4-Chlorotoluene	50.000	52.422	-4.8	98	0.00
83 T	tert-Butylbenzene	50.000	53.274	-6.5	100	0.00
84 T	1,2,4-Trimethylbenzene	50.000	52.990	-6.0	98	0.00
85 T	sec-Butylbenzene	50.000	53.903	-7.8	99	0.00
86 T	p-Isopropyltoluene	50.000	54.337	-8.7	101	0.00
87 T	1,3-Dichlorobenzene	50.000	50.521	-1.0	98	0.00
88 T	1,4-Dichlorobenzene	50.000	49.117	1.8	97	0.00
89 T	n-Butylbenzene	50.000	55.189	-10.4	99	0.00
90 T	Hexachloroethane	50.000	54.066	-8.1	102	0.00
91 T	1,2-Dichlorobenzene	50.000	49.938	0.1	98	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	53.226	-6.5	98	0.00
93 T	1,2,4-Trichlorobenzene	50.000	51.935	-3.9	101	0.00
94 T	Hexachlorobutadiene	50.000	49.961	0.1	98	0.00
95 T	Naphthalene	50.000	52.765	-5.5	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
Data File : VX045532.D
Acq On : 01 Apr 2025 19:48
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX040225

Quant Time: Apr 02 03:14:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	50.396	-0.8	97	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: ENTA05

Lab Code: CHEM Case No.: Q1822 SAS No.: Q1822 SDG No.: Q1822

Instrument ID: MSVOA_X Calibration Date/Time: 04/16/2025 09:28

Lab File ID: VX045797.D Init. Calib. Date(s): 04/01/2025 04/01/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 17:06 19:02

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Methyl tert-butyl Ether	2.100	2.349		11.86	20
Carbon Tetrachloride	0.518	0.503		-2.9	20
Chloroform	1.306	1.328		1.68	20
1,1,1-Trichloroethane	1.105	1.092		-1.18	20
Benzene	1.463	1.414		-3.35	20
Toluene	0.885	0.862		-2.6	20
Tetrachloroethene	0.353	0.315		-10.77	20
Ethyl Benzene	1.914	1.888		-1.36	20
m/p-Xylenes	0.696	0.691		-0.72	20
o-Xylene	0.686	0.694		1.17	20
1,2-Dichloroethane-d4	0.914	0.994		8.75	20
Dibromofluoromethane	0.355	0.371		4.51	20
Toluene-d8	1.238	1.235		-0.24	20
4-Bromofluorobenzene	0.451	0.505		11.97	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041625\
 Data File : VX045797.D
 Acq On : 16 Apr 2025 09:28
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050

Manual Integrations
 APPROVED

Reviewed By : John Carlone 04/17/2025
 Supervised By : Mahesh Dadoda 04/17/2025

Quant Time: Apr 17 01:34:51 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	85794	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	6.751	114	152447	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	137129	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	66178	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	85258	54.341	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 108.680%			
35) Dibromofluoromethane	5.379	113	56514	52.250	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 104.500%			
50) Toluene-d8	8.647	98	188345	49.889	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 99.780%			
62) 4-Bromofluorobenzene	11.079	95	77007	56.000	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 112.000%			
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.166	85	56864	44.319	ug/l	99
3) Chloromethane	1.301	50	58067	44.053	ug/l	100
4) Vinyl Chloride	1.374	62	51042	42.313	ug/l	100
5) Bromomethane	1.593	94	25102	43.887	ug/l	97
6) Chloroethane	1.666	64	31371	49.017	ug/l	97
7) Trichlorofluoromethane	1.874	101	80612	44.814	ug/l	99
8) Diethyl Ether	2.130	74	31315	51.853	ug/l	98
9) 1,1,2-Trichlorotrifluo...	2.319	101	49923	47.363	ug/l	97
10) Methyl Iodide	2.441	142	62671	47.668	ug/l	98
11) Tert butyl alcohol	2.977	59	58600	277.917	ug/l	100
12) 1,1-Dichloroethene	2.306	96	45913	44.578	ug/l	98
13) Acrolein	2.233	56	81349	280.165	ug/l	98
14) Allyl chloride	2.654	41	99816	51.077	ug/l	99
15) Acrylonitrile	3.062	53	178859	268.722	ug/l	99
16) Acetone	2.386	43	168189	261.060	ug/l	99
17) Carbon Disulfide	2.502	76	98988	38.913	ug/l	97
18) Methyl Acetate	2.703	43	92549	62.394	ug/l	98
19) Methyl tert-butyl Ether	3.111	73	201566	55.945	ug/l	99
20) Methylene Chloride	2.782	84	60009	49.754	ug/l	95
21) trans-1,2-Dichloroethene	3.081	96	47526	45.167	ug/l	96
22) Diisopropyl ether	3.757	45	207871	53.999	ug/l	97
23) Vinyl Acetate	3.715	43	915079	275.132	ug/l	100
24) 1,1-Dichloroethane	3.599	63	108165	49.684	ug/l	99
25) 2-Butanone	4.550	43	258726	274.348	ug/l	99
26) 2,2-Dichloropropane	4.465	77	82001	58.298	ug/l	99
27) cis-1,2-Dichloroethene	4.483	96	63840	49.799	ug/l	98
28) Bromochloromethane	4.891	49	52003	49.430	ug/l	99
29) Tetrahydrofuran	5.001	42	162217	265.314	ug/l	100
30) Chloroform	5.086	83	113975	50.879	ug/l	99
31) Cyclohexane	5.458	56	82171	42.695	ug/l	98
32) 1,1,1-Trichloroethane	5.373	97	93712	49.439	ug/l	98
36) 1,1-Dichloropropene	5.684	75	66526	45.553	ug/l	99
37) Ethyl Acetate	4.708	43	95198	51.727	ug/l	99
38) Carbon Tetrachloride	5.666	117	76701	48.596	ug/l	98
39) Methylcyclohexane	7.373	83	80833	45.154	ug/l	98
40) Benzene	6.031	78	215607	48.340	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041625\
 Data File : VX045797.D
 Acq On : 16 Apr 2025 09:28
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050

Manual Integrations
 APPROVED

Reviewed By : John Carlone 04/17/2025
 Supervised By : Mahesh Dadoda 04/17/2025

Quant Time: Apr 17 01:34:51 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.916	41	55377	57.028	ug/l	98
42) 1,2-Dichloroethane	6.080	62	99725	54.134	ug/l	99
43) Isopropyl Acetate	6.336	43	158110	56.683	ug/l	100
44) Trichloroethene	7.117	130	50885	47.999	ug/l	93
45) 1,2-Dichloropropane	7.421	63	57978	52.098	ug/l	94
46) Dibromomethane	7.574	93	45348	53.083	ug/l	99
47) Bromodichloromethane	7.818	83	91931	53.886	ug/l	95
48) Methyl methacrylate	7.690	41	82951	57.597	ug/l	98
49) 1,4-Dioxane	7.659	88	27700	1063.052	ug/l	98
51) 4-Methyl-2-Pentanone	8.568	43	523522	283.119	ug/l	100
52) Toluene	8.714	92	131394	48.685	ug/l	99
53) t-1,3-Dichloropropene	8.976	75	85387	51.952	ug/l	99
54) cis-1,3-Dichloropropene	8.360	75	92710	57.828	ug/l	97
55) 1,1,2-Trichloroethane	9.147	97	57736	53.685	ug/l	96
56) Ethyl methacrylate	9.116	69	97572	58.623	ug/l	98
57) 1,3-Dichloropropane	9.305	76	100890	53.872	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.238	63	240911	286.126	ug/l	100
59) 2-Hexanone	9.427	43	383752	280.249	ug/l	100
60) Dibromochloromethane	9.519	129	67592	57.650	ug/l	97
61) 1,2-Dibromoethane	9.604	107	59717	54.852	ug/l	99
64) Tetrachloroethene	9.269	164	43149	44.568	ug/l	97
65) Chlorobenzene	10.073	112	150112	51.229	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.159	131	54651	53.721	ug/l	98
67) Ethyl Benzene	10.189	91	258942	49.341	ug/l	99
68) m/p-Xylenes	10.299	106	189565	99.313	ug/l	99
69) o-Xylene	10.640	106	95182	50.620	ug/l	97
70) Styrene	10.653	104	168709	54.362	ug/l	99
71) Bromoform	10.799	173	43602	57.401	ug/l #	96
73) Isopropylbenzene	10.957	105	248367	46.854	ug/l	99
74) N-amyl acetate	10.841	43	135817	53.562	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.207	83	91338	49.050	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	80513m	49.877	ug/l	
77) Bromobenzene	11.195	156	59517	48.593	ug/l	97
78) n-propylbenzene	11.299	91	289136	47.354	ug/l	100
79) 2-Chlorotoluene	11.360	91	186070	47.669	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	214243	48.876	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	25224	53.033	ug/l	99
82) 4-Chlorotoluene	11.451	91	216767	49.808	ug/l	99
83) tert-Butylbenzene	11.713	119	209176	48.200	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	220858	50.200	ug/l	99
85) sec-Butylbenzene	11.890	105	259172	48.623	ug/l	100
86) p-Isopropyltoluene	12.006	119	216106	49.177	ug/l	100
87) 1,3-Dichlorobenzene	11.969	146	108860	48.827	ug/l	99
88) 1,4-Dichlorobenzene	12.036	146	110841	49.039	ug/l	98
89) n-Butylbenzene	12.329	91	191975	50.378	ug/l	100
90) Hexachloroethane	12.536	117	38783	51.353	ug/l	99
91) 1,2-Dichlorobenzene	12.335	146	111786	50.431	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	22270	57.185	ug/l	97
93) 1,2,4-Trichlorobenzene	13.585	180	67177	54.446	ug/l	96
94) Hexachlorobutadiene	13.725	225	25870	48.658	ug/l	98
95) Naphthalene	13.774	128	249200	54.274	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	67128	51.958	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041625\
Data File : VX045797.D
Acq On : 16 Apr 2025 09:28
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/17/2025
Supervised By :Mahesh Dadoda 04/17/2025

Quant Time: Apr 17 01:34:51 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


```
Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041625\
Data File : VX045797.D
Acq On    : 16 Apr 2025   09:28
Operator  : JC/MD
Sample    : VSTDCCC050
Misc      : 5.0mL/MSVOA_X/WATER
ALS Vial  : 2      Sample Multiplier: 1
```

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations

Reviewed By :John Carlone 04/17/2025
Supervised By :Mahesh Dadoda 04/17/2025

Quant Time: Apr 17 01:34:51 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041625\
 Data File : VX045797.D
 Acq On : 16 Apr 2025 09:28
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 17 01:34:51 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	90	-0.01
2 T	Dichlorodifluoromethane	0.748	0.663	11.4	72	0.00
3 P	Chloromethane	0.768	0.677	11.8	77	0.00
4 C	Vinyl Chloride	0.703	0.595	15.4#	74	0.00
5 T	Bromomethane	0.333	0.293	12.0	79	0.00
6 T	Chloroethane	0.373	0.366	1.9	82	0.00
7 T	Trichlorofluoromethane	1.048	0.940	10.3	78	0.00
8 T	Diethyl Ether	0.352	0.365	-3.7	92	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.614	0.582	5.2	84	0.00
10 T	Methyl Iodide	0.766	0.730	4.7	82	0.00
11 T	Tert butyl alcohol	0.123	0.137	-11.4	97	0.00
12 CM	1,1-Dichloroethene	0.600	0.535	10.8#	80	0.00
13 T	Acrolein	0.169	0.190	-12.4	102	0.00
14 T	Allyl chloride	1.139	1.163	-2.1	89	0.00
15 T	Acrylonitrile	0.388	0.417	-7.5	92	0.00
16 T	Acetone	0.375	0.392	-4.5	93	0.00
17 T	Carbon Disulfide	1.483	1.154	22.2	67	0.00
18 T	Methyl Acetate	0.864	1.079	-24.9	113	0.00
19 T	Methyl tert-butyl Ether	2.100	2.349	-11.9	99	0.00
20 T	Methylene Chloride	0.703	0.699	0.6	88	0.00
21 T	trans-1,2-Dichloroethene	0.613	0.554	9.6	80	0.00
22 T	Diisopropyl ether	2.243	2.423	-8.0	95	0.00
23 T	Vinyl Acetate	1.938	2.133	-10.1	92	0.00
24 P	1,1-Dichloroethane	1.269	1.261	0.6	89	0.00
25 T	2-Butanone	0.550	0.603	-9.6	92	0.00
26 T	2,2-Dichloropropane	0.820	0.956	-16.6	101	0.00
27 T	cis-1,2-Dichloroethene	0.747	0.744	0.4	89	0.00
28 T	Bromochloromethane	0.613	0.606	1.1	91	0.00
29 T	Tetrahydrofuran	0.356	0.378	-6.2	91	0.00
30 C	Chloroform	1.306	1.328	-1.7#	91	0.00
31 T	Cyclohexane	1.122	0.958	14.6	77	0.00
32 T	1,1,1-Trichloroethane	1.105	1.092	1.2	88	0.00
33 S	1,2-Dichloroethane-d4	0.914	0.994	-8.8	103	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	91	0.00
35 S	Dibromofluoromethane	0.355	0.371	-4.5	97	0.00
36 T	1,1-Dichloropropene	0.479	0.436	9.0	82	0.00
37 T	Ethyl Acetate	0.604	0.624	-3.3	91	0.00
38 T	Carbon Tetrachloride	0.518	0.503	2.9	85	0.00
39 T	Methylcyclohexane	0.587	0.530	9.7	77	0.00
40 TM	Benzene	1.463	1.414	3.3	87	0.00
41 T	Methacrylonitrile	0.318	0.363	-14.2	94	0.00
42 TM	1,2-Dichloroethane	0.604	0.654	-8.3	95	0.00
43 T	Isopropyl Acetate	0.915	1.037	-13.3	96	0.00
44 TM	Trichloroethene	0.348	0.334	4.0	86	0.00
45 C	1,2-Dichloropropane	0.365	0.380	-4.1#	92	0.00
46 T	Dibromomethane	0.280	0.297	-6.1	92	0.00
47 T	Bromodichloromethane	0.560	0.603	-7.7	96	0.00
48 T	Methyl methacrylate	0.472	0.544	-15.3	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041625\
 Data File : VX045797.D
 Acq On : 16 Apr 2025 09:28
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 17 01:34:51 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.009	0.0	89	0.00
50 S	Toluene-d8	1.238	1.235	0.2	92	0.00
51 T	4-Methyl-2-Pentanone	0.606	0.687	-13.4	94	0.00
52 CM	Toluene	0.885	0.862	2.6#	86	0.00
53 T	t-1,3-Dichloropropene	0.467	0.560	-19.9	100	0.00
54 T	cis-1,3-Dichloropropene	0.526	0.608	-15.6	97	0.00
55 T	1,1,2-Trichloroethane	0.353	0.379	-7.4	96	0.00
56 T	Ethyl methacrylate	0.546	0.640	-17.2	97	0.00
57 T	1,3-Dichloropropane	0.614	0.662	-7.8	95	0.00
58 T	2-Chloroethyl Vinyl ether	0.276	0.316	-14.5	96	0.00
59 T	2-Hexanone	0.449	0.503	-12.0	93	0.00
60 T	Dibromochloromethane	0.385	0.443	-15.1	98	0.00
61 T	1,2-Dibromoethane	0.357	0.392	-9.8	95	0.00
62 S	4-Bromofluorobenzene	0.451	0.505	-12.0	101	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	91	0.00
64 T	Tetrachloroethene	0.353	0.315	10.8	83	0.00
65 PM	Chlorobenzene	1.068	1.095	-2.5	92	0.00
66 T	1,1,1,2-Tetrachloroethane	0.371	0.399	-7.5	98	0.00
67 C	Ethyl Benzene	1.914	1.888	1.4#	86	0.00
68 T	m/p-Xylenes	0.696	0.691	0.7	87	0.00
69 T	o-Xylene	0.686	0.694	-1.2	88	0.00
70 T	Styrene	1.132	1.230	-8.7	92	0.00
71 P	Bromoform	0.277	0.318	-14.8	98	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
73 T	Isopropylbenzene	4.005	3.753	6.3	86	0.00
74 T	N-amyl acetate	1.916	2.052	-7.1	96	0.00
75 P	1,1,2,2-Tetrachloroethane	1.407	1.380	1.9	96	0.00
76 T	1,2,3-Trichloropropane	1.220	1.217	0.2	96	0.00
77 T	Bromobenzene	0.925	0.899	2.8	92	0.00
78 T	n-propylbenzene	4.613	4.369	5.3	86	0.00
79 T	2-Chlorotoluene	2.949	2.812	4.6	91	0.00
80 T	1,3,5-Trimethylbenzene	3.312	3.237	2.3	89	0.00
81 T	trans-1,4-Dichloro-2-butene	0.359	0.381	-6.1	99	0.00
82 T	4-Chlorotoluene	3.288	3.276	0.4	91	0.00
83 T	tert-Butylbenzene	3.279	3.161	3.6	89	0.00
84 T	1,2,4-Trimethylbenzene	3.324	3.337	-0.4	91	0.00
85 T	sec-Butylbenzene	4.027	3.916	2.8	88	0.00
86 T	p-Isopropyltoluene	3.320	3.266	1.6	89	0.00
87 T	1,3-Dichlorobenzene	1.684	1.645	2.3	93	0.00
88 T	1,4-Dichlorobenzene	1.708	1.675	1.9	95	0.00
89 T	n-Butylbenzene	2.879	2.901	-0.8	88	0.00
90 T	Hexachloroethane	0.571	0.586	-2.6	94	0.00
91 T	1,2-Dichlorobenzene	1.675	1.689	-0.8	97	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.294	0.337	-14.6	103	0.00
93 T	1,2,4-Trichlorobenzene	0.932	1.015	-8.9	103	0.00
94 T	Hexachlorobutadiene	0.402	0.391	2.7	93	0.00
95 T	Naphthalene	3.469	3.766	-8.6	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041625\
Data File : VX045797.D
Acq On : 16 Apr 2025 09:28
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Apr 17 01:34:51 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.976	1.014	-3.9	98	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041625\
 Data File : VX045797.D
 Acq On : 16 Apr 2025 09:28
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 17 01:34:51 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	90	-0.01
2 T	Dichlorodifluoromethane	0.748	0.663	11.4	72	0.00
3 P	Chloromethane	0.768	0.677	11.8	77	0.00
4 C	Vinyl Chloride	0.703	0.595	15.4#	74	0.00
5 T	Bromomethane	0.333	0.293	12.0	79	0.00
6 T	Chloroethane	0.373	0.366	1.9	82	0.00
7 T	Trichlorofluoromethane	1.048	0.940	10.3	78	0.00
8 T	Diethyl Ether	0.352	0.365	-3.7	92	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.614	0.582	5.2	84	0.00
10 T	Methyl Iodide	0.766	0.730	4.7	82	0.00
11 T	Tert butyl alcohol	0.123	0.137	-11.4	97	0.00
12 CM	1,1-Dichloroethene	0.600	0.535	10.8#	80	0.00
13 T	Acrolein	0.169	0.190	-12.4	102	0.00
14 T	Allyl chloride	1.139	1.163	-2.1	89	0.00
15 T	Acrylonitrile	0.388	0.417	-7.5	92	0.00
16 T	Acetone	0.375	0.392	-4.5	93	0.00
17 T	Carbon Disulfide	1.483	1.154	22.2	67	0.00
18 T	Methyl Acetate	0.864	1.079	-24.9	113	0.00
19 T	Methyl tert-butyl Ether	2.100	2.349	-11.9	99	0.00
20 T	Methylene Chloride	0.703	0.699	0.6	88	0.00
21 T	trans-1,2-Dichloroethene	0.613	0.554	9.6	80	0.00
22 T	Diisopropyl ether	2.243	2.423	-8.0	95	0.00
23 T	Vinyl Acetate	1.938	2.133	-10.1	92	0.00
24 P	1,1-Dichloroethane	1.269	1.261	0.6	89	0.00
25 T	2-Butanone	0.550	0.603	-9.6	92	0.00
26 T	2,2-Dichloropropane	0.820	0.956	-16.6	101	0.00
27 T	cis-1,2-Dichloroethene	0.747	0.744	0.4	89	0.00
28 T	Bromochloromethane	0.613	0.606	1.1	91	0.00
29 T	Tetrahydrofuran	0.356	0.378	-6.2	91	0.00
30 C	Chloroform	1.306	1.328	-1.7#	91	0.00
31 T	Cyclohexane	1.122	0.958	14.6	77	0.00
32 T	1,1,1-Trichloroethane	1.105	1.092	1.2	88	0.00
33 S	1,2-Dichloroethane-d4	0.914	0.994	-8.8	103	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	91	0.00
35 S	Dibromofluoromethane	0.355	0.371	-4.5	97	0.00
36 T	1,1-Dichloropropene	0.479	0.436	9.0	82	0.00
37 T	Ethyl Acetate	0.604	0.624	-3.3	91	0.00
38 T	Carbon Tetrachloride	0.518	0.503	2.9	85	0.00
39 T	Methylcyclohexane	0.587	0.530	9.7	77	0.00
40 TM	Benzene	1.463	1.414	3.3	87	0.00
41 T	Methacrylonitrile	0.318	0.363	-14.2	94	0.00
42 TM	1,2-Dichloroethane	0.604	0.654	-8.3	95	0.00
43 T	Isopropyl Acetate	0.915	1.037	-13.3	96	0.00
44 TM	Trichloroethene	0.348	0.334	4.0	86	0.00
45 C	1,2-Dichloropropane	0.365	0.380	-4.1#	92	0.00
46 T	Dibromomethane	0.280	0.297	-6.1	92	0.00
47 T	Bromodichloromethane	0.560	0.603	-7.7	96	0.00
48 T	Methyl methacrylate	0.472	0.544	-15.3	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041625\
 Data File : VX045797.D
 Acq On : 16 Apr 2025 09:28
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 17 01:34:51 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.009	0.0	89	0.00
50 S	Toluene-d8	1.238	1.235	0.2	92	0.00
51 T	4-Methyl-2-Pentanone	0.606	0.687	-13.4	94	0.00
52 CM	Toluene	0.885	0.862	2.6#	86	0.00
53 T	t-1,3-Dichloropropene	0.467	0.560	-19.9	100	0.00
54 T	cis-1,3-Dichloropropene	0.526	0.608	-15.6	97	0.00
55 T	1,1,2-Trichloroethane	0.353	0.379	-7.4	96	0.00
56 T	Ethyl methacrylate	0.546	0.640	-17.2	97	0.00
57 T	1,3-Dichloropropane	0.614	0.662	-7.8	95	0.00
58 T	2-Chloroethyl Vinyl ether	0.276	0.316	-14.5	96	0.00
59 T	2-Hexanone	0.449	0.503	-12.0	93	0.00
60 T	Dibromochloromethane	0.385	0.443	-15.1	98	0.00
61 T	1,2-Dibromoethane	0.357	0.392	-9.8	95	0.00
62 S	4-Bromofluorobenzene	0.451	0.505	-12.0	101	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	91	0.00
64 T	Tetrachloroethene	0.353	0.315	10.8	83	0.00
65 PM	Chlorobenzene	1.068	1.095	-2.5	92	0.00
66 T	1,1,1,2-Tetrachloroethane	0.371	0.399	-7.5	98	0.00
67 C	Ethyl Benzene	1.914	1.888	1.4#	86	0.00
68 T	m/p-Xylenes	0.696	0.691	0.7	87	0.00
69 T	o-Xylene	0.686	0.694	-1.2	88	0.00
70 T	Styrene	1.132	1.230	-8.7	92	0.00
71 P	Bromoform	0.277	0.318	-14.8	98	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
73 T	Isopropylbenzene	4.005	3.753	6.3	86	0.00
74 T	N-amyl acetate	1.916	2.052	-7.1	96	0.00
75 P	1,1,2,2-Tetrachloroethane	1.407	1.380	1.9	96	0.00
76 T	1,2,3-Trichloropropane	1.220	1.217	0.2	96	0.00
77 T	Bromobenzene	0.925	0.899	2.8	92	0.00
78 T	n-propylbenzene	4.613	4.369	5.3	86	0.00
79 T	2-Chlorotoluene	2.949	2.812	4.6	91	0.00
80 T	1,3,5-Trimethylbenzene	3.312	3.237	2.3	89	0.00
81 T	trans-1,4-Dichloro-2-butene	0.359	0.381	-6.1	99	0.00
82 T	4-Chlorotoluene	3.288	3.276	0.4	91	0.00
83 T	tert-Butylbenzene	3.279	3.161	3.6	89	0.00
84 T	1,2,4-Trimethylbenzene	3.324	3.337	-0.4	91	0.00
85 T	sec-Butylbenzene	4.027	3.916	2.8	88	0.00
86 T	p-Isopropyltoluene	3.320	3.266	1.6	89	0.00
87 T	1,3-Dichlorobenzene	1.684	1.645	2.3	93	0.00
88 T	1,4-Dichlorobenzene	1.708	1.675	1.9	95	0.00
89 T	n-Butylbenzene	2.879	2.901	-0.8	88	0.00
90 T	Hexachloroethane	0.571	0.586	-2.6	94	0.00
91 T	1,2-Dichlorobenzene	1.675	1.689	-0.8	97	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.294	0.337	-14.6	103	0.00
93 T	1,2,4-Trichlorobenzene	0.932	1.015	-8.9	103	0.00
94 T	Hexachlorobutadiene	0.402	0.391	2.7	93	0.00
95 T	Naphthalene	3.469	3.766	-8.6	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041625\
Data File : VX045797.D
Acq On : 16 Apr 2025 09:28
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Apr 17 01:34:51 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.976	1.014	-3.9	98	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6




QC SAMPLE DATA

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX040225\
Data File : VX045524.D
Acq On : 01 Apr 2025 16:15
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Title : SW846 8260
Last Update : Wed Apr 02 03:11:43 2025



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632



Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	23.6	7184	PASS
75	95	30	60	58.2	17768	PASS
95	95	100	100	100.0	30504	PASS
96	95	5	9	6.3	1916	PASS
173	174	0.00	2	1.2	253	PASS
174	95	50	100	66.8	20365	PASS
175	174	5	9	7.8	1596	PASS
176	174	95	101	97.8	19923	PASS
177	176	5	9	6.8	1352	PASS

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041625\
Data File : VX045796.D
Acq On : 16 Apr 2025 08:26
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Title : SW846 8260
Last Update : Wed Apr 02 03:11:43 2025



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	22.6	8703	PASS
75	95	30	60	56.1	21597	PASS
95	95	100	100	100.0	38499	PASS
96	95	5	9	6.3	2426	PASS
173	174	0.00	2	1.3	341	PASS
174	95	50	100	67.0	25776	PASS
175	174	5	9	7.5	1933	PASS
176	174	95	101	96.5	24861	PASS
177	176	5	9	6.6	1642	PASS

Report of Analysis

Client:	ENTACT	Date Collected:	
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	
Client Sample ID:	VX0416WBL01	SDG No.:	Q1822
Lab Sample ID:	VX0416WBL01	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group4
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045799.D	1		04/16/25 11:47	VX041625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	1.00	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	1.00	ug/L
67-66-3	Chloroform	0.25	U	0.25	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.20	U	0.20	1.00	ug/L
71-43-2	Benzene	0.15	U	0.15	1.00	ug/L
108-88-3	Toluene	0.14	U	0.14	1.00	ug/L
127-18-4	Tetrachloroethene	0.23	U	0.23	1.00	ug/L
100-41-4	Ethyl Benzene	0.13	U	0.13	1.00	ug/L
1330-20-7	Total Xylenes	0.36	U	0.36	3.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	53.0		70 (74) - 130 (125)	106%	SPK: 50
1868-53-7	Dibromofluoromethane	50.5		70 (75) - 130 (124)	101%	SPK: 50
2037-26-5	Toluene-d8	49.6		70 (86) - 130 (113)	99%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.9		70 (77) - 130 (121)	104%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	72500	5.544			
540-36-3	1,4-Difluorobenzene	142000	6.757			
3114-55-4	Chlorobenzene-d5	129000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	56300	12.018			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041625\
 Data File : VX045799.D
 Acq On : 16 Apr 2025 11:47
 Operator : JC/MD
 Sample : VX0416WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0416WBL01

Quant Time: Apr 17 01:35:24 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	72471	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	141858	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	129459	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	56266	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	70188	52.960	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.920%
35) Dibromofluoromethane	5.379	113	50855	50.527	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.060%
50) Toluene-d8	8.647	98	174406	49.646	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.300%
62) 4-Bromofluorobenzene	11.079	95	66459	51.937	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	103.880%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041625\
Data File : VX045799.D
Acq On : 16 Apr 2025 11:47
Operator : JC/MD
Sample : VX0416WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0416WBL01

Quant Time: Apr 17 01:35:24 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration



#1
Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.544 min Scan# 731

Delta R.T. -0.006 min

Lab File: VX045799.D

Acq: 16 Apr 2025 11:47

Instrument :

MSVOA_X

ClientSampleId :





VX0416WBL01

Tgt Ion:168 Resp: 72471

Ion Ratio Lower Upper

168 100

99 73.4 52.3 78.5



#33

1,2-Dichloroethane-d4

Concen: 52.960 ug/l

RT: 5.952 min Scan# 798

Delta R.T. -0.000 min

Lab File: VX045799.D




Acq: 16 Apr 2025 11:47


Tgt Ion: 65 Resp: 70188

Ion Ratio Lower Upper

65 100

67 49.1 0.0 99.0









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.757 min Scan# 91
Delta R.T. -0.000 min
Lab File: VX045799.D
Acq: 16 Apr 2025 11:47




Instrument :
MSVOA_X
ClientSampleId :
VX0416WBL01


Tgt Ion:114 Resp: 141858
Ion Ratio Lower Upper
114 100
63 24.0 0.0 46.8
88 17.4 0.0 35.4



#35
Dibromofluoromethane
Concen: 50.527 ug/l
RT: 5.379 min Scan# 704
Delta R.T. -0.000 min
Lab File: VX045799.D
Acq: 16 Apr 2025 11:47

Tgt Ion:113 Resp: 50855
Ion Ratio Lower Upper
113 100
111 102.9 81.8 122.6
192 16.6 13.8 20.6









#50
Toluene-d8
Concen: 49.646 ug/l
RT: 8.647 min Scan# 1240
Delta R.T. -0.000 min
Lab File: VX045799.D
Acq: 16 Apr 2025 11:47




Instrument :
MSVOA_X
ClientSampleId :
VX0416WBL01


Tgt Ion: 98 Resp: 174406
Ion Ratio Lower Upper
98 100
100 65.1 52.2 78.4



#62
4-Bromofluorobenzene
Concen: 51.937 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX045799.D
Acq: 16 Apr 2025 11:47

Tgt Ion: 95 Resp: 66459
Ion Ratio Lower Upper
95 100
174 68.5 0.0 135.8
176 64.9 0.0 131.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. -0.000 min
Lab File: VX045799.D
Acq: 16 Apr 2025 11:47




Instrument :
MSVOA_X
ClientSampleId :
VX0416WBL01

Tgt Ion	Ratio	Lower	Upper
117	100		
82	61.8	49.2	73.8
119	32.2	25.1	37.7



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VX045799.D
Acq: 16 Apr 2025 11:47

Tgt Ion	Ratio	Lower	Upper
152	100		
115	66.4	46.9	140.7
150	157.0	0.0	349.4



Report of Analysis

Client:	ENTACT	Date Collected:	
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	
Client Sample ID:	VX0416WBS01	SDG No.:	Q1822
Lab Sample ID:	VX0416WBS01	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group4
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045800.D	1		04/16/25 12:11	VX041625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
1634-04-4	Methyl tert-butyl Ether	20.4		0.16	1.00	ug/L
56-23-5	Carbon Tetrachloride	19.9		0.25	1.00	ug/L
67-66-3	Chloroform	19.7		0.25	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	20.1		0.20	1.00	ug/L
71-43-2	Benzene	18.9		0.15	1.00	ug/L
108-88-3	Toluene	19.2		0.14	1.00	ug/L
127-18-4	Tetrachloroethene	19.2		0.23	1.00	ug/L
100-41-4	Ethyl Benzene	19.6		0.13	1.00	ug/L
1330-20-7	Total Xylenes	58.9		0.36	3.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	53.5		70 (74) - 130 (125)	107%	SPK: 50
1868-53-7	Dibromofluoromethane	52.4		70 (75) - 130 (124)	105%	SPK: 50
2037-26-5	Toluene-d8	51.6		70 (86) - 130 (113)	103%	SPK: 50
460-00-4	4-Bromofluorobenzene	54.2		70 (77) - 130 (121)	108%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	91300	5.544			
540-36-3	1,4-Difluorobenzene	163000	6.757			
3114-55-4	Chlorobenzene-d5	145000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	66600	12.018			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041625\
 Data File : VX045800.D
 Acq On : 16 Apr 2025 12:11
 Operator : JC/MD
 Sample : VX0416WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0416WBS01

Manual Integrations
 APPROVED

Reviewed By : John Carlone 04/17/2025
 Supervised By : Mahesh Dadoda 04/17/2025

Quant Time: Apr 17 01:35:35 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	91256	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	163123	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	144629	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	66591	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	89259	53.486	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 106.980%			
35) Dibromofluoromethane	5.379	113	60634	52.390	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 104.780%			
50) Toluene-d8	8.647	98	208286	51.561	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 103.120%			
62) 4-Bromofluorobenzene	11.079	95	79824	54.249	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 108.500%			
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.166	85	26223	19.215	ug/l	97
3) Chloromethane	1.301	50	24106	17.194	ug/l	100
4) Vinyl Chloride	1.374	62	22830	17.793	ug/l	97
5) Bromomethane	1.593	94	10735	17.645	ug/l	100
6) Chloroethane	1.666	64	13251	19.465	ug/l	97
7) Trichlorofluoromethane	1.874	101	37615	19.659	ug/l	100
8) Diethyl Ether	2.130	74	12179	18.960	ug/l	95
9) 1,1,2-Trichlorotrifluo...	2.319	101	22629	20.183	ug/l	97
10) Methyl Iodide	2.447	142	24390	17.441	ug/l	99
11) Tert butyl alcohol	2.971	59	22536	100.482	ug/l	100
12) 1,1-Dichloroethene	2.313	96	20145	18.388	ug/l	99
13) Acrolein	2.233	56	22864	74.030	ug/l	99
14) Allyl chloride	2.654	41	41079	19.762	ug/l	99
15) Acrylonitrile	3.062	53	68066	96.143	ug/l	99
16) Acetone	2.386	43	64882	94.681	ug/l	98
17) Carbon Disulfide	2.502	76	40473	14.958	ug/l	98
18) Methyl Acetate	2.703	43	37513	23.776	ug/l	98
19) Methyl tert-butyl Ether	3.111	73	78318	20.436	ug/l	100
20) Methylene Chloride	2.782	84	24113	18.796	ug/l	96
21) trans-1,2-Dichloroethene	3.081	96	20286	18.125	ug/l	97
22) Diisopropyl ether	3.764	45	81859	19.992	ug/l	99
23) Vinyl Acetate	3.721	43	339728	96.030	ug/l	99
24) 1,1-Dichloroethane	3.605	63	44275	19.120	ug/l	99
25) 2-Butanone	4.562	43	99057	98.751	ug/l	99
26) 2,2-Dichloropropane	4.465	77	33716	22.535	ug/l	100
27) cis-1,2-Dichloroethene	4.489	96	26190	19.207	ug/l	96
28) Bromochloromethane	4.891	49	22857	20.426	ug/l	98
29) Tetrahydrofuran	5.001	42	63198	97.177	ug/l	99
30) Chloroform	5.087	83	47018	19.733	ug/l	99
31) Cyclohexane	5.465	56	38188	18.654	ug/l	91
32) 1,1,1-Trichloroethane	5.379	97	40463	20.069	ug/l	100
36) 1,1-Dichloropropene	5.690	75	28998	18.557	ug/l	99
37) Ethyl Acetate	4.715	43	36476	18.523	ug/l	97
38) Carbon Tetrachloride	5.666	117	33560	19.871	ug/l	95
39) Methylcyclohexane	7.373	83	36471	19.040	ug/l	94
40) Benzene	6.031	78	90157	18.891	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041625\
 Data File : VX045800.D
 Acq On : 16 Apr 2025 12:11
 Operator : JC/MD
 Sample : VX0416WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0416WBS01

Manual Integrations
 APPROVED

Reviewed By : John Carlone 04/17/2025
 Supervised By : Mahesh Dadoda 04/17/2025

Quant Time: Apr 17 01:35:35 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	21777	20.959	ug/l	96
42) 1,2-Dichloroethane	6.086	62	40085	20.336	ug/l	100
43) Isopropyl Acetate	6.342	43	59835	20.047	ug/l	99
44) Trichloroethene	7.123	130	21854	19.266	ug/l	97
45) 1,2-Dichloropropane	7.428	63	22835	19.176	ug/l	96
46) Dibromomethane	7.580	93	17700	19.363	ug/l	96
47) Bromodichloromethane	7.818	83	35805	19.614	ug/l	99
48) Methyl methacrylate	7.690	41	30857	20.023	ug/l	97
49) 1,4-Dioxane	7.659	88	11317	405.891	ug/l	95
51) 4-Methyl-2-Pentanone	8.574	43	205623	103.922	ug/l	100
52) Toluene	8.714	92	55575	19.244	ug/l	99
53) t-1,3-Dichloropropene	8.976	75	30071	18.634	ug/l	98
54) cis-1,3-Dichloropropene	8.366	75	34942	20.369	ug/l	97
55) 1,1,2-Trichloroethane	9.147	97	23308	20.254	ug/l	95
56) Ethyl methacrylate	9.116	69	36340	20.405	ug/l	98
57) 1,3-Dichloropropane	9.305	76	40377	20.149	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.238	63	85031	94.380	ug/l	99
59) 2-Hexanone	9.427	43	151213	103.202	ug/l	100
60) Dibromochloromethane	9.519	129	25122	20.024	ug/l	98
61) 1,2-Dibromoethane	9.610	107	23511	20.182	ug/l	98
64) Tetrachloroethene	9.269	164	19633	19.227	ug/l	94
65) Chlorobenzene	10.080	112	59838	19.362	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.159	131	21536	20.072	ug/l	98
67) Ethyl Benzene	10.189	91	108314	19.569	ug/l	99
68) m/p-Xylenes	10.299	106	78456	38.972	ug/l	97
69) o-Xylene	10.640	106	39382	19.858	ug/l	100
70) Styrene	10.653	104	64978	19.852	ug/l	99
71) Bromoform	10.799	173	15526	19.380	ug/l #	97
73) Isopropylbenzene	10.957	105	103860	19.471	ug/l	100
74) N-amyl acetate	10.842	43	50004	19.598	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.213	83	36361	19.405	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	31787m	19.570	ug/l	
77) Bromobenzene	11.195	156	23662	19.199	ug/l	99
78) n-propylbenzene	11.305	91	119070	19.380	ug/l	100
79) 2-Chlorotoluene	11.360	91	76695	19.526	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	87621	19.865	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	8601	17.971	ug/l	95
82) 4-Chlorotoluene	11.451	91	85739	19.578	ug/l	99
83) tert-Butylbenzene	11.713	119	87028	19.929	ug/l	97
84) 1,2,4-Trimethylbenzene	11.750	105	88156	19.913	ug/l	99
85) sec-Butylbenzene	11.890	105	107627	20.067	ug/l	99
86) p-Isopropyltoluene	12.006	119	88318	19.973	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	43357	19.326	ug/l	100
88) 1,4-Dichlorobenzene	12.036	146	43462	19.110	ug/l	96
89) n-Butylbenzene	12.329	91	73529	19.176	ug/l	100
90) Hexachloroethane	12.536	117	14606	19.220	ug/l	99
91) 1,2-Dichlorobenzene	12.335	146	43884	19.675	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	12.945	75	8271	21.106	ug/l	94
93) 1,2,4-Trichlorobenzene	13.585	180	24532	19.760	ug/l	97
94) Hexachlorobutadiene	13.725	225	10725	20.047	ug/l	99
95) Naphthalene	13.774	128	90877	19.670	ug/l	99
96) 1,2,3-Trichlorobenzene	13.963	180	24818	19.090	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041625\
Data File : VX045800.D
Acq On : 16 Apr 2025 12:11
Operator : JC/MD
Sample : VX0416WBS01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0416WBS01

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/17/2025
Supervised By :Mahesh Dadoda 04/17/2025

Quant Time: Apr 17 01:35:35 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041625\
 Data File : VX045800.D
 Acq On : 16 Apr 2025 12:11
 Operator : JC/MD
 Sample : VX0416WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0416WBS01

Manual Integrations APPROVED

Reviewed By : John Carlone 04/17/2025
 Supervised By : Mahesh Dadoda 04/17/2025

Quant Time: Apr 17 01:35:35 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration



Report of Analysis

Client:	ENTACT			Date Collected:	
Project:	540 Degraw St, Brooklyn, NY - E9309			Date Received:	
Client Sample ID:	VX0416WBSD01			SDG No.:	Q1822
Lab Sample ID:	VX0416WBSD01			Matrix:	Water
Analytical Method:	SW8260			% Solid:	0
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000 uL
Soil Aliquot Vol:			uL	Test:	VOCMS Group4
GC Column:	DB-624UI	ID :	0.18	Level :	LOW
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045801.D	1		04/16/25 12:39	VX041625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
1634-04-4	Methyl tert-butyl Ether	21.0		0.16	1.00	ug/L
56-23-5	Carbon Tetrachloride	19.8		0.25	1.00	ug/L
67-66-3	Chloroform	20.0		0.25	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	19.9		0.20	1.00	ug/L
71-43-2	Benzene	19.1		0.15	1.00	ug/L
108-88-3	Toluene	19.3		0.14	1.00	ug/L
127-18-4	Tetrachloroethene	19.4		0.23	1.00	ug/L
100-41-4	Ethyl Benzene	19.8		0.13	1.00	ug/L
1330-20-7	Total Xylenes	59.9		0.36	3.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	54.1		70 (74) - 130 (125)	108%	SPK: 50
1868-53-7	Dibromofluoromethane	53.4		70 (75) - 130 (124)	107%	SPK: 50
2037-26-5	Toluene-d8	51.1		70 (86) - 130 (113)	102%	SPK: 50
460-00-4	4-Bromofluorobenzene	54.4		70 (77) - 130 (121)	109%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	87900	5.543			
540-36-3	1,4-Difluorobenzene	156000	6.756			
3114-55-4	Chlorobenzene-d5	138000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	65200	12.018			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041625\
 Data File : VX045801.D
 Acq On : 16 Apr 2025 12:39
 Operator : JC/MD
 Sample : VX0416WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0416WBSD01

Manual Integrations
 APPROVED

Reviewed By : John Carlone 04/17/2025
 Supervised By : Mahesh Dadoda 04/17/2025

Quant Time: Apr 17 01:35:54 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	87856	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.756	114	155557	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	137506	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	65216	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	86933	54.108	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 108.220%			
35) Dibromofluoromethane	5.379	113	58987	53.446	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 106.900%			
50) Toluene-d8	8.646	98	196683	51.056	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 102.120%			
62) 4-Bromofluorobenzene	11.079	95	76265	54.351	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 108.700%			
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.166	85	24851	18.914	ug/l	99
3) Chloromethane	1.306	50	23070	17.091	ug/l	96
4) Vinyl Chloride	1.373	62	21187	17.151	ug/l	99
5) Bromomethane	1.593	94	10130	17.295	ug/l	99
6) Chloroethane	1.672	64	12287	18.748	ug/l	100
7) Trichlorofluoromethane	1.879	101	35715	19.389	ug/l	98
8) Diethyl Ether	2.129	74	11900	19.242	ug/l	94
9) 1,1,2-Trichlorotrifluo...	2.318	101	21338	19.768	ug/l	99
10) Methyl Iodide	2.446	142	23618	17.542	ug/l	97
11) Tert butyl alcohol	2.964	59	23467	108.683	ug/l	100
12) 1,1-Dichloroethene	2.312	96	18997	18.012	ug/l	96
13) Acrolein	2.233	56	23340	78.496	ug/l	99
14) Allyl chloride	2.660	41	38368	19.173	ug/l	99
15) Acrylonitrile	3.062	53	71328	104.650	ug/l	99
16) Acetone	2.379	43	66208	100.355	ug/l	99
17) Carbon Disulfide	2.501	76	38642	14.834	ug/l	99
18) Methyl Acetate	2.702	43	37874	24.934	ug/l	100
19) Methyl tert-butyl Ether	3.111	73	77510	21.008	ug/l	98
20) Methylene Chloride	2.782	84	23309	18.872	ug/l	98
21) trans-1,2-Dichloroethene	3.080	96	19618	18.207	ug/l	99
22) Diisopropyl ether	3.757	45	79262	20.107	ug/l #	86
23) Vinyl Acetate	3.720	43	339884	99.793	ug/l	100
24) 1,1-Dichloroethane	3.605	63	42449	19.041	ug/l	97
25) 2-Butanone	4.550	43	102829	106.479	ug/l	99
26) 2,2-Dichloropropane	4.464	77	31433	21.823	ug/l	99
27) cis-1,2-Dichloroethene	4.482	96	24812	18.900	ug/l	98
28) Bromochloromethane	4.897	49	22529	20.912	ug/l	100
29) Tetrahydrofuran	5.001	42	63966	102.164	ug/l	99
30) Chloroform	5.080	83	45977	20.043	ug/l	97
31) Cyclohexane	5.458	56	35724	18.126	ug/l	96
32) 1,1,1-Trichloroethane	5.379	97	38638	19.906	ug/l	98
36) 1,1-Dichloropropene	5.684	75	28481	19.112	ug/l	99
37) Ethyl Acetate	4.714	43	36657	19.520	ug/l	98
38) Carbon Tetrachloride	5.671	117	31865	19.785	ug/l	95
39) Methylcyclohexane	7.372	83	35094	19.212	ug/l	96
40) Benzene	6.031	78	86831	19.079	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041625\
 Data File : VX045801.D
 Acq On : 16 Apr 2025 12:39
 Operator : JC/MD
 Sample : VX0416WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0416WBSD01

Manual Integrations
 APPROVED

Reviewed By : John Carlone 04/17/2025
 Supervised By : Mahesh Dadoda 04/17/2025

Quant Time: Apr 17 01:35:54 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.915	41	22229	22.434	ug/l	97
42) 1,2-Dichloroethane	6.086	62	39385	20.952	ug/l	99
43) Isopropyl Acetate	6.336	43	57906	20.344	ug/l	99
44) Trichloroethene	7.116	130	20068	18.552	ug/l	89
45) 1,2-Dichloropropane	7.427	63	22646	19.943	ug/l	99
46) Dibromomethane	7.580	93	17433	19.999	ug/l	97
47) Bromodichloromethane	7.817	83	34562	19.854	ug/l	99
48) Methyl methacrylate	7.689	41	30732	20.912	ug/l	98
49) 1,4-Dioxane	7.659	88	11765	442.482	ug/l	98
51) 4-Methyl-2-Pentanone	8.567	43	206961	109.686	ug/l	98
52) Toluene	8.713	92	53091	19.278	ug/l	99
53) t-1,3-Dichloropropene	8.976	75	29771	19.258	ug/l	99
54) cis-1,3-Dichloropropene	8.360	75	33125	20.249	ug/l	95
55) 1,1,2-Trichloroethane	9.146	97	22846	20.818	ug/l	95
56) Ethyl methacrylate	9.116	69	35457	20.877	ug/l	99
57) 1,3-Dichloropropane	9.305	76	39129	20.476	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.238	63	84260	98.074	ug/l	100
59) 2-Hexanone	9.427	43	153170	109.622	ug/l	100
60) Dibromochloromethane	9.518	129	24174	20.206	ug/l	99
61) 1,2-Dibromoethane	9.610	107	23320	20.992	ug/l	97
64) Tetrachloroethene	9.268	164	18792	19.357	ug/l	94
65) Chlorobenzene	10.079	112	57965	19.728	ug/l	98
66) 1,1,1,2-Tetrachloroethane	10.158	131	20528	20.123	ug/l	99
67) Ethyl Benzene	10.189	91	104457	19.849	ug/l	99
68) m/p-Xylenes	10.299	106	76653	40.048	ug/l	99
69) o-Xylene	10.640	106	37427	19.850	ug/l	99
70) Styrene	10.652	104	62215	19.992	ug/l	98
71) Bromoform	10.798	173	15183	19.933	ug/l #	96
73) Isopropylbenzene	10.963	105	99759	19.097	ug/l	100
74) N-amyl acetate	10.841	43	49240	19.705	ug/l	98
75) 1,1,2,2-Tetrachloroethane	11.213	83	35966	19.599	ug/l	100
76) 1,2,3-Trichloropropane	11.237	75	31084m	19.540	ug/l	
77) Bromobenzene	11.195	156	23388	19.377	ug/l	98
78) n-propylbenzene	11.298	91	115002	19.113	ug/l	100
79) 2-Chlorotoluene	11.359	91	74729	19.427	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	84999	19.677	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.018	75	8674	18.506	ug/l	94
82) 4-Chlorotoluene	11.451	91	85155	19.855	ug/l	99
83) tert-Butylbenzene	11.713	119	85181	19.918	ug/l	99
84) 1,2,4-Trimethylbenzene	11.749	105	85190	19.649	ug/l	100
85) sec-Butylbenzene	11.890	105	105375	20.061	ug/l	100
86) p-Isopropyltoluene	12.006	119	84535	19.521	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	42176	19.196	ug/l	100
88) 1,4-Dichlorobenzene	12.036	146	42153	18.925	ug/l	94
89) n-Butylbenzene	12.329	91	73547	19.585	ug/l	100
90) Hexachloroethane	12.536	117	14766	19.840	ug/l	97
91) 1,2-Dichlorobenzene	12.335	146	42843	19.613	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	12.944	75	8361	21.786	ug/l	95
93) 1,2,4-Trichlorobenzene	13.585	180	23654	19.454	ug/l	98
94) Hexachlorobutadiene	13.725	225	10799	20.611	ug/l	98
95) Naphthalene	13.774	128	91763	20.280	ug/l	100
96) 1,2,3-Trichlorobenzene	13.963	180	24916	19.570	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041625\
Data File : VX045801.D
Acq On : 16 Apr 2025 12:39
Operator : JC/MD
Sample : VX0416WBSD01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0416WBSD01

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/17/2025
Supervised By :Mahesh Dadoda 04/17/2025

Quant Time: Apr 17 01:35:54 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041625\
Data File : VX045801.D
Acq On : 16 Apr 2025 12:39
Operator : JC/MD
Sample : VX0416WBSD01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0416WBSD01

Manual Integrations APPROVED

Reviewed By :John Carlone 04/17/2025
Supervised By :Mahesh Dadoda 04/17/2025

Quant Time: Apr 17 01:35:54 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration



Manual Integration Report

Sequence:	VX040225	Instrument	MSVOA_x
-----------	----------	------------	---------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDICC001	VX045525.D	1,2,3-Trichloropropane	Amit	4/2/2025 2:14:20 PM	MMDadoda	4/2/2025 2:16:05 PM	Peak Integrated by Software
VSTDICC001	VX045525.D	1,4-Dichlorobenzene	Amit	4/2/2025 2:14:20 PM	MMDadoda	4/2/2025 2:16:05 PM	Peak Integrated by Software
VSTDICC001	VX045525.D	1,4-Dioxane	Amit	4/2/2025 2:14:20 PM	MMDadoda	4/2/2025 2:16:05 PM	Peak Integrated by Software
VSTDICC001	VX045525.D	Ethyl methacrylate	Amit	4/2/2025 2:14:20 PM	MMDadoda	4/2/2025 2:16:05 PM	Peak Integrated by Software
VSTDICC001	VX045525.D	Methacrylonitrile	Amit	4/2/2025 2:14:20 PM	MMDadoda	4/2/2025 2:16:05 PM	Peak Integrated by Software
VSTDICC005	VX045526.D	1,2,3-Trichloropropane	Amit	4/2/2025 2:14:22 PM	MMDadoda	4/2/2025 2:16:07 PM	Peak Integrated by Software
VSTDICC020	VX045527.D	1,2,3-Trichloropropane	Amit	4/2/2025 2:14:25 PM	MMDadoda	4/2/2025 2:16:09 PM	Peak Integrated by Software
VSTDICCC050	VX045528.D	1,2,3-Trichloropropane	Amit	4/2/2025 2:14:26 PM	MMDadoda	4/2/2025 2:16:11 PM	Peak Integrated by Software
VSTDICC100	VX045529.D	1,2,3-Trichloropropane	Amit	4/2/2025 2:14:28 PM	MMDadoda	4/2/2025 2:16:15 PM	Peak Integrated by Software
VSTDICC150	VX045530.D	1,2,3-Trichloropropane	Amit	4/2/2025 2:14:30 PM	MMDadoda	4/2/2025 2:16:46 PM	Peak Integrated by Software
VSTDICV050	VX045532.D	1,2,3-Trichloropropane	Amit	4/2/2025 2:14:32 PM	MMDadoda	4/2/2025 2:17:50 PM	Peak Integrated by Software
VSTDCCC050	VX045534.D	1,2,3-Trichloropropane	JOHN	4/3/2025 10:04:10 AM	MMDadoda	4/3/2025 1:12:02 PM	Peak Integrated by Software
VSTDCCC050	VX045557.D	1,2,3-Trichloropropane	JOHN	4/3/2025 10:04:58 AM	MMDadoda	4/3/2025 1:12:08 PM	Peak Integrated by Software



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:

VX040225

Instrument

MSVOA_x

Sample ID

File ID

Parameter

Review By

Review On

Supervised
By

Supervised On

Reason

Manual Integration Report

Sequence:

VX041625

Instrument

MSVOA_x

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDCCC050	VX045797.D	1,2,3-Trichloropropane	JOHN	4/17/2025 9:02:04 AM	MMDadoda	4/17/2025 1:41:09 PM	Peak Integrated by Software
VX0416WBS01	VX045800.D	1,2,3-Trichloropropane	JOHN	4/17/2025 9:02:08 AM	MMDadoda	4/17/2025 1:41:06 PM	Peak Integrated by Software
VX0416WBSD0 1	VX045801.D	1,2,3-Trichloropropane	JOHN	4/17/2025 9:02:13 AM	MMDadoda	4/17/2025 1:41:05 PM	Peak Integrated by Software
VSTDCCC050	VX045815.D	1,2,3-Trichloropropane	JOHN	4/17/2025 9:02:44 AM	MMDadoda	4/17/2025 1:40:57 PM	Peak Integrated by Software

Instrument ID: **MSVOA_X**

Daily Analysis Runlog For Sequence/QC Batch ID # VX040225

Review By	John Carlone	Review On	4/2/2025 9:42:46 AM
Supervise By	Amit Patel	Supervise On	4/2/2025 2:14:39 PM
SubDirectory	VX040225	HP Acquire Method	HP Processing Method 82X040225W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP133559,VP133563		
Initial Calibration Stds	VP133569,VP133570,VP133571,VP133572,VP133573,VP133574		
CCC	VP133564,VP133565		
Internal Standard/PEM			
ICV/I.BLK	VP133575		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VX045524.D	01 Apr 2025 16:15	JC/MD	Ok
2	VSTDICCC001	VX045525.D	01 Apr 2025 17:06	JC/MD	Ok,M
3	VSTDICCC005	VX045526.D	01 Apr 2025 17:29	JC/MD	Ok,M
4	VSTDICCC020	VX045527.D	01 Apr 2025 17:52	JC/MD	Ok,M
5	VSTDICCC050	VX045528.D	01 Apr 2025 18:15	JC/MD	Ok,M
6	VSTDICCC100	VX045529.D	01 Apr 2025 18:38	JC/MD	Ok,M
7	VSTDICCC150	VX045530.D	01 Apr 2025 19:02	JC/MD	Ok,M
8	IBLK	VX045531.D	01 Apr 2025 19:25	JC/MD	Ok
9	VSTDICV050	VX045532.D	01 Apr 2025 19:48	JC/MD	Ok,M
10	BFB	VX045533.D	02 Apr 2025 09:30	JC/MD	Ok
11	VSTDICCC050	VX045534.D	02 Apr 2025 10:02	JC/MD	Ok,M
12	VX0402MBL01	VX045535.D	02 Apr 2025 10:30	JC/MD	Ok
13	VX0402WBL01	VX045536.D	02 Apr 2025 10:53	JC/MD	Ok
14	VX0402WBS01	VX045537.D	02 Apr 2025 11:16	JC/MD	Ok,M
15	VX0402WBSD01	VX045538.D	02 Apr 2025 11:44	JC/MD	Ok,M
16	Q1697-01	VX045539.D	02 Apr 2025 12:07	JC/MD	Dilution
17	Q1697-02	VX045540.D	02 Apr 2025 12:30	JC/MD	Dilution
18	IBLK	VX045541.D	02 Apr 2025 12:54	JC/MD	Ok
19	Q1697-01DL	VX045542.D	02 Apr 2025 13:21	JC/MD	Ok
20	Q1697-02DL	VX045543.D	02 Apr 2025 13:44	JC/MD	Ok
21	Q1697-03	VX045544.D	02 Apr 2025 14:07	JC/MD	Not Ok

Instrument ID: **MSVOA_X**

Daily Analysis Runlog For Sequence/QC Batch ID # VX040225

Review By	John Carlone	Review On	4/2/2025 9:42:46 AM
Supervise By	Amit Patel	Supervise On	4/2/2025 2:14:39 PM
SubDirectory	VX040225	HP Acquire Method	HP Processing Method 82X040225W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP133559,VP133563		
Initial Calibration Stds	VP133569,VP133570,VP133571,VP133572,VP133573,VP133574		
CCC	VP133564,VP133565		
Internal Standard/PEM			
ICV/I.BLK	VP133575		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	Q1697-04	VX045545.D	02 Apr 2025 14:31	JC/MD	Not Ok
23	Q1697-05	VX045546.D	02 Apr 2025 14:54	JC/MD	Not Ok
24	Q1697-07	VX045547.D	02 Apr 2025 15:18	JC/MD	Not Ok
25	IBLK	VX045548.D	02 Apr 2025 15:41	JC/MD	Ok
26	Q1697-03	VX045549.D	02 Apr 2025 16:04	JC/MD	Ok
27	Q1697-04	VX045550.D	02 Apr 2025 16:27	JC/MD	Ok
28	Q1697-05	VX045551.D	02 Apr 2025 16:51	JC/MD	Ok
29	Q1697-06	VX045552.D	02 Apr 2025 17:14	JC/MD	ReRun
30	Q1623-01	VX045553.D	02 Apr 2025 17:37	JC/MD	Ok
31	Q1623-02	VX045554.D	02 Apr 2025 18:01	JC/MD	Ok
32	Q1623-03	VX045555.D	02 Apr 2025 18:24	JC/MD	Ok
33	Q1623-04	VX045556.D	02 Apr 2025 18:47	JC/MD	Ok
34	VSTDCCC050	VX045557.D	02 Apr 2025 19:10	JC/MD	Ok,M

M : Manual Integration

Instrument ID: **MSVOA_X**

Daily Analysis Runlog For Sequence/QC Batch ID # VX041625

Review By	John Carlone	Review On	4/17/2025 9:03:49 AM
Supervise By	Mahesh Dadoda	Supervise On	4/17/2025 1:41:16 PM
SubDirectory	VX041625	HP Acquire Method	HP Processing Method 82X040225W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP133677		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133678,VP133679		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VX045796.D	16 Apr 2025 08:26	JC/MD	Ok
2	VSTDCCC050	VX045797.D	16 Apr 2025 09:28	JC/MD	Ok,M
3	VX0416MBL01	VX045798.D	16 Apr 2025 11:24	JC/MD	Ok
4	VX0416WBL01	VX045799.D	16 Apr 2025 11:47	JC/MD	Ok
5	VX0416WBS01	VX045800.D	16 Apr 2025 12:11	JC/MD	Ok,M
6	VX0416WBSD01	VX045801.D	16 Apr 2025 12:39	JC/MD	Ok,M
7	PB167604TB	VX045802.D	16 Apr 2025 13:02	JC/MD	Ok
8	PB167604ZHE#61	VX045803.D	16 Apr 2025 13:25	JC/MD	Ok,M
9	PB167604ZHE#62	VX045804.D	16 Apr 2025 13:48	JC/MD	Ok
10	PB167604ZHE#63	VX045805.D	16 Apr 2025 14:11	JC/MD	Ok,M
11	PB167604ZHE#64	VX045806.D	16 Apr 2025 14:35	JC/MD	Ok
12	PB167604ZHE#65	VX045807.D	16 Apr 2025 14:58	JC/MD	Ok
13	Q1800-01	VX045808.D	16 Apr 2025 15:21	JC/MD	Ok
14	Q1803-01	VX045809.D	16 Apr 2025 15:44	JC/MD	Ok
15	Q1803-02	VX045810.D	16 Apr 2025 16:08	JC/MD	Ok
16	Q1808-02	VX045811.D	16 Apr 2025 16:31	JC/MD	Ok,M
17	Q1808-04	VX045812.D	16 Apr 2025 16:54	JC/MD	Ok,M
18	IBLK	VX045813.D	16 Apr 2025 17:17	JC/MD	Ok
19	Q1822-01	VX045814.D	16 Apr 2025 17:41	JC/MD	Ok
20	VSTDCCC050	VX045815.D	16 Apr 2025 18:04	JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX040225

Review By	John Carlone	Review On	4/2/2025 9:42:46 AM
Supervise By	Amit Patel	Supervise On	4/2/2025 2:14:39 PM
SubDirectory	VX040225	HP Acquire Method	HP Processing Method 82X040225W.M

STD. NAME	STD REF.#
Tune/Reschk	VP133559,VP133563
Initial Calibration Stds	VP133569,VP133570,VP133571,VP133572,VP133573,VP133574
CCC	VP133564,VP133565
Internal Standard/PEM	
ICV/I.BLK	VP133575
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX045524.D	01 Apr 2025 16:15		JC/MD	Ok
2	VSTDIC001	VSTDIC001	VX045525.D	01 Apr 2025 17:06	%D failed for Comp. #53 in 01PPB	JC/MD	Ok,M
3	VSTDIC005	VSTDIC005	VX045526.D	01 Apr 2025 17:29		JC/MD	Ok,M
4	VSTDIC020	VSTDIC020	VX045527.D	01 Apr 2025 17:52		JC/MD	Ok,M
5	VSTDIC050	VSTDIC050	VX045528.D	01 Apr 2025 18:15		JC/MD	Ok,M
6	VSTDIC100	VSTDIC100	VX045529.D	01 Apr 2025 18:38		JC/MD	Ok,M
7	VSTDIC150	VSTDIC150	VX045530.D	01 Apr 2025 19:02		JC/MD	Ok,M
8	IBLK	IBLK	VX045531.D	01 Apr 2025 19:25		JC/MD	Ok
9	VSTDICV050	ICVVX040225	VX045532.D	01 Apr 2025 19:48		JC/MD	Ok,M
10	BFB	BFB	VX045533.D	02 Apr 2025 09:30		JC/MD	Ok
11	VSTDIC050	VSTDIC050	VX045534.D	02 Apr 2025 10:02	pH#Lot#V12668	JC/MD	Ok,M
12	VX0402MBL01	VX0402MBL01	VX045535.D	02 Apr 2025 10:30		JC/MD	Ok
13	VX0402WBL01	VX0402WBL01	VX045536.D	02 Apr 2025 10:53		JC/MD	Ok
14	VX0402WBS01	VX0402WBS01	VX045537.D	02 Apr 2025 11:16		JC/MD	Ok,M
15	VX0402WBSD01	VX0402WBSD01	VX045538.D	02 Apr 2025 11:44		JC/MD	Ok,M
16	Q1697-01	MW-19B-72-040125	VX045539.D	02 Apr 2025 12:07	vial A pH<2 Need 200X	JC/MD	Dilution
17	Q1697-02	IW-01-55-040125	VX045540.D	02 Apr 2025 12:30	vial A pH<2 Need 10X	JC/MD	Dilution

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX040225

Review By	John Carlone	Review On	4/2/2025 9:42:46 AM
Supervise By	Amit Patel	Supervise On	4/2/2025 2:14:39 PM
SubDirectory	VX040225	HP Acquire Method	HP Processing Method 82X040225W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP133559,VP133563		
Initial Calibration Stds	VP133569,VP133570,VP133571,VP133572,VP133573,VP133574		
CCC	VP133564,VP133565		
Internal Standard/PEM			
ICV/I.BLK	VP133575		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	IBLK	IBLK	VX045541.D	02 Apr 2025 12:54		JC/MD	Ok
19	Q1697-01DL	MW-19B-72-040125DL	VX045542.D	02 Apr 2025 13:21	vial B pH<2	JC/MD	Ok
20	Q1697-02DL	IW-01-55-040125DL	VX045543.D	02 Apr 2025 13:44	vial B pH<2	JC/MD	Ok
21	Q1697-03	IW-02-55-040125	VX045544.D	02 Apr 2025 14:07	Need lower dilution	JC/MD	Not Ok
22	Q1697-04	IW-02-55-040125-FD	VX045545.D	02 Apr 2025 14:31	Need lower dilution	JC/MD	Not Ok
23	Q1697-05	IW-03-55-040125	VX045546.D	02 Apr 2025 14:54	Need lower dilution	JC/MD	Not Ok
24	Q1697-07	MW-19B-72-040125	VX045547.D	02 Apr 2025 15:18	Not On Login	JC/MD	Not Ok
25	IBLK	IBLK	VX045548.D	02 Apr 2025 15:41		JC/MD	Ok
26	Q1697-03	IW-02-55-040125	VX045549.D	02 Apr 2025 16:04	vial A pH<2	JC/MD	Ok
27	Q1697-04	IW-02-55-040125-FD	VX045550.D	02 Apr 2025 16:27	vial A pH<2	JC/MD	Ok
28	Q1697-05	IW-03-55-040125	VX045551.D	02 Apr 2025 16:51	vial A pH<2	JC/MD	Ok
29	Q1697-06	TB-01-040125	VX045552.D	02 Apr 2025 17:14	vial A pH<2 TB;Hit of comp.#44	JC/MD	ReRun
30	Q1623-01	Storage-Blank-SOIL-RE	VX045553.D	02 Apr 2025 17:37	vial A pH<2	JC/MD	Ok
31	Q1623-02	Storage-Blank-WATER	VX045554.D	02 Apr 2025 18:01	vial A pH<2	JC/MD	Ok
32	Q1623-03	Storage-Blank-WATER	VX045555.D	02 Apr 2025 18:24	vial A pH<2	JC/MD	Ok
33	Q1623-04	Storage-Blank-SAMPLE	VX045556.D	02 Apr 2025 18:47	vial A pH<2	JC/MD	Ok
34	VSTDCCC050	VSTDCCC050EC	VX045557.D	02 Apr 2025 19:10		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX041625

Review By	John Carlone	Review On	4/17/2025 9:03:49 AM
Supervise By	Mahesh Dadoda	Supervise On	4/17/2025 1:41:16 PM
SubDirectory	VX041625	HP Acquire Method	HP Processing Method 82X040225W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133677
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133678,VP133679

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX045796.D	16 Apr 2025 08:26		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX045797.D	16 Apr 2025 09:28	pH#Lot#V12668	JC/MD	Ok,M
3	VX0416MBL01	VX0416MBL01	VX045798.D	16 Apr 2025 11:24		JC/MD	Ok
4	VX0416WBL01	VX0416WBL01	VX045799.D	16 Apr 2025 11:47		JC/MD	Ok
5	VX0416WBS01	VX0416WBS01	VX045800.D	16 Apr 2025 12:11		JC/MD	Ok,M
6	VX0416WBSD01	VX0416WBSD01	VX045801.D	16 Apr 2025 12:39		JC/MD	Ok,M
7	PB167604TB	PB167604TB	VX045802.D	16 Apr 2025 13:02		JC/MD	Ok
8	PB167604ZHE#61	PB167604ZHE#61	VX045803.D	16 Apr 2025 13:25		JC/MD	Ok,M
9	PB167604ZHE#62	PB167604ZHE#62	VX045804.D	16 Apr 2025 13:48		JC/MD	Ok
10	PB167604ZHE#63	PB167604ZHE#63	VX045805.D	16 Apr 2025 14:11		JC/MD	Ok,M
11	PB167604ZHE#64	PB167604ZHE#64	VX045806.D	16 Apr 2025 14:35		JC/MD	Ok
12	PB167604ZHE#65	PB167604ZHE#65	VX045807.D	16 Apr 2025 14:58		JC/MD	Ok
13	Q1800-01	WC-A4-01-G	VX045808.D	16 Apr 2025 15:21	vial A pH#5.0	JC/MD	Ok
14	Q1803-01	WEST-BAY	VX045809.D	16 Apr 2025 15:44	vial A pH#5.0	JC/MD	Ok
15	Q1803-02	FUEL-MONITORING	VX045810.D	16 Apr 2025 16:08	vial A pH#5.0	JC/MD	Ok
16	Q1808-02	OILY-SOIL-PILE	VX045811.D	16 Apr 2025 16:31	vial A pH#5.0	JC/MD	Ok,M
17	Q1808-04	LAW-25-0060	VX045812.D	16 Apr 2025 16:54	vial A pH#5.0	JC/MD	Ok,M
18	IBLK	IBLK	VX045813.D	16 Apr 2025 17:17		JC/MD	Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX041625

Review By	John Carlone	Review On	4/17/2025 9:03:49 AM
Supervise By	Maresh Dadoda	Supervise On	4/17/2025 1:41:16 PM
SubDirectory	VX041625	HP Acquire Method	HP Processing Method 82X040225W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133677
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133678,VP133679

19	Q1822-01	TW-WTS-06	VX045814.D	16 Apr 2025 17:41	vial A pH<2	JC/MD	Ok
20	VSTDCCC050	VSTDCCC050EC	VX045815.D	16 Apr 2025 18:04		JC/MD	Ok,M

M : Manual Integration



SHIPPING DOCUMENTS

CHAIN OF CUSTODY RECORD

Alliance Project Number:

Q1822

COC Number:

Page 1 of 1

CLIENT INFORMATION

COMPANY: ENTACT, LLC
ADDRESS: 150 Bay Street, Suite 806
CITY: Jersey City STATE: NJ ZIP: 07302
ATTENTION: Jarod Stanfield
PHONE: 570-886-0442 FAX:

PROJECT INFORMATION

PROJECT NAME: 540 Degraw St Brooklyn, NY
PROJECT #: E9309 LOCATION: Brooklyn, NY
PROJECT MANAGER: Jarod Stanfield
E-MAIL: jstanfield@entact.com
PHONE: 570-886-0442 FAX:

BILLING INFORMATION

BILL TO: ENTACT, LLC PO# E9309
ADDRESS: 999 Oakmont Plaza Drive, Suite 300
CITY: Westmont STATE: IL ZIP: 60559
ATTENTION: Wendy Murray PHONE: 800-936-8228

DATA TURNAROUND INFORMATION

FAX: 5 DAYS*
HARD COPY: 5 DAYS*
EDD 5 DAYS*
* TO BE APPROVED BY ALLIANCE
STANDARD TURNAROUND TIME IS 10 BUSINESS DAYS

DATA DELIVERABLE INFORMATION

- ☐ RESULTS ONLY ☐ USEPA CLP
☐ RESULTS + QC ☐ New York State ASP "B"
☐ New Jersey REDUCED ☐ New York State ASP "A"
☐ New Jersey CLP ☐ Other _____
☐ EDD Format _____

ANALYSIS

SVOC-TCL BNA-20	Flash Point	PCB	BOD5	TSS	VOC-TCLVOA- 10	Metals ICP-TAL			
1	2	3	4	5	6	7	8	9	

PRESERVATIVES

E	E	E	E	E	A	B			
1	2	3	4	5	6	7	8	9	

COMMENTS

<-- Specify Preservatives
A-HCl B-HNO3
C-H2SO4 D-NaOH
E-ICE F-Other

CHEMTECH
SAMPLE
ID

PROJECT
SAMPLE IDENTIFICATION

SAMPLE
MATRIX

SAMPLE
TYPE
COMP GRAB

SAMPLE
COLLECTION
DATE TIME

of Bottles

1.	TW-WTS-06	Surface Water		X	4/15	15:30	7
2.							
3.							
4.							
5.							
6.							
7.							
8.							
9.							
10.							

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE PROSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER
1. Jarod Stanfield

DATE/TIME
04/15/25
15:30

RECEIVED BY

1.  1300 4-15-25

RELINQUISHED BY

DATE/TIME

RECEIVED BY

2.

DATE/TIME

RECEIVED BY

3.

DATE/TIME

RECEIVED FOR LAB BY

1445 4-16-25

Conditions of bottles or coolers at receipt: ☐ Compliant ☐ Non Compliant ☐ Cooler Temp 4.6° ☐ Ice in Cooler? _____

Comments:

Temp 4.6° C Adjustment factor +1.0 IR Gun #1

Page _____ of _____

SHIPPED VIA: CLIENT: ☐ Hand Delivered ☐ Overnight
ALLIANCE: ☐ Picked Up ☐ Overnight

Shipment Complete

☐ YES ☐ NO

WHITE - ALLIANCE COPY FOR RETURN TO CLIENT

YELLOW - ALLIANCE COPY

PINK - SAMPLER COPY

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q1822	ENTA05	Order Date : 4/16/2025 1:17:00 PM	Project Mgr :
Client Name : ENTACT		Project Name : 540 Degraw St, Brooklyn, N	Report Type : Level 1
Client Contact : Jarod Stanfield		Receive DateTime : 4/16/2025 12:00:00 AM	EDD Type : Excel NJ
Invoice Name : ENTACT		Purchase Order : 14:48	Hard Copy Date :
Invoice Contact : Jarod Stanfield			Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q1822-01	TW-WTS-06	Water	04/15/2025	15:30	VOCMS Group4		8260-Low		5 Bus. Days

Relinquished By :

Date / Time : 4-16-25 1500

Received By :

Date / Time :

Storage Area : VOA Refridgerator Room