

Surrogate Summary

SDG No.: Q1822

Client: ENTACT

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1822-01	TW-WTS-06	1,2-Dichloroethane-d4	50	54.3	109	70 (74)	130 (125)
		Dibromofluoromethane	50	50.3	101	70 (75)	130 (124)
		Toluene-d8	50	50.3	101	70 (86)	130 (113)
		4-Bromofluorobenzene	50	50.4	101	70 (77)	130 (121)
VX0416WBL01	VX0416WBL01	1,2-Dichloroethane-d4	50	53.0	106	70 (74)	130 (125)
		Dibromofluoromethane	50	50.5	101	70 (75)	130 (124)
		Toluene-d8	50	49.6	99	70 (86)	130 (113)
		4-Bromofluorobenzene	50	51.9	104	70 (77)	130 (121)
VX0416WBS01	VX0416WBS01	1,2-Dichloroethane-d4	50	53.5	107	70 (74)	130 (125)
		Dibromofluoromethane	50	52.4	105	70 (75)	130 (124)
		Toluene-d8	50	51.6	103	70 (86)	130 (113)
		4-Bromofluorobenzene	50	54.3	108	70 (77)	130 (121)
VX0416WBSD0	VX0416WBSD01	1,2-Dichloroethane-d4	50	54.1	108	70 (74)	130 (125)
		Dibromofluoromethane	50	53.5	107	70 (75)	130 (124)
		Toluene-d8	50	51.1	102	70 (86)	130 (113)
		4-Bromofluorobenzene	50	54.4	109	70 (77)	130 (121)

