

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1822

Client: ENTACT

Analytical Method: SW8260-Low

Datafile : VX045801.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0416WBSD01	Methyl tert-butyl Ether	20	21.0	ug/L	105	3		70 (78)	130 (114)	20 (20)
	Carbon Tetrachloride	20	19.8	ug/L	99	1		70 (77)	130 (113)	20 (20)
	Chloroform	20	20.0	ug/L	100	1		70 (79)	130 (113)	20 (20)
	1,1,1-Trichloroethane	20	19.9	ug/L	100	1		70 (80)	130 (108)	20 (20)
	Benzene	20	19.1	ug/L	96	1		70 (82)	130 (109)	20 (20)
	Toluene	20	19.3	ug/L	97	1		70 (82)	130 (110)	20 (20)
	Tetrachloroethene	20	19.4	ug/L	97	1		70 (67)	130 (123)	20 (20)
	Ethyl Benzene	20	19.8	ug/L	99	1		70 (83)	130 (109)	20 (20)
	m/p-Xylenes	40	40.0	ug/L	100	2		70 (82)	130 (110)	20 (20)
	o-Xylene	20	19.9	ug/L	100	0		70 (83)	130 (109)	20 (20)