

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ENTA05
 Lab Code: CHEM Case No.: Q1822 SAS No.: Q1822 SDG NO.: Q1822
 Lab File ID: VX045797.D Date Analyzed: 04/16/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:28
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	66178	12.018				
UPPER LIMIT	132356	12.518				
LOWER LIMIT	33089	11.518				
EPA SAMPLE NO.						
TW-WTS-06	52878	12.02				
VX0416WBL01	56266	12.02				
VX0416WBS01	66591	12.02				
VX0416WBSD01	65216	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.