

Initial and Continuing Calibration Verification

Client: Weston Solutions

SDG No.: Q1827

Project: Joint Base McGuire Dix Lakehurst JBMDL - DoD

RunNo.: LB135495

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV1						
TOC		mg/L	1030	1000	103	90-110	03/14/2025
Sample ID:	CCV1						
TOC		mg/L	998	1000	100	90-110	04/21/2025
Sample ID:	CCV2						
TOC		mg/L	990	1000	99	90-110	04/21/2025
Sample ID:	CCV3						
TOC		mg/L	1030	1000	103	90-110	04/21/2025
Sample ID:	CCV4						
TOC		mg/L	962	1000	96	90-110	04/22/2025
Sample ID:	CCV5						
TOC		mg/L	968	1000	97	90-110	04/22/2025