

Initial and Continuing Calibration Blank Summary

Client: Weston Solutions

SDG No.: Q1827

Project: Joint Base McGuire Dix Lakehurst JBMDL - DoD

RunNo.: LB135495

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 TOC	mg/L	< 125.0000	125.0000	U	32.1	250	03/14/2025
Sample ID: CCB1 TOC	mg/L	< 125.0000	125.0000	U	32.1	250	04/21/2025
Sample ID: CCB2 TOC	mg/L	< 125.0000	125.0000	U	32.1	250	04/21/2025
Sample ID: CCB3 TOC	mg/L	< 125.0000	125.0000	U	32.1	250	04/21/2025
Sample ID: CCB4 TOC	mg/L	< 125.0000	125.0000	U	32.1	250	04/22/2025
Sample ID: CCB5 TOC	mg/L	< 125.0000	125.0000	U	32.1	250	04/22/2025