

Matrix Spike Summary

Client:	Weston Solutions	SDG No.:	Q1827
Project:	Joint Base McGuire Dix Lakehurst JBMDL - DoD	Sample ID:	Q1827-04
Client ID:	SS050P-SD41-041525-00MS	Percent Solids for Spike Sample:	73.3

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
TOC	mg/Kg	75-125	11700	OR	10900		1000	1	80		04/21/2025

Matrix Spike Summary

Client:	Weston Solutions	SDG No.:	Q1827
Project:	Joint Base McGuire Dix Lakehurst JBMDL - DoD	Sample ID:	Q1827-04
Client ID:	SS050P-SD41-041525-00MSD	Percent Solids for Spike Sample:	73.3

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
TOC	mg/Kg	75-125	12100	OR	10900		1000	1	120		04/21/2025