

Analytical Summary Report

Analysis Method: SM2120 B
Parameter: Color
Run Number: LB135481

Reviewed By: Iwona
Supervisor Review By: jignesh
pH Meter ID: WC pH Meter-1

Reagent/Standard	Lot/Log #
Calibration std, 0 color unit	WP112741
Calibration std, 5 color unit	WP112742
Calibration std, 10 color unit	WP112743
Calibration std, 15 color unit	WP112752
Calibration std, 20 color unit	WP112744
Calibration std, 25 color unit	WP112745
Calibration std, 30 color unit	WP112746
Calibration std, 40 color unit	WP112747
Calibration std, 50 color unit	WP112748
Calibration std, 60 color unit	WP112749
Calibration std, 70 color unit	WP112750
Calibration std, 100 color unit	WP112751

Seq	LabID	ClientID	DL	pH Reading	pH Adjustment	Result (cu)	Anal Date	Anal Time
1	LB135481BL	LB135481BL	1	5.03	N/A	0.00	04/18/2025	09:35
2	Q1831-01	EFFLUENT-COMPOSITE	20	7.93	N/A	40.00	04/18/2025	09:40
3	Q1831-01DUP	EFFLUENT-COMPOSITEDU	20	7.92	N/A	40.00	04/18/2025	09:45

WORKLIST(Hardcopy Internal Chain)

LB135481

WorkList Name : COLOR-041825

WorkList ID : 189021

Department : Wet-Chemistry

Date : 04-18-2025 09:16:56

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1831-01	EFFLUENT-COMPOSITE	Water	Color	Cool 4 deg C	M&MMO1	L31	04/17/2025	SM2120 B

Date/Time 04/18/25 09:25
Raw Sample Received by: 12 (WC)
Raw Sample Relinquished by: HCCOC

Date/Time 04/18/25 10:00
Raw Sample Received by: HCCOC
Raw Sample Relinquished by: 12 (WC)