

Initial and Continuing Calibration Verification

Client: Mars Chocolate North America, LLC

SDG No.: Q1831

Project: W.T.P. Final Effluent 2025

RunNo.: LB135482

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Ammonia as N	mg/L	0.98	1	98	90-110	04/18/2025
Sample ID: CCV1 Ammonia as N	mg/L	0.96	1	96	90-110	04/18/2025
Sample ID: CCV2 Ammonia as N	mg/L	0.95	1	95	90-110	04/18/2025
Sample ID: CCV3 Ammonia as N	mg/L	1	1	100	90-110	04/18/2025
Sample ID: CCV4 Ammonia as N	mg/L	1	1	100	90-110	04/18/2025

Initial and Continuing Calibration Verification

Client: Mars Chocolate North America, LLC

SDG No.: Q1831

Project: W.T.P. Final Effluent 2025

RunNo.: LB135487

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Phosphorus, Total	mg/L	0.487	0.50	97	90-110	04/18/2025
Sample ID: CCV1 Phosphorus, Total	mg/L	0.500	0.50	100	90-110	04/18/2025
Sample ID: CCV2 Phosphorus, Total	mg/L	0.493	0.50	99	90-110	04/18/2025
Sample ID: CCV3 Phosphorus, Total	mg/L	0.499	0.50	100	90-110	04/18/2025

Initial and Continuing Calibration Verification

Client: Mars Chocolate North America, LLC

SDG No.: Q1831

Project: W.T.P. Final Effluent 2025

RunNo.: LB135532

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV COD	mg/L	49.329	50	99	95-105	01/22/2025
Sample ID: CCV1 COD	mg/L	50.319	50	101	95-105	04/21/2025
Sample ID: CCV2 COD	mg/L	49.329	50	99	95-105	04/21/2025

Initial and Continuing Calibration Verification

Client: Mars Chocolate North America, LLC

SDG No.: Q1831

Project: W.T.P. Final Effluent 2025

RunNo.: LB135532

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
---------	-------	--------	------------	---------------	---------------------------	------------------