

Matrix Spike Summary

Client:	Ardmore Chemical	SDG No.:	Q1833
Project:	PVSC Monthly 2025	Sample ID:	Q1833-01
Client ID:	EFF-WASTE-WATERMS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Cyanide	mg/L	75-125	0.035		0.0014	J	0.04	1	84		04/21/2025

Matrix Spike Summary

Client:	Ardmore Chemical	SDG No.:	Q1833
Project:	PVSC Monthly 2025	Sample ID:	Q1833-01
Client ID:	EFF-WASTE-WATERMSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Cyanide	mg/L	75-125	0.037		0.0014	J	0.04	1	89		04/21/2025