



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	Ardmore Chemical				SDG No.:	Q1833			
Contract:	ARDM01	Lab Code:	CHEM		Case No.:	Q1833	SAS No.:	Q1833	
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB97	Mercury	0.20	+/-0.20	U	0.20	CV	04/22/2025	14:02	LB135518

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Ardmore Chemical

SDG No.: Q1833

Contract: ARDM01

Lab Code: CHEM

Case No.: Q1833

SAS No.: Q1833

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB12	Mercury	0.20	+/-0.20	U	0.20	CV	04/22/2025	14:07	LB135518
CCB13	Mercury	0.20	+/-0.20	U	0.20	CV	04/22/2025	14:36	LB135518
CCB14	Mercury	0.20	+/-0.20	U	0.20	CV	04/22/2025	14:45	LB135518

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	Ardmore Chemical				SDG No.:	Q1833			
Contract:	ARDM01	Lab Code:	CHEM		Case No.:	Q1833		SAS No.:	Q1833
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Lead	12.0	+/-12.0	U	12.0	P	04/22/2025	13:00	LB135523
	Zinc	40.0	+/-40.0	U	40.0	P	04/22/2025	13:00	LB135523

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Ardmore Chemical</u>		SDG No.: <u>Q1833</u>							
Contract: <u>ARDM01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1833</u>	SAS No.: <u>Q1833</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Lead	12.0	+/-12.0	U	12.0	P	04/22/2025	13:30	LB135523
	Zinc	40.0	+/-40.0	U	40.0	P	04/22/2025	13:30	LB135523
CCB02	Lead	12.0	+/-12.0	U	12.0	P	04/22/2025	14:20	LB135523
	Zinc	40.0	+/-40.0	U	40.0	P	04/22/2025	14:20	LB135523
CCB03	Lead	12.0	+/-12.0	U	12.0	P	04/22/2025	15:12	LB135523
	Zinc	40.0	+/-40.0	U	40.0	P	04/22/2025	15:12	LB135523
CCB04	Lead	12.0	+/-12.0	U	12.0	P	04/22/2025	16:03	LB135523
	Zinc	40.0	+/-40.0	U	40.0	P	04/22/2025	16:03	LB135523
CCB05	Lead	12.0	+/-12.0	U	12.0	P	04/22/2025	17:02	LB135523
	Zinc	40.0	+/-40.0	U	40.0	P	04/22/2025	17:02	LB135523
CCB06	Lead	12.0	+/-12.0	U	12.0	P	04/22/2025	17:55	LB135523
	Zinc	40.0	+/-40.0	U	40.0	P	04/22/2025	17:55	LB135523
CCB07	Lead	12.0	+/-12.0	U	12.0	P	04/22/2025	19:00	LB135523
	Zinc	40.0	+/-40.0	U	40.0	P	04/22/2025	19:00	LB135523
CCB08	Lead	12.0	+/-12.0	U	12.0	P	04/22/2025	19:46	LB135523
	Zinc	40.0	+/-40.0	U	40.0	P	04/22/2025	19:46	LB135523

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	<u>Ardmore Chemical</u>				SDG No.:	<u>Q1833</u>			
Contract:	<u>ARDM01</u>	Lab Code:	<u>CHEM</u>	Case No.:	<u>Q1833</u>	SAS No.:	<u>Q1833</u>		

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
-----------	---------	----------------	---------------------	--------------	------	---	------------------	------------------	---------------
