

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF042125

Review By	Rahul	Review On	4/22/2025 10:39:20 AM
Supervise By		Supervise On	
SubDirectory	BF042125	HP Acquire Method	BNA_F
		HP Processing Method	BF042125

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12660,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF142198.D	21 Apr 2025 10:43		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BF142199.D	21 Apr 2025 11:13		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BF142200.D	21 Apr 2025 11:41	Compounds #32,54,56,65,85 removed from 5ppm	RC/JU	Ok,NS
4	SSTDICC010	SSTDICC010	BF142201.D	21 Apr 2025 12:09	Compound #32 kept on LR	RC/JU	Ok,NS
5	SSTDICC020	SSTDICC020	BF142202.D	21 Apr 2025 12:37	Com#69,77 failed in this calibration	RC/JU	Ok
6	SSTDICCC040	SSTDICCC040	BF142203.D	21 Apr 2025 13:06	This calibration is good for 8270E DOD except Com#69,77 and good for 625.1 method	RC/JU	Ok
7	SSTDICC050	SSTDICC050	BF142204.D	21 Apr 2025 13:33	This calibration is used for 625.1 method	RC/JU	Ok
8	SSTDICC060	SSTDICC060	BF142205.D	21 Apr 2025 14:02	Compound #69 Removed from 60PPM	RC/JU	Ok
9	SSTDICC080	SSTDICC080	BF142206.D	21 Apr 2025 14:30	Compound #45,69,79 Removed from 80PPM	RC/JU	Ok
10	SSTDICV040	ICVBF042125	BF142207.D	21 Apr 2025 18:12		RC/JU	Ok
11	PB167645BL	PB167645BL	BF142208.D	21 Apr 2025 19:09	Method is Fail for Com# 69, Analyzed for contamination check	RC/JU	Not Ok

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF042225

Review By	Rahul	Review On	4/23/2025 11:57:50 AM
Supervise By		Supervise On	
SubDirectory	BF042225	HP Acquire Method	BNA_F
		HP Processing Method	BF042125
STD. NAME	STD REF.#		
Tune/Reschk	SP6757		
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729		
CCC	SP6725		
Internal Standard/PEM	S12660,10ul/1000ul sample		
ICV/I.BLK	SP6770		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF142209.D	22 Apr 2025 08:26		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF142210.D	22 Apr 2025 14:11		RC/JU	Ok
3	PB167672BS	PB167672BS	BF142211.D	22 Apr 2025 14:44		RC/JU	Ok,NS
4	PB167672BSD	PB167672BSD	BF142212.D	22 Apr 2025 15:13		RC/JU	Ok,NS
5	PB167672BL	PB167672BL	BF142213.D	22 Apr 2025 15:41		RC/JU	Ok
6	Q1833-02	EFF-WASTE-WATER	BF142214.D	22 Apr 2025 16:21	Need 10X dilution	RC/JU	Dilution
7	Q1833-02DL	EFF-WASTE-WATERD	BF142215.D	22 Apr 2025 16:50		RC/JU	Ok

M : Manual Integration