

Instrument ID: **BNA_F**

Daily Analysis Runlog For Sequence/QC Batch ID # BF042125

Review By	Rahul	Review On	4/22/2025 10:39:20 AM
Supervise By		Supervise On	
SubDirectory	BF042125	HP Acquire Method	BNA_F
		HP Processing Method	BF042125
STD. NAME	STD REF.#		
Tune/Reschk	SP6757		
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729		
CCC	SP6725		
Internal Standard/PEM	S12660,10ul/1000ul sample		
ICV/I.BLK	SP6770		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BF142198.D	21 Apr 2025 10:43	RC/JU	Ok
2	SSTDICC2.5	BF142199.D	21 Apr 2025 11:13	RC/JU	Ok
3	SSTDICC005	BF142200.D	21 Apr 2025 11:41	RC/JU	Ok,NS
4	SSTDICC010	BF142201.D	21 Apr 2025 12:09	RC/JU	Ok,NS
5	SSTDICC020	BF142202.D	21 Apr 2025 12:37	RC/JU	Ok
6	SSTDICCC040	BF142203.D	21 Apr 2025 13:06	RC/JU	Ok
7	SSTDICC050	BF142204.D	21 Apr 2025 13:33	RC/JU	Ok
8	SSTDICC060	BF142205.D	21 Apr 2025 14:02	RC/JU	Ok
9	SSTDICC080	BF142206.D	21 Apr 2025 14:30	RC/JU	Ok
10	SSTDICV040	BF142207.D	21 Apr 2025 18:12	RC/JU	Ok
11	PB167645BL	BF142208.D	21 Apr 2025 19:09	RC/JU	Not Ok

M : Manual Integration