

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance

Contract: ARDM01

Lab Code: ACE

SDG No.: Q1833

Instrument ID: BNA_F

Calibration Date(s): 04/21/2025 04/21/2025

Calibration Time(s): 11:13 14:30

LAB FILE ID:		RRF2.5 = BF142199.D			RRF005 = BF142200.D		RRF010 = BF142201.D	
		RRF020 = BF142202.D			RRF040 = BF142203.D		RRF050 = BF142204.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
n-Nitrosodimethylamine		0.356	0.367	0.412	0.382	0.395	0.382	4.9
2-Fluorophenol		1.110	1.107	1.166	1.048	1.044	1.067	6.3
Phenol-d6		1.396	1.432	1.457	1.293	1.293	1.336	7.1
Phenol		1.457	1.483	1.532	1.383	1.372	1.409	6.4
bis(2-Chloroethyl)ether		1.097	1.094	1.128	1.018	1.043	1.063	4.4
2-Chlorophenol		1.254	1.271	1.335	1.207	1.196	1.221	6.1
2,2-oxybis(1-Chloropropane)		1.097	1.082	1.108	0.998	1.011	1.030	6.7
n-Nitroso-di-n-propylamine	0.779	0.841	0.807	0.791	0.713	0.721	0.755	7.5
Nitrobenzene-d5		0.324	0.333	0.344	0.313	0.308	0.318	5.1
Hexachloroethane		0.509	0.502	0.518	0.465	0.467	0.482	5.8
Nitrobenzene		0.318	0.319	0.336	0.316	0.310	0.315	3.6
Isophorone		0.549	0.541	0.562	0.523	0.522	0.535	3.1
2-Nitrophenol		0.134	0.151	0.169	0.172	0.174	0.166	10.5
2,4-Dimethylphenol		0.201	0.206	0.225	0.217	0.220	0.216	4.3
bis(2-Chloroethoxy)methane		0.358	0.364	0.372	0.345	0.344	0.351	4.2
2,4-Dichlorophenol		0.281	0.292	0.309	0.295	0.296	0.293	3.1
1,2,4-Trichlorobenzene		0.362	0.350	0.365	0.336	0.333	0.343	5.0
Naphthalene		1.079	1.063	1.070	0.917	0.882	0.945	13.6
Hexachlorobutadiene		0.232	0.231	0.242	0.224	0.218	0.226	4.1
4-Chloro-3-methylphenol		0.292	0.297	0.307	0.294	0.292	0.295	2.3
Hexachlorocyclopentadiene		0.128	0.161	0.214	0.226	0.216	0.201	20.0
2,4,6-Trichlorophenol		0.337	0.367	0.409	0.389	0.389	0.382	6.2
2-Fluorobiphenyl		1.486	1.417	1.372	1.101	1.014	1.227	18.2
2-Chloronaphthalene		1.225	1.176	1.228	1.082	1.046	1.113	8.8
Dimethylphthalate		1.352	1.322	1.401	1.250	1.218	1.282	6.2
Acenaphthylene		1.790	1.701	1.771	1.530	1.474	1.573	11.7
2,6-Dinitrotoluene		0.274	0.283	0.309	0.291	0.286	0.288	3.9
Acenaphthene		1.230	1.168	1.233	1.099	1.082	1.134	7.0
2,4-Dinitrophenol			0.094	0.131	0.143	0.153	0.141	18.6
4-Nitrophenol			0.156	0.206	0.205	0.216	0.202	11.8
2,4-Dinitrotoluene		0.340	0.367	0.398	0.376	0.375	0.372	5.4
Diethylphthalate		1.257	1.266	1.327	1.187	1.155	1.197	8.0
4-Chlorophenyl-phenylether		0.737	0.710	0.731	0.661	0.639	0.677	7.3
Fluorene		1.381	1.339	1.363	1.187	1.156	1.230	10.7

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance

Contract: ARDM01

Lab Code: ACE

SDG No.: Q1833

Instrument ID: BNA_F

Calibration Date(s): 04/21/2025 04/21/2025

Calibration Time(s): 11:13 14:30

LAB FILE ID:		RRF2.5 = BF142199.D			RRF005 = BF142200.D		RRF010 = BF142201.D	
		RRF020 = BF142202.D			RRF040 = BF142203.D		RRF050 = BF142204.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
4,6-Dinitro-2-methylphenol			0.088	0.110	0.117	0.123	0.117	13.9
n-Nitrosodiphenylamine		0.651	0.640	0.648	0.594	0.587	0.613	5.3
Azobenzene		1.147	1.110	1.139	1.018	1.002	1.043	9.0
2,4,6-Tribromophenol		0.246	0.265	0.289	0.274	0.280	0.273	5.4
4-Bromophenyl-phenylether		0.261	0.262	0.272	0.259	0.257	0.263	1.9
Hexachlorobenzene		0.323	0.318	0.326	0.304	0.304	0.313	2.9
Pentachlorophenol		0.100	0.136	0.165	0.169	0.178	0.160	19.5
Phenanthrene		1.127	1.120	1.110	0.952	0.916	0.993	12.5
Anthracene		1.114	1.102	1.104	0.967	0.926	0.993	11.5
Di-n-butylphthalate		0.862	0.981	1.038	0.909	0.881	0.898	10.1
Fluoranthene		1.176	1.190	1.180	0.991	0.955	1.032	14.5
Benzidine		0.230	0.196	0.166	0.385	0.301	0.258	28.3
Pyrene		1.783	1.726	1.849	1.592	1.501	1.588	13.5
Terphenyl-d14		1.390	1.331	1.382	1.106	1.015	1.194	16.6
Butylbenzylphthalate		0.258	0.308	0.403	0.402	0.433	0.384	19.0
3,3-Dichlorobenzidine		0.272	0.329	0.372	0.370	0.386	0.361	12.9
Benzo (a) anthracene		1.197	1.211	1.357	1.207	1.199	1.211	5.9
Chrysene		1.259	1.232	1.229	1.133	1.119	1.167	6.1
Bis(2-ethylhexyl)phthalate		0.351	0.405	0.513	0.539	0.565	0.514	19.7
Di-n-octyl phthalate			0.642	0.831	0.934	0.972	0.911	16.9
Benzo (b) fluoranthene		0.993	1.065	1.303	1.134	1.138	1.122	8.5
Benzo (k) fluoranthene		1.273	1.226	1.114	1.055	1.060	1.108	10.5
Benzo (a) pyrene		0.989	0.999	1.069	0.990	1.002	1.004	3.2
Indeno (1,2,3-cd) pyrene		1.378	1.344	1.475	1.405	1.413	1.403	2.9
Dibenzo (a,h) anthracene		1.116	1.084	1.215	1.164	1.158	1.145	3.7
Benzo (g,h,i) perylene		1.176	1.139	1.239	1.203	1.201	1.192	2.7

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1