

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: ARDM01
Lab Code: CHEM Case No.: Q1833 SAS No.: Q1833 SDG No.: Q1833
Instrument ID: MSVOA_N Calibration Date/Time: 04/23/2025 09:45
Lab File ID: VN086373.D Init. Calib. Date(s): 04/23/2025 04/23/2025
Heated Purge: (Y/N) N Init. Calib. Time(s): 09:22 13:17
GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
Acrolein	100.000	83.90	16.104	20