

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q1835

Project: Pre Treatment Plant 2025

Run No.: LB135485

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135485BS							
Orthophosphate as P	mg/L	0.5	0.48		97	1	90-110	04/18/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q1835

Project: Pre Treatment Plant 2025

Run No.: LB135493

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135493BS							
BOD5	mg/L	198	208		105	1	84.6-115.4	04/18/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q1835

Project: Pre Treatment Plant 2025

Run No.: LB135500

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135500BS							
TSS	mg/L	550	532		97	1	90-110	04/21/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q1835

Project: Pre Treatment Plant 2025

Run No.: LB135516

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135516BS							
Oil and Grease	mg/L	20.0	16.8		84	1	78-114	04/22/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q1835

Project: Pre Treatment Plant 2025

Run No.: LB135482

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB167618BS							
Ammonia as N	mg/L	1	0.95		95	1	90-110	04/18/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q1835

Project: Pre Treatment Plant 2025

Run No.: LB135487

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB167651BS							
Phosphorus, Total	mg/L	0.50	0.47		95	1	90-110	04/18/2025