

Initial and Continuing Calibration Verification

Client: Vartan Manufacturing Corp.

SDG No.: Q1836

Project: NYCDEP Permit No. 25-P3067-1

RunNo.: LB135574

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Cyanide	ICV1	mg/L	0.096	0.099	97	85-115	04/28/2025
Sample ID: Cyanide	CCV1	mg/L	0.25	0.25	100	90-110	04/28/2025
Sample ID: Cyanide	CCV2	mg/L	0.24	0.25	96	90-110	04/28/2025
Sample ID: Cyanide	CCV3	mg/L	0.26	0.25	104	90-110	04/28/2025