

Initial and Continuing Calibration Blank Summary

Client: Vartan Manufacturing Corp.

SDG No.: Q1836

Project: NYCDEP Permit No. 25-P3067-1

RunNo.: LB135574

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Cyanide	mg/L	< 0.0025	0.0025	U	0.0012	0.005	04/28/2025
Sample ID: CCB1 Cyanide	mg/L	< 0.0025	0.0025	U	0.0012	0.005	04/28/2025
Sample ID: CCB2 Cyanide	mg/L	0.0013	0.0025	J	0.0012	0.005	04/28/2025
Sample ID: CCB3 Cyanide	mg/L	0.0012	0.0025	J	0.0012	0.005	04/28/2025