

Matrix Spike Summary

Client:	Vartan Manufacturing Corp.	SDG No.:	Q1836
Project:	NYCDEP Permit No. 25-P3067-1	Sample ID:	Q1836-08
Client ID:	CYANIDE-4MS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Cyanide	mg/L	75-125	0.036		0.0012	U	0.04	1	90		04/28/2025

Matrix Spike Summary

Client:	Vartan Manufacturing Corp.	SDG No.:	Q1836
Project:	NYCDEP Permit No. 25-P3067-1	Sample ID:	Q1836-08
Client ID:	CYANIDE-4MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Cyanide	mg/L	75-125	0.037		0.0012	U	0.04	1	92		04/28/2025