

Laboratory Control Sample Summary

Client: Vartan Manufacturing Corp.

SDG No.: Q1836

Project: NYCDEP Permit No. 25-P3067-1

Run No.: LB135574

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB167752BS							
Cyanide	mg/L	0.1	0.095		95	1	85-115	04/28/2025