



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Vartan Manufacturing Corp. **SDG No.:** Q1836
Contract: F&GE01 **Lab Code:** CHEM **Case No.:** Q1836 **SAS No.:** Q1836
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.041	0.05	82	40 - 160	CV	04/30/2025	10:08	LB135601
CRI01	Cadmium	6.34	6.0	106	40 - 160	P	04/30/2025	16:41	LB135614
	Chromium	9.88	10.0	99	40 - 160	P	04/30/2025	16:41	LB135614
	Copper	23.2	20.0	116	40 - 160	P	04/30/2025	16:41	LB135614
	Lead	13.5	12.0	112	40 - 160	P	04/30/2025	16:41	LB135614
	Molybdenum	220	200	110	40 - 160	P	04/30/2025	16:41	LB135614
	Nickel	41.3	40.0	103	40 - 160	P	04/30/2025	16:41	LB135614
	Silver	10.8	10.0	108	40 - 160	P	04/30/2025	16:41	LB135614
	Zinc	42.8	40.0	107	40 - 160	P	04/30/2025	16:41	LB135614