

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1837 SAS No.: Q1837 SDG NO.: Q1837

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/09/2025 04/09/2025

Client Sample No.: CCAL01 Date Analyzed: 04/22/2025

Lab Sample No.: AR1660CCC400 Data File : PQ070221.D Time Analyzed: 10:17

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.484	4.384	4.584	209106200.000	209800000.000	0.3
Aroclor-1016-2	4.503	4.403	4.603	327268100.000	333931000.000	2.0
Aroclor-1016-3	4.561	4.461	4.661	205892100.000	213396000.000	3.6
Aroclor-1016-4	4.651	4.551	4.751	168530800.000	172401000.000	2.3
Aroclor-1016-5	4.935	4.835	5.035	164363500.000	169563000.000	3.2
Aroclor-1260-1	6.025	5.925	6.125	289545900.000	295051000.000	1.9
Aroclor-1260-2	6.281	6.180	6.380	336546600.000	338938000.000	0.7
Aroclor-1260-3	6.632	6.531	6.731	257191800.000	265666000.000	3.3
Aroclor-1260-4	6.846	6.745	6.945	284683900.000	293216000.000	3.0
Aroclor-1260-5	7.152	7.051	7.251	548455700.000	550642000.000	0.4
DCB	8.503	8.402	8.602	3392103000.000	3573900000.000	5.4
TCX	3.404	3.304	3.504	7302210000.000	7472250000.000	2.4

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1837 SAS No.: Q1837 SDG NO.: Q1837

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/09/2025 04/09/2025

Client Sample No.: CCAL01 Date Analyzed: 04/22/2025

Lab Sample No.: AR1660CCC400 Data File : PQ070221.D Time Analyzed: 10:17

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	3.784	3.685	3.885	220583700.000	219817000.000	-0.3
Aroclor-1016-2	3.800	3.700	3.900	333637400.000	341206000.000	2.3
Aroclor-1016-3	3.960	3.861	4.061	184921400.000	194424000.000	5.1
Aroclor-1016-4	4.008	3.909	4.109	161588800.000	170684000.000	5.6
Aroclor-1016-5	4.203	4.105	4.305	195663400.000	204099000.000	4.3
Aroclor-1260-1	5.197	5.098	5.298	354254700.000	371117000.000	4.8
Aroclor-1260-2	5.386	5.287	5.487	426945600.000	443939000.000	4.0
Aroclor-1260-3	5.528	5.429	5.629	402319200.000	412148000.000	2.4
Aroclor-1260-4	5.988	5.890	6.090	352395600.000	362852000.000	3.0
Aroclor-1260-5	6.232	6.134	6.334	796660000.000	826287000.000	3.7
DCB	7.544	7.446	7.646	554060000.000	569178000.000	2.7
TCX	2.782	2.683	2.883	7031452000.000	7377570000.000	4.9