

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1837 SAS No.: Q1837 SDG NO.: Q1837

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/09/2025 04/09/2025

Client Sample No.: CCAL02 Date Analyzed: 04/22/2025

Lab Sample No.: AR1660CCC400 Data File : PQ070225.D Time Analyzed: 12:31

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.485	4.384	4.584	209106200.000	208512000.000	-0.3
Aroclor-1016-2	4.504	4.403	4.603	327268100.000	330383000.000	1.0
Aroclor-1016-3	4.561	4.461	4.661	205892100.000	208499000.000	1.3
Aroclor-1016-4	4.651	4.551	4.751	168530800.000	170659000.000	1.3
Aroclor-1016-5	4.935	4.835	5.035	164363500.000	165737000.000	0.8
Aroclor-1260-1	6.025	5.925	6.125	289545900.000	291130000.000	0.5
Aroclor-1260-2	6.281	6.180	6.380	336546600.000	338292000.000	0.5
Aroclor-1260-3	6.632	6.531	6.731	257191800.000	262038000.000	1.9
Aroclor-1260-4	6.846	6.745	6.945	284683900.000	289915000.000	1.8
Aroclor-1260-5	7.152	7.051	7.251	548455700.000	555183000.000	1.2
DCB	8.503	8.402	8.602	3392103000.000	3569510000.000	5.2
TCX	3.404	3.304	3.504	7302210000.000	7378300000.000	1.1

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1837 SAS No.: Q1837 SDG NO.: Q1837

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/09/2025 04/09/2025

Client Sample No.: CCAL02 Date Analyzed: 04/22/2025

Lab Sample No.: AR1660CCC400 Data File : PQ070225.D Time Analyzed: 12:31

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	3.783	3.685	3.885	220583700.000	218223000.000	-1.1
Aroclor-1016-2	3.799	3.700	3.900	333637400.000	334793000.000	0.3
Aroclor-1016-3	3.960	3.861	4.061	184921400.000	188798000.000	2.1
Aroclor-1016-4	4.008	3.909	4.109	161588800.000	165260000.000	2.3
Aroclor-1016-5	4.203	4.105	4.305	195663400.000	199664000.000	2.0
Aroclor-1260-1	5.197	5.098	5.298	354254700.000	357664000.000	1.0
Aroclor-1260-2	5.386	5.287	5.487	426945600.000	428407000.000	0.3
Aroclor-1260-3	5.527	5.429	5.629	402319200.000	402124000.000	0.0
Aroclor-1260-4	5.988	5.890	6.090	352395600.000	358499000.000	1.7
Aroclor-1260-5	6.233	6.134	6.334	796660000.000	789675000.000	-0.9
DCB	7.545	7.446	7.646	554060000.000	565919000.000	2.2
TCX	2.782	2.683	2.883	7031452000.000	7174390000.000	2.1