

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: Chem02

Lab Code: CHEM Case No.: Q1837 SAS No.: Q1837 SDG No.:

Instrument ID: MSVOA_X Calibration Date/Time: 04/23/2025 08:56

Lab File ID: VX045899.D Init. Calib. Date(s): 04/01/2025 04/01/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 17:06 19:02

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
t-1,3-Dichloropropene	50.000	53.53	-7.068001	20