

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: Chem02
 Lab Code: CHEM Case No.: Q1837 SAS No.: Q1837 SDG NO.:
 Lab File ID: VX045899.D Date Analyzed: 04/23/2025
 Instrument ID: MSVOA_X Time Analyzed: 08:56
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	67565	12.018				
UPPER LIMIT	135130	12.518				
LOWER LIMIT	33782.5	11.518				
EPA SAMPLE NO.						
STORAGE-BLANK-SOIL-REF-6	50851	12.02				
STORAGE-BLANK-WATER-REF-5	52258	12.02				
STORAGE-BLANK-WATER-REF-4	44728	12.02				
VX0423WBL01	56196	12.02				
VX0423WBS01	70247	12.02				
VX0423WBSD01	61538	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.