



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: Weston Solutions

SDG No.: Q1838

Project: Joint Base McGuire Dix Lakehurst JBMDL - DoD

RunNo.: LB135539

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: TOC	ICV1	mg/L	1030	1000	103	90-110	03/14/2025
Sample ID: TOC	CCV1	mg/L	951	1000	95	90-110	04/24/2025
Sample ID: TOC	CCV2	mg/L	982	1000	98	90-110	04/24/2025
Sample ID: TOC	CCV3	mg/L	1010	1000	101	90-110	04/24/2025