

Duplicate Sample Summary

Client: Weston Solutions	SDG No.: Q1838
Project: Joint Base McGuire Dix Lakehurst JBMDL - DoD	Sample ID: Q1838-06
Client ID: SS050P-SD33-041625-00MSD	Percent Solids for Spike Sample: 78.4

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
TOC	mg/Kg	+/-20	7090		7310		1	3		04/24/2025