

Initial and Continuing Calibration Verification

Client: Weston Solutions

SDG No.: Q1839

Project: Joint Base McGuire Dix Lakehurst JBMDL - DoD

RunNo.: LB135517

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: TOC	ICV1	mg/L	1030	1000	103	90-110	03/14/2025
Sample ID: TOC	CCV1	mg/L	1060	1000	106	90-110	04/23/2025
Sample ID: TOC	CCV2	mg/L	1030	1000	103	90-110	04/23/2025
Sample ID: TOC	CCV3	mg/L	962	1000	96	90-110	04/23/2025
Sample ID: TOC	CCV4	mg/L	1020	1000	102	90-110	04/23/2025