

Report of Analysis

Client:	GFE LLC	Date Collected:	
Project:	82-64 Austin St, Kew Gardens NY	Date Received:	
Client Sample ID:	VL0421ABS01	SDG No.:	Q1843
Lab Sample ID:	VL0421ABS01	Matrix:	Air
Analytical Method:	TO-15	Test:	VOCMS Group2
Sample Wt/Vol:	400	Units:	mL

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL042397.D	1		04/21/25 10:14	VL042125

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
TARGETS							
75-01-4	Vinyl Chloride	8.60	22.0		0.030	0.080	ug/m3
75-34-3	1,1-Dichloroethane	9.60	38.9		0.53	2.02	ug/m3
156-59-2	cis-1,2-Dichloroethene	9.10	36.1		0.36	1.98	ug/m3
71-55-6	1,1,1-Trichloroethane	9.30	50.7		0.050	0.16	ug/m3
79-01-6	Trichloroethene	7.80	41.9		0.11	0.16	ug/m3
127-18-4	Tetrachloroethene	7.40	50.2		0.14	0.20	ug/m3
SURROGATES							
460-00-4	1-Bromo-4-Fluorobenzene	10.8			65 - 135	108%	SPK: 10
INTERNAL STANDARDS							
74-97-5	Bromochloromethane	118000		2.781			
540-36-3	1,4-Difluorobenzene	338000		3.949			
3114-55-4	Chlorobenzene-d5	294000		8.875			

U = Not Detected

RL = Reporting Limit

MDL = Method Detection Limit

E = Value Exceeds Calibration Range

D = Dilution

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

Q = indicates LCS control criteria did not meet requirements