

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1843

Client: GFE LLC

Analytical Method: SWTO-15

Datafile : VL042397.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VL0421ABS01	Vinyl Chloride	10	8.60	ppbv	86			70	130	
	1,1-Dichloroethane	10	9.60	ppbv	96			70	130	
	cis-1,2-Dichloroethene	10	9.10	ppbv	91			70	130	
	1,1,1-Trichloroethane	10	9.30	ppbv	93			70	130	
	Trichloroethene	10	7.80	ppbv	78			70	130	
	Tetrachloroethene	10	7.40	ppbv	74			70	130	