

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: GFEL01

Lab Code: CHEM Case No.: Q1844 SAS No.: Q1844 SDG No.: Q1844

Instrument ID: MSVOA_L Calibration Date/Time: 04/21/2025 08:41

Lab File ID: VL042395.D Init. Calib. Date(s): 04/17/2025 04/17/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 09:44 12:54

GC Column: RTX-1 ID: 0.32 (mm)

COMPOUND	RRF	RRF010	MIN RRF	%D	MAX%D
Vinyl Chloride	0.761	0.665		-12.61	30
1,1-Dichloroethane	1.388	1.305		-5.98	30
cis-1,2-Dichloroethene	1.553	1.406		-9.47	30
1,1,1-Trichloroethane	2.599	2.413		-7.16	30
Trichloroethene	0.471	0.368		-21.87	30
Tetrachloroethene	0.417	0.312		-25.18	30
1-Bromo-4-Fluorobenzene	0.822	0.882		7.3	30

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.