

Report of Analysis

Client:	PSEG	Date Collected:	04/21/25
Project:	Central Ave Substation	Date Received:	04/21/25
Client Sample ID:	CONCRETE-PILE-N-C-LOT	SDG No.:	Q1850
Lab Sample ID:	Q1850-01	Matrix:	SOIL
Analytical Method:	SW8151A	% Solid:	100
Sample Wt/Vol:	30.07	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	Herbicide
GPC Factor :	1.0	Injection Volume :	
Prep Method :	8151A		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PS029946.D	1	04/23/25 09:00	04/23/25 22:39	PB167710

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
1918-00-9	DICAMBA	0.067	U	0.0077	0.067	mg/Kg
120-36-5	DICHLORPROP	0.067	U	0.013	0.067	mg/Kg
94-75-7	2,4-D	0.067	U	0.0090	0.067	mg/Kg
93-72-1	2,4,5-TP (Silvex)	0.067	U	0.0090	0.067	mg/Kg
93-76-5	2,4,5-T	0.067	U	0.0087	0.067	mg/Kg
94-82-6	2,4-DB	0.067	U	0.024	0.067	mg/Kg
88-85-7	DINOSEB	0.067	U	0.011	0.067	mg/Kg
SURROGATES						
19719-28-9	2,4-DCAA	337		10 - 141	67%	SPK: 500

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit