

Continuing Calibration Report for SequenceID : FG042225AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FG015733.D

Aliphatic C9-C12	6745431.000	60.000	3.145	6.786	112423.850	112740.457	0.281
Aliphatic C12-C16	4676946.000	40.000	6.787	10.235	116923.650	115442.820	-1.283
Aliphatic C16-C21	7192390.000	60.000	10.236	13.612	119873.167	118965.713	-0.763
Aliphatic C21-C28	9129977.000	80.000	13.613	17.283	114124.713	113091.503	-0.914
Aliphatic C28-C40	11998647.000	120.000	17.284	22.194	99988.725	98657.746	-1.349
Aliphatic EPH	39743391.000	360.000	3.145	22.194	110398.308	109462.035	-0.855