

Continuing Calibration Report for SequenceID : FG042225AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FG015745.D

Aliphatic C9-C12	6114796.000	60.000	3.145	6.786	101913.267	112740.457	9.604
Aliphatic C12-C16	4281915.000	40.000	6.787	10.235	107047.875	115442.820	7.272
Aliphatic C16-C21	6771038.000	60.000	10.236	13.612	112850.633	118965.713	5.140
Aliphatic C21-C28	8957528.000	80.000	13.613	17.283	111969.100	113091.503	0.992
Aliphatic C28-C40	12702469.000	120.000	17.284	22.194	105853.908	98657.746	-7.294
Aliphatic EPH	38827746.000	360.000	3.145	22.194	107854.850	109462.035	1.468