

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01
 Lab Code: CHEM Case No.: Q1851 SAS No.: Q1851 SDG NO.: Q1851
 Lab File ID: VY021935.D Date Analyzed: 04/21/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:08
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	190380	7.71	283274	8.62	258729	11.42
UPPER LIMIT	380760	8.207	566548	9.116	517458	11.92
LOWER LIMIT	95190	7.207	141637	8.116	129365	10.92
EPA SAMPLE NO.						
GP5	202834	7.71	379948	8.62	314728	11.42
GP7	209157	7.71	398558	8.62	345335	11.42
VY0421SBL01	238181	7.71	431702	8.62	370399	11.42
VY0421SBS01	181890	7.71	285604	8.62	261318	11.42
VY0421SBSD01	184070	7.71	287552	8.62	258721	11.41

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.