

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01
 Lab Code: CHEM Case No.: Q1851 SAS No.: Q1851 SDG NO.: Q1851
 Lab File ID: VY021935.D Date Analyzed: 04/21/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:08
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	137455	13.347				
UPPER LIMIT	274910	13.847				
LOWER LIMIT	68727.5	12.847				
EPA SAMPLE NO.						
GP5	108735	13.35				
GP7	119952	13.35				
VY0421SBL01	140852	13.35				
VY0421SBS01	139221	13.35				
VY0421SBSD01	136375	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.