

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01
 Lab Code: CHEM Case No.: Q1851 SAS No.: Q1851 SDG NO.: Q1851
 Lab File ID: VY021962.D Date Analyzed: 04/23/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:57
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	227723	13.347				
UPPER LIMIT	455446	13.847				
LOWER LIMIT	113862	12.847				
EPA SAMPLE NO.						
GP6	115067	13.35				
VY0423SBL01	194434	13.35				
VY0423SBS01	221603	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.