



Contract:	<u>POWE02</u>						
Lab Code:	<u>CHEM</u>	Case No.:	<u>Q1858</u>	SAS No.:	<u>Q1858</u>	SDG NO.:	<u>Q1858</u>
Instrument ID:	<u>ECD_O</u>	Calibration Date(s):		<u>04/10/2025</u>	<u>04/10/2025</u>		
		Calibration Times:		<u>09:36</u>	<u>17:52</u>		

LAB FILE ID:		RT 1000 =	<u>PO110349.D</u>	RT 750 =	<u>PO110350.D</u>
RT 500 =	PO110351.D	RT 250 =	PO110352.D	RT 050 =	PO110353.D

[illegible]

RETENTION TIMES OF INITIAL CALIBRATION

Contract: POWE02

Lab Code: CHEM Case No.: Q1858 SAS No.: Q1858 SDG NO.: Q1858

Instrument ID: ECD_O Calibration Date(s): 04/10/2025 04/10/2025

Calibration Times: 09:36 17:52

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:	RT 1000 = <u>PO110349.D</u>	RT 750 = <u>PO110350.D</u>
RT 500 = <u>PO110351.D</u>	RT 250 = <u>PO110352.D</u>	RT 050 = <u>PO110353.D</u>

COMPOUND		RT 1000	RT 750	RT 500	RT 250	RT 050	MEAN RT	RT WINDOW	
								FROM	TO
Aroclor-1016-1	(1)	4.77	4.77	4.77	4.77	4.77	4.77	4.67	4.87
Aroclor-1016-2	(2)	4.78	4.79	4.78	4.78	4.79	4.78	4.68	4.88
Aroclor-1016-3	(3)	4.96	4.96	4.96	4.96	4.96	4.96	4.86	5.06
Aroclor-1016-4	(4)	5.00	5.00	5.00	5.00	5.00	5.00	4.90	5.10
Aroclor-1016-5	(5)	5.22	5.22	5.22	5.21	5.22	5.22	5.12	5.32
Aroclor-1260-1	(1)	6.25	6.25	6.25	6.25	6.25	6.25	6.15	6.35
Aroclor-1260-2	(2)	6.43	6.43	6.43	6.43	6.43	6.43	6.33	6.53
Aroclor-1260-3	(3)	6.59	6.59	6.59	6.59	6.59	6.59	6.49	6.69
Aroclor-1260-4	(4)	7.06	7.06	7.06	7.06	7.06	7.06	6.96	7.16
Aroclor-1260-5	(5)	7.30	7.30	7.30	7.30	7.30	7.30	7.20	7.40
Decachlorobiphenyl		8.69	8.69	8.68	8.68	8.68	8.68	8.58	8.78
Tetrachloro-m-xylene		3.69	3.69	3.69	3.69	3.69	3.69	3.59	3.79
Decachlorobiphenyl		8.68	8.68	8.68	8.68	8.68	8.68	8.58	8.78
Tetrachloro-m-xylene		3.69	3.69	3.68	3.69	3.68	3.69	3.59	3.79
Decachlorobiphenyl		8.68	8.69	8.69	8.68	8.68	8.68	8.58	8.78
Tetrachloro-m-xylene		3.69	3.69	3.69	3.69	3.68	3.69	3.59	3.79
Aroclor-1254-1	(1)	5.57	5.57	5.57	5.57	5.57	5.57	5.47	5.67
Aroclor-1254-2	(2)	5.71	5.71	5.71	5.71	5.71	5.71	5.61	5.81
Aroclor-1254-3	(3)	6.12	6.12	6.12	6.12	6.12	6.12	6.02	6.22
Aroclor-1254-4	(4)	6.34	6.34	6.34	6.34	6.34	6.34	6.24	6.44
Aroclor-1254-5	(5)	6.76	6.76	6.76	6.76	6.76	6.76	6.66	6.86
Decachlorobiphenyl		8.68	8.68	8.68	8.68	8.68	8.68	8.58	8.78
Tetrachloro-m-xylene		3.69	3.69	3.69	3.68	3.69	3.69	3.59	3.79
Decachlorobiphenyl		8.68	8.68	8.68	8.68	8.68	8.68	8.58	8.78
Tetrachloro-m-xylene		3.68	3.68	3.69	3.69	3.68	3.68	3.58	3.78