

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1858 SAS No.: Q1858 SDG NO.: Q1858

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/10/2025 04/10/2025

Client Sample No.: CCAL01 Date Analyzed: 04/23/2025

Lab Sample No.: AR1660CCC500 Data File : PO110638.D Time Analyzed: 08:44

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.776	4.679	4.879	549.930	500.000	10.0
Aroclor-1016-2	4.796	4.698	4.898	551.570	500.000	10.3
Aroclor-1016-3	4.852	4.755	4.955	537.620	500.000	7.5
Aroclor-1016-4	4.972	4.875	5.075	547.360	500.000	9.5
Aroclor-1016-5	5.228	5.132	5.332	518.290	500.000	3.7
Aroclor-1260-1	6.269	6.172	6.372	576.780	500.000	15.4
Aroclor-1260-2	6.458	6.360	6.560	550.750	500.000	10.2
Aroclor-1260-3	6.826	6.729	6.929	509.730	500.000	1.9
Aroclor-1260-4	7.086	6.988	7.188	498.950	500.000	-0.2
Aroclor-1260-5	7.328	7.230	7.430	503.370	500.000	0.7
Decachlorobiphenyl	8.730	8.632	8.832	45.790	50.000	-8.4
Tetrachloro-m-xylene	3.686	3.588	3.788	55.810	50.000	11.6

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1858 SAS No.: Q1858 SDG NO.: Q1858

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/10/2025 04/10/2025

Client Sample No.: CCAL01 Date Analyzed: 04/23/2025

Lab Sample No.: AR1660CCC500 Data File : PO110638.D Time Analyzed: 08:44

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.762	4.666	4.866	525.760	500.000	5.2
Aroclor-1016-2	4.781	4.684	4.884	522.330	500.000	4.5
Aroclor-1016-3	4.957	4.860	5.060	513.380	500.000	2.7
Aroclor-1016-4	4.999	4.902	5.102	421.440	500.000	-15.7
Aroclor-1016-5	5.212	5.115	5.315	583.490	500.000	16.7
Aroclor-1260-1	6.242	6.146	6.346	530.130	500.000	6.0
Aroclor-1260-2	6.431	6.334	6.534	512.290	500.000	2.5
Aroclor-1260-3	6.583	6.487	6.687	494.610	500.000	-1.1
Aroclor-1260-4	7.055	6.957	7.157	494.230	500.000	-1.2
Aroclor-1260-5	7.295	7.197	7.397	495.190	500.000	-1.0
Decachlorobiphenyl	8.681	8.584	8.784	47.490	50.000	-5.0
Tetrachloro-m-xylene	3.683	3.586	3.786	54.270	50.000	8.5