

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1858 SAS No.: Q1858 SDG NO.: Q1858

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/10/2025 04/10/2025

Client Sample No.: CCAL02 Date Analyzed: 04/23/2025

Lab Sample No.: AR1660CCC500 Data File : PO110652.D Time Analyzed: 16:20

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.776	4.679	4.879	554.870	500.000	11.0
Aroclor-1016-2	4.795	4.698	4.898	557.710	500.000	11.5
Aroclor-1016-3	4.852	4.755	4.955	543.560	500.000	8.7
Aroclor-1016-4	4.972	4.875	5.075	552.400	500.000	10.5
Aroclor-1016-5	5.229	5.132	5.332	538.640	500.000	7.7
Aroclor-1260-1	6.269	6.172	6.372	567.990	500.000	13.6
Aroclor-1260-2	6.458	6.360	6.560	529.560	500.000	5.9
Aroclor-1260-3	6.825	6.729	6.929	518.020	500.000	3.6
Aroclor-1260-4	7.086	6.988	7.188	511.830	500.000	2.4
Aroclor-1260-5	7.328	7.230	7.430	508.430	500.000	1.7
Decachlorobiphenyl	8.729	8.632	8.832	46.500	50.000	-7.0
Tetrachloro-m-xylene	3.686	3.588	3.788	56.370	50.000	12.7

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1858 SAS No.: Q1858 SDG NO.: Q1858

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/10/2025 04/10/2025

Client Sample No.: CCAL02 Date Analyzed: 04/23/2025

Lab Sample No.: AR1660CCC500 Data File : PO110652.D Time Analyzed: 16:20

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.763	4.666	4.866	528.500	500.000	5.7
Aroclor-1016-2	4.782	4.684	4.884	525.700	500.000	5.1
Aroclor-1016-3	4.957	4.860	5.060	518.670	500.000	3.7
Aroclor-1016-4	5.000	4.902	5.102	420.850	500.000	-15.8
Aroclor-1016-5	5.212	5.115	5.315	503.960	500.000	0.8
Aroclor-1260-1	6.243	6.146	6.346	528.930	500.000	5.8
Aroclor-1260-2	6.431	6.334	6.534	506.730	500.000	1.3
Aroclor-1260-3	6.584	6.487	6.687	490.180	500.000	-2.0
Aroclor-1260-4	7.055	6.957	7.157	492.840	500.000	-1.4
Aroclor-1260-5	7.296	7.197	7.397	490.200	500.000	-2.0
Decachlorobiphenyl	8.681	8.584	8.784	47.120	50.000	-5.8
Tetrachloro-m-xylene	3.684	3.586	3.786	54.770	50.000	9.5