

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1858 SAS No.: Q1858 SDG NO.: Q1858

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/10/2025 04/10/2025

Client Sample No.: CCAL03 Date Analyzed: 04/23/2025

Lab Sample No.: AR1660CCC500 Data File : PO110667.D Time Analyzed: 22:50

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.776	4.679	4.879	559.600	500.000	11.9
Aroclor-1016-2	4.796	4.698	4.898	564.500	500.000	12.9
Aroclor-1016-3	4.852	4.755	4.955	550.930	500.000	10.2
Aroclor-1016-4	4.972	4.875	5.075	561.080	500.000	12.2
Aroclor-1016-5	5.229	5.132	5.332	539.150	500.000	7.8
Aroclor-1260-1	6.269	6.172	6.372	583.100	500.000	16.6
Aroclor-1260-2	6.458	6.360	6.560	555.900	500.000	11.2
Aroclor-1260-3	6.826	6.729	6.929	536.310	500.000	7.3
Aroclor-1260-4	7.086	6.988	7.188	528.490	500.000	5.7
Aroclor-1260-5	7.328	7.230	7.430	524.450	500.000	4.9
Decachlorobiphenyl	8.727	8.632	8.832	46.930	50.000	-6.1
Tetrachloro-m-xylene	3.687	3.588	3.788	57.040	50.000	14.1

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1858 SAS No.: Q1858 SDG NO.: Q1858

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/10/2025 04/10/2025

Client Sample No.: CCAL03 Date Analyzed: 04/23/2025

Lab Sample No.: AR1660CCC500 Data File : PO110667.D Time Analyzed: 22:50

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.763	4.666	4.866	537.030	500.000	7.4
Aroclor-1016-2	4.782	4.684	4.884	529.750	500.000	6.0
Aroclor-1016-3	4.958	4.860	5.060	529.630	500.000	5.9
Aroclor-1016-4	5.000	4.902	5.102	405.680	500.000	-18.9
Aroclor-1016-5	5.212	5.115	5.315	509.900	500.000	2.0
Aroclor-1260-1	6.244	6.146	6.346	540.930	500.000	8.2
Aroclor-1260-2	6.431	6.334	6.534	520.700	500.000	4.1
Aroclor-1260-3	6.584	6.487	6.687	497.360	500.000	-0.5
Aroclor-1260-4	7.055	6.957	7.157	492.790	500.000	-1.4
Aroclor-1260-5	7.296	7.197	7.397	493.880	500.000	-1.2
Decachlorobiphenyl	8.680	8.584	8.784	47.310	50.000	-5.4
Tetrachloro-m-xylene	3.684	3.586	3.786	55.340	50.000	10.7