

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: POWE02

Lab Code: CHEM Case No.: Q1858 SAS No.: Q1858 SDG NO.: Q1858

Instrument ID: ECD_D

Calibration Date(s): 04/18/2025 04/18/2025

Calibration Times: 13:56 14:51

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:		CF 100 =	<u>PD088124.D</u>	CF 075 =	<u>PD088125.D</u>		
CF 050 =		<u>PD088126.D</u>	CF 025 =	<u>PD088127.D</u>	CF 005 =	<u>PD088128.D</u>	
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
4,4'-DDD	2657600000	2587000000	2495010000	2376000000	2459180000	2514960000	4
4,4'-DDE	3466910000	3527170000	3240150000	3071100000	3185920000	3298250000	6
4,4'-DDT	2923480000	2868140000	2755010000	2629860000	2711210000	2777540000	4
Aldrin	4191470000	4069870000	3911790000	3719150000	3855640000	3949580000	5
Decachlorobiphenyl	3080820000	3141130000	3178140000	3290360000	3850090000	3308110000	9
Dieldrin	3750160000	3656120000	3530390000	3371440000	3534380000	3568500000	4
Tetrachloro-m-xylene	1982340000	2006790000	1938680000	1923660000	2135510000	1997400000	4

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: POWE02

Lab Code: CHEM Case No.: Q1858 SAS No.: Q1858 SDG NO.: Q1858

Instrument ID: ECD_D

Calibration Date(s): 04/18/2025 04/18/2025

Calibration Times: 13:56 14:51

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:		CF 100 =	<u>PD088124.D</u>	CF 075 =	<u>PD088125.D</u>		
CF 050 =		<u>PD088126.D</u>	CF 025 =	<u>PD088127.D</u>	CF 005 =	<u>PD088128.D</u>	
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
4,4'-DDD	15154700000	15403900000	15792200000	16361000000	19219200000	16386200000	10
4,4'-DDE	18345500000	18580000000	18872600000	19581800000	23090100000	19694000000	10
4,4'-DDT	16431600000	16496500000	16745700000	17063800000	18344900000	17016500000	5
Aldrin	19439700000	19604700000	20000000000	20715500000	24254000000	20802800000	10
Decachlorobiphenyl	16767000000	17098200000	17470300000	18387600000	22674800000	18479600000	13
Dieldrin	18536800000	18713100000	19146400000	19886400000	23381600000	19932900000	10
Tetrachloro-m-xylene	13615300000	13685800000	14010800000	14551200000	17245000000	14621600000	10