

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1858 SAS No.: Q1858 SDG NO.: Q1858

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL01 Date Analyzed: 04/23/2025

Lab Sample No.: PSTDCCC050 Data File : PD088233.D Time Analyzed: 09:55

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.715	6.606	6.806	52.650	50.000	5.3
4,4'-DDE	6.206	6.097	6.297	51.060	50.000	2.1
4,4'-DDT	7.031	6.922	7.122	47.170	50.000	-5.7
Aldrin	5.281	5.173	5.373	52.870	50.000	5.7
Decachlorobiphenyl	9.085	8.975	9.175	45.570	50.000	-8.9
Dieldrin	6.357	6.249	6.449	52.670	50.000	5.3
Tetrachloro-m-xylene	3.558	3.452	3.652	51.590	50.000	3.2

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1858 SAS No.: Q1858 SDG NO.: Q1858

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL01 Date Analyzed: 04/23/2025

Lab Sample No.: PSTDCCC050 Data File : PD088233.D Time Analyzed: 09:55

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.934	5.834	6.034	47.660	50.000	-4.7
4,4'-DDE	5.379	5.280	5.480	46.970	50.000	-6.1
4,4'-DDT	6.188	6.088	6.288	43.900	50.000	-12.2
Aldrin	4.371	4.273	4.473	47.670	50.000	-4.7
Decachlorobiphenyl	8.077	7.977	8.177	42.400	50.000	-15.2
Dieldrin	5.516	5.417	5.617	47.120	50.000	-5.8
Tetrachloro-m-xylene	2.881	2.783	2.983	47.500	50.000	-5.0