

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1858 SAS No.: Q1858 SDG NO.: Q1858

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL04 Date Analyzed: 04/28/2025

Lab Sample No.: PSTDCCC050 Data File : PD088303.D Time Analyzed: 15:10

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
4,4'-DDD	6.714	6.606	6.806	52.640	50.000	5.3
4,4'-DDE	6.204	6.097	6.297	50.750	50.000	1.5
4,4'-DDT	7.030	6.922	7.122	46.560	50.000	-6.9
Aldrin	5.280	5.173	5.373	53.300	50.000	6.6
Decachlorobiphenyl	9.083	8.975	9.175	45.920	50.000	-8.2
Dieldrin	6.356	6.249	6.449	52.730	50.000	5.5
Tetrachloro-m-xylene	3.558	3.452	3.652	51.300	50.000	2.6

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1858 SAS No.: Q1858 SDG NO.: Q1858

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL04 Date Analyzed: 04/28/2025

Lab Sample No.: PSTDCCC050 Data File : PD088303.D Time Analyzed: 15:10

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.933	5.834	6.034	47.040	50.000	-5.9
4,4'-DDE	5.379	5.280	5.480	47.880	50.000	-4.2
4,4'-DDT	6.188	6.088	6.288	43.610	50.000	-12.8
Aldrin	4.372	4.273	4.473	48.020	50.000	-4.0
Decachlorobiphenyl	8.077	7.977	8.177	40.440	50.000	-19.1
Dieldrin	5.517	5.417	5.617	47.610	50.000	-4.8
Tetrachloro-m-xylene	2.881	2.783	2.983	47.960	50.000	-4.1