

### CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1858 SAS No.: Q1858 SDG NO.: Q1858

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL03 Date Analyzed: 04/23/2025

Lab Sample No.: PSTDCCC050 Data File : PD088257.D Time Analyzed: 21:58

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D    |
|----------------------|-------|-----------|-------|--------------------|-------------------|-------|
|                      |       | FROM      | TO    |                    |                   |       |
| 4,4'-DDD             | 6.706 | 6.606     | 6.806 | 59.130             | 50.000            | 18.3  |
| 4,4'-DDE             | 6.195 | 6.097     | 6.297 | 56.460             | 50.000            | 12.9  |
| 4,4'-DDT             | 7.022 | 6.922     | 7.122 | 43.360             | 50.000            | -13.3 |
| Aldrin               | 5.273 | 5.173     | 5.373 | 58.610             | 50.000            | 17.2  |
| Decachlorobiphenyl   | 9.074 | 8.975     | 9.175 | 48.850             | 50.000            | -2.3  |
| Dieldrin             | 6.348 | 6.249     | 6.449 | 58.100             | 50.000            | 16.2  |
| Tetrachloro-m-xylene | 3.552 | 3.452     | 3.652 | 56.540             | 50.000            | 13.1  |

### CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1858 SAS No.: Q1858 SDG NO.: Q1858

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/18/2025 04/18/2025

Client Sample No.: CCAL03 Date Analyzed: 04/23/2025

Lab Sample No.: PSTDCCC050 Data File : PD088257.D Time Analyzed: 21:58

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D    |
|----------------------|-------|-----------|-------|--------------------|-------------------|-------|
|                      |       | FROM      | TO    |                    |                   |       |
| 4,4'-DDD             | 5.932 | 5.834     | 6.034 | 52.870             | 50.000            | 5.7   |
| 4,4'-DDE             | 5.377 | 5.280     | 5.480 | 50.860             | 50.000            | 1.7   |
| 4,4'-DDT             | 6.187 | 6.088     | 6.288 | 40.850             | 50.000            | -18.3 |
| Aldrin               | 4.373 | 4.273     | 4.473 | 52.190             | 50.000            | 4.4   |
| Decachlorobiphenyl   | 8.076 | 7.977     | 8.177 | 43.690             | 50.000            | -12.6 |
| Dieldrin             | 5.516 | 5.417     | 5.617 | 51.090             | 50.000            | 2.2   |
| Tetrachloro-m-xylene | 2.884 | 2.783     | 2.983 | 51.730             | 50.000            | 3.5   |