

Instrument ID: BNA_M

Daily Analysis Runlog For Sequence/QC Batch ID # BM040825

Review By	anahy	Review On	4/9/2025 10:38:14 AM
Supervise By	Jagrut	Supervise On	4/9/2025 4:50:06 PM
SubDirectory	BM040825	HP Acquire Method	BNA_M
		HP Processing Method	bm040825

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12657,10ul/1000ul sample
ICV/I.BLK	SP6686
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BM049847.D	08 Apr 2025 12:55		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BM049848.D	08 Apr 2025 13:35		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BM049849.D	08 Apr 2025 14:14	Compound #32,54,65 removed from 5ppm	RC/JU	Ok,M
4	SSTDICC010	SSTDICC010	BM049850.D	08 Apr 2025 14:53	Compound #32,54 kept on LR	RC/JU	Ok,M
5	SSTDICC020	SSTDICC020	BM049851.D	08 Apr 2025 15:32	This calibration is fail for Com#77	RC/JU	Ok
6	SSTDICCC040	SSTDICCC040	BM049852.D	08 Apr 2025 16:12	This calibration is good for 8270E DOD except Com#77 and good for 625.1 method	RC/JU	Ok
7	SSTDICC050	SSTDICC050	BM049853.D	08 Apr 2025 17:30		RC/JU	Ok
8	SSTDICC060	SSTDICC060	BM049854.D	08 Apr 2025 19:28	Compound #69 Removed from 60PPM	RC/JU	Ok,M
9	SSTDICC080	SSTDICC080	BM049855.D	08 Apr 2025 20:07	Compound #69 Removed from 80PPM	RC/JU	Ok,M
10	SSTDICV040	ICVBM040825	BM049856.D	08 Apr 2025 20:47		RC/JU	Ok
11	PB167474BL	PB167474BL	BM049857.D	08 Apr 2025 21:26	Use for Q1730 only	RC/JU	Not Ok

M : Manual Integration

Instrument ID: BNA_M

Daily Analysis Runlog For Sequence/QC Batch ID # BM042325

Review By	Rahul	Review On	4/24/2025 9:48:24 AM		
Supervise By	Jagrut	Supervise On	4/24/2025 2:07:47 PM		
SubDirectory	BM042325	HP Acquire Method	BNA_M	HP Processing Method	BM040825

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12660,10ul/1000ul sample
ICV/I.BLK	SP6686
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BM050006.D	23 Apr 2025 09:18		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BM050007.D	23 Apr 2025 10:38		RC/JU	Ok
3	PB167694BL	PB167694BL	BM050008.D	23 Apr 2025 11:17		RC/JU	Ok
4	PB167694BS	PB167694BS	BM050009.D	23 Apr 2025 11:57		RC/JU	Ok,M
5	PB167694BSD	PB167694BSD	BM050010.D	23 Apr 2025 12:36		RC/JU	Ok,M
6	Q1848-03DL	ETGI-275DL	BM050011.D	23 Apr 2025 13:19		RC/JU	Ok,M
7	Q1852-07	72-12013	BM050012.D	23 Apr 2025 13:58		RC/JU	Ok,M
8	Q1852-07MS	72-12013MS	BM050013.D	23 Apr 2025 14:37		RC/JU	Ok,M
9	Q1852-07MSD	72-12013MSD	BM050014.D	23 Apr 2025 15:16		RC/JU	Ok,M
10	Q1858-01	COMP-1	BM050015.D	23 Apr 2025 15:55		RC/JU	Ok
11	Q1858-02	COMP-2	BM050016.D	23 Apr 2025 16:35		RC/JU	Ok,M
12	Q1858-03	COMP-3	BM050017.D	23 Apr 2025 17:14		RC/JU	Ok
13	Q1859-02	COMP-2	BM050018.D	23 Apr 2025 17:53		RC/JU	Ok
14	Q1859-03	COMP-3	BM050019.D	23 Apr 2025 18:32		RC/JU	Ok
15	Q1859-01	COMP-1	BM050020.D	23 Apr 2025 19:11		RC/JU	Ok
16	Q1853-01	EO-02-04222025	BM050021.D	23 Apr 2025 19:50		RC/JU	Ok,M
17	Q1854-01	NB-08-04222025	BM050022.D	23 Apr 2025 20:29		RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP041425

Review By	Rahul	Review On	4/15/2025 9:55:40 AM
Supervise By	Jagrut	Supervise On	4/15/2025 10:53:04 AM
SubDirectory	BP041425	HP Acquire Method	BNA_P
		HP Processing Method	bp041425

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12659,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP024274.D	14 Apr 2025 10:25		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BP024275.D	14 Apr 2025 11:06		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BP024276.D	14 Apr 2025 11:47	Compound #32,54,65 removed from 5ppm	RC/JU	Ok,M
4	SSTDICC010	SSTDICC010	BP024277.D	14 Apr 2025 12:27		RC/JU	Ok,M
5	SSTDICC020	SSTDICC020	BP024278.D	14 Apr 2025 13:08		RC/JU	Ok,M
6	SSTDICCC040	SSTDICCC040	BP024279.D	14 Apr 2025 13:49	This calibration is fail for Com#77	RC/JU	Ok
7	SSTDICC050	SSTDICC050	BP024280.D	14 Apr 2025 15:10	This calibration is good for 8270E, 8270 DOD and 625.1 methods except Compound#77	RC/JU	Ok,M
8	SSTDICC060	SSTDICC060	BP024281.D	14 Apr 2025 16:32	Compound #69 Removed from 60PPM	RC/JU	Ok,M
9	SSTDICC080	SSTDICC080	BP024282.D	14 Apr 2025 17:13	Compound #69 Removed from 80PPM	RC/JU	Ok,M
10	SSTDICV040	ICVBP041425	BP024283.D	14 Apr 2025 18:35		RC/JU	Ok,M
11	PB167488TB	PB167488TB	BP024284.D	14 Apr 2025 19:16		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP042425

Review By	Jagrut	Review On	4/25/2025 10:24:06 AM
Supervise By	Rahul	Supervise On	4/25/2025 10:24:26 AM
SubDirectory	BP042425	HP Acquire Method	BNA_P
		HP Processing Method	bp041425

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12660,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP024406.D	24 Apr 2025 11:10		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BP024407.D	24 Apr 2025 11:50		RC/JU	Ok,M
3	PB167711BL	PB167711BL	BP024408.D	24 Apr 2025 12:31		RC/JU	Ok
4	PB167711BS	PB167711BS	BP024409.D	24 Apr 2025 13:12		RC/JU	Ok,M
5	Q1852-02	ETGI-354	BP024410.D	24 Apr 2025 16:35	Internal Standard Fail	RC/JU	ReRun
6	Q1852-02MS	ETGI-354MS	BP024411.D	24 Apr 2025 17:16		RC/JU	Ok,M
7	Q1852-02MSD	ETGI-354MSD	BP024412.D	24 Apr 2025 17:56		RC/JU	Ok,M
8	Q1852-04	72-11977	BP024413.D	24 Apr 2025 18:37		RC/JU	Ok
9	Q1852-06	ETGI-278	BP024414.D	24 Apr 2025 19:18	Internal Standard Fail	RC/JU	ReRun
10	Q1852-08	72-12013	BP024415.D	24 Apr 2025 19:58	Internal Standard Fail	RC/JU	ReRun

M : Manual Integration