

LAB CHRONICLE

OrderID:	Q1858	OrderDate:	4/22/2025 2:54:00 PM
Client:	Kleinfelder	Project:	Henry Lea School
Contact:	Mark Warchol	Location:	L41,VOA Ref. #2 Soil

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1858-01	COMP-1	SOIL	SVOCMS Group1	8270E	04/21/25	04/23/25	04/23/25	04/22/25
Q1858-02	COMP-2	SOIL	SVOCMS Group1	8270E	04/21/25	04/23/25	04/23/25	04/22/25
Q1858-03	COMP-3	SOIL	SVOCMS Group1	8270E	04/21/25	04/23/25	04/23/25	04/22/25

Hit Summary Sheet SW-846

SDG No.: Q1858
Client: Kleinfelder

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID : COMP-1								
Q1858-01	COMP-1	SOIL	Phenanthrene	97.100	J	24.9	200	ug/Kg
Q1858-01	COMP-1	SOIL	Pyrene	96.100	J	42.8	200	ug/Kg
Total Svoc :				193.20				
Total Concentration:				193.20				
Client ID : COMP-2								
Q1858-02	COMP-2	SOIL	Phenanthrene	220.000		25.7	210	ug/Kg
Q1858-02	COMP-2	SOIL	Pyrene	280.000		44.2	210	ug/Kg
Q1858-02	COMP-2	SOIL	Benzo(a)anthracene	160.000	J	28.3	210	ug/Kg
Q1858-02	COMP-2	SOIL	Chrysene	170.000	J	24.5	210	ug/Kg
Q1858-02	COMP-2	SOIL	Benzo(b)fluoranthene	200.000	J	23.3	210	ug/Kg
Q1858-02	COMP-2	SOIL	Benzo(a)pyrene	150.000	J	36.2	210	ug/Kg
Q1858-02	COMP-2	SOIL	Indeno(1,2,3-cd)pyrene	86.100	J	35.8	210	ug/Kg
Q1858-02	COMP-2	SOIL	Benzo(g,h,i)perylene	110.000	J	31.6	210	ug/Kg
Total Svoc :				1,376.10				
Total Concentration:				1,376.10				
Client ID : COMP-3								
Q1858-03	COMP-3	SOIL	Phenanthrene	150.000	J	22.9	190	ug/Kg
Q1858-03	COMP-3	SOIL	Pyrene	92.500	J	39.4	190	ug/Kg
Total Svoc :				242.50				
Total Concentration:				242.50				



SAMPLE DATA

Report of Analysis

Client:	Kleinfelder	Date Collected:	04/21/25
Project:	Henry Lea School	Date Received:	04/22/25
Client Sample ID:	COMP-1	SDG No.:	Q1858
Lab Sample ID:	Q1858-01	Matrix:	SOIL
Analytical Method:	SW8270	% Solid:	83.9
Sample Wt/Vol:	30.06	Units:	g
Soil Aliquot Vol:			uL
Extraction Type :		Decanted :	N
Injection Volume :		GPC Factor :	1.0
Prep Method :	SW3541	Level :	LOW
		Test:	SVOCMS Group1
		Final Vol:	1000
		GPC Cleanup :	N
		PH :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM050015.D	1	04/23/25 09:20	04/23/25 15:55	PB167711

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
91-20-3	Naphthalene	27.0	U	27.0	200	ug/Kg
86-73-7	Fluorene	30.1	U	30.1	200	ug/Kg
85-01-8	Phenanthrene	97.1	J	24.9	200	ug/Kg
120-12-7	Anthracene	39.6	U	39.6	200	ug/Kg
129-00-0	Pyrene	96.1	J	42.8	200	ug/Kg
56-55-3	Benzo(a)anthracene	27.4	U	27.4	200	ug/Kg
218-01-9	Chrysene	23.7	U	23.7	200	ug/Kg
205-99-2	Benzo(b)fluoranthene	22.6	U	22.6	200	ug/Kg
50-32-8	Benzo(a)pyrene	35.1	U	35.1	200	ug/Kg
193-39-5	Indeno(1,2,3-cd)pyrene	34.6	U	34.6	200	ug/Kg
191-24-2	Benzo(g,h,i)perylene	30.6	U	30.6	200	ug/Kg
SURROGATES						
4165-60-0	Nitrobenzene-d5	56.5		18 - 107	56%	SPK: 100
321-60-8	2-Fluorobiphenyl	52.9		20 - 109	53%	SPK: 100
1718-51-0	Terphenyl-d14	54.7		10 - 105	55%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	268000	7.763			
1146-65-2	Naphthalene-d8	920000	10.557			
15067-26-2	Acenaphthene-d10	612000	14.41			
1517-22-2	Phenanthrene-d10	1240000	17.156			
1719-03-5	Chrysene-d12	1320000	21.397			
1520-96-3	Perylene-d12	1400000	24.403			

Report of Analysis

Client:	Kleinfelder	Date Collected:	04/21/25
Project:	Henry Lea School	Date Received:	04/22/25
Client Sample ID:	COMP-1	SDG No.:	Q1858
Lab Sample ID:	Q1858-01	Matrix:	SOIL
Analytical Method:	SW8270	% Solid:	83.9
Sample Wt/Vol:	30.06	Units:	g
Soil Aliquot Vol:			uL
Extraction Type :		Decanted :	N
Injection Volume :		GPC Factor :	1.0
Prep Method :	SW3541	Final Vol:	1000
		Test:	SVOCMS Group1
		Level :	LOW
		GPC Cleanup :	N
		PH :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM050015.D	1	04/23/25 09:20	04/23/25 15:55	PB167711

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Kleinfelder	Date Collected:	04/21/25
Project:	Henry Lea School	Date Received:	04/22/25
Client Sample ID:	COMP-2	SDG No.:	Q1858
Lab Sample ID:	Q1858-02	Matrix:	SOIL
Analytical Method:	SW8270	% Solid:	81.3
Sample Wt/Vol:	30.03 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM050016.D	1	04/23/25 09:20	04/23/25 16:35	PB167711

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
91-20-3	Naphthalene	27.9	U	27.9	210	ug/Kg
86-73-7	Fluorene	31.1	U	31.1	210	ug/Kg
85-01-8	Phenanthrene	220		25.7	210	ug/Kg
120-12-7	Anthracene	40.9	U	40.9	210	ug/Kg
129-00-0	Pyrene	280		44.2	210	ug/Kg
56-55-3	Benzo(a)anthracene	160	J	28.3	210	ug/Kg
218-01-9	Chrysene	170	J	24.5	210	ug/Kg
205-99-2	Benzo(b)fluoranthene	200	J	23.3	210	ug/Kg
50-32-8	Benzo(a)pyrene	150	J	36.2	210	ug/Kg
193-39-5	Indeno(1,2,3-cd)pyrene	86.1	J	35.8	210	ug/Kg
191-24-2	Benzo(g,h,i)perylene	110	J	31.6	210	ug/Kg
SURROGATES						
4165-60-0	Nitrobenzene-d5	62.9		18 - 107	63%	SPK: 100
321-60-8	2-Fluorobiphenyl	57.2		20 - 109	57%	SPK: 100
1718-51-0	Terphenyl-d14	58.4		10 - 105	58%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	281000	7.763			
1146-65-2	Naphthalene-d8	963000	10.557			
15067-26-2	Acenaphthene-d10	636000	14.41			
1517-22-2	Phenanthrene-d10	1310000	17.157			
1719-03-5	Chrysene-d12	1370000	21.398			
1520-96-3	Perylene-d12	1420000	24.397			

Report of Analysis

Client:	Kleinfelder	Date Collected:	04/21/25
Project:	Henry Lea School	Date Received:	04/22/25
Client Sample ID:	COMP-2	SDG No.:	Q1858
Lab Sample ID:	Q1858-02	Matrix:	SOIL
Analytical Method:	SW8270	% Solid:	81.3
Sample Wt/Vol:	30.03	Units:	g
Soil Aliquot Vol:			uL
Extraction Type :		Decanted :	N
Injection Volume :		GPC Factor :	1.0
Prep Method :	SW3541	Final Vol:	1000
		Test:	SVOCMS Group1
		Level :	LOW
		GPC Cleanup :	N
		PH :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM050016.D	1	04/23/25 09:20	04/23/25 16:35	PB167711

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Kleinfelder	Date Collected:	04/21/25
Project:	Henry Lea School	Date Received:	04/22/25
Client Sample ID:	COMP-3	SDG No.:	Q1858
Lab Sample ID:	Q1858-03	Matrix:	SOIL
Analytical Method:	SW8270	% Solid:	91.1
Sample Wt/Vol:	30.07 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM050017.D	1	04/23/25 09:20	04/23/25 17:14	PB167711

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
91-20-3	Naphthalene	24.9	U	24.9	190	ug/Kg
86-73-7	Fluorene	27.7	U	27.7	190	ug/Kg
85-01-8	Phenanthrene	150	J	22.9	190	ug/Kg
120-12-7	Anthracene	36.5	U	36.5	190	ug/Kg
129-00-0	Pyrene	92.5	J	39.4	190	ug/Kg
56-55-3	Benzo(a)anthracene	25.2	U	25.2	190	ug/Kg
218-01-9	Chrysene	21.8	U	21.8	190	ug/Kg
205-99-2	Benzo(b)fluoranthene	20.8	U	20.8	190	ug/Kg
50-32-8	Benzo(a)pyrene	32.3	U	32.3	190	ug/Kg
193-39-5	Indeno(1,2,3-cd)pyrene	31.9	U	31.9	190	ug/Kg
191-24-2	Benzo(g,h,i)perylene	28.1	U	28.1	190	ug/Kg
SURROGATES						
4165-60-0	Nitrobenzene-d5	49.8		18 - 107	50%	SPK: 100
321-60-8	2-Fluorobiphenyl	47.3		20 - 109	47%	SPK: 100
1718-51-0	Terphenyl-d14	54.1		10 - 105	54%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	296000	7.763			
1146-65-2	Naphthalene-d8	1030000	10.557			
15067-26-2	Acenaphthene-d10	691000	14.41			
1517-22-2	Phenanthrene-d10	1440000	17.157			
1719-03-5	Chrysene-d12	1430000	21.397			
1520-96-3	Perylene-d12	1480000	24.397			

Report of Analysis

Client:	Kleinfelder	Date Collected:	04/21/25
Project:	Henry Lea School	Date Received:	04/22/25
Client Sample ID:	COMP-3	SDG No.:	Q1858
Lab Sample ID:	Q1858-03	Matrix:	SOIL
Analytical Method:	SW8270	% Solid:	91.1
Sample Wt/Vol:	30.07	Units:	g
Soil Aliquot Vol:			uL
Extraction Type :		Decanted :	N
Injection Volume :		GPC Factor :	1.0
Prep Method :	SW3541	Level :	LOW
		Final Vol:	1000
		Test:	SVOCMS Group1
		GPC Cleanup :	N
		PH :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM050017.D	1	04/23/25 09:20	04/23/25 17:14	PB167711

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



QC SUMMARY

Surrogate Summary

SW-846

SDG No.: Q1858

Client: Kleinfelder

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB167711BL	PB167711BL	Nitrobenzene-d5	100	91.5	92		18	107
		2-Fluorobiphenyl	100	92.2	92		20	109
		Terphenyl-d14	100	98.5	98		10	105
PB167711BS	PB167711BS	Nitrobenzene-d5	100	90.4	90		18	107
		2-Fluorobiphenyl	100	89.9	90		20	109
		Terphenyl-d14	100	101	101		10	105
Q1852-07MS	72-12013MS	Nitrobenzene-d5	100	69.4	69		18	107
		2-Fluorobiphenyl	100	67.1	67		20	109
		Terphenyl-d14	100	65.1	65		10	105
Q1852-07MSD	72-12013MSD	Nitrobenzene-d5	100	68.6	69		18	107
		2-Fluorobiphenyl	100	63.7	64		20	109
		Terphenyl-d14	100	63.1	63		10	105
Q1858-01	COMP-1	Nitrobenzene-d5	100	56.5	56		18	107
		2-Fluorobiphenyl	100	52.9	53		20	109
		Terphenyl-d14	100	54.7	55		10	105
Q1858-02	COMP-2	Nitrobenzene-d5	100	62.9	63		18	107
		2-Fluorobiphenyl	100	57.2	57		20	109
		Terphenyl-d14	100	58.4	58		10	105
Q1858-03	COMP-3	Nitrobenzene-d5	100	49.8	50		18	107
		2-Fluorobiphenyl	100	47.3	47		20	109
		Terphenyl-d14	100	54.1	54		10	105

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1858

Client: Kleinfelder

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID: Q1852-07MS		Client Sample ID: 72-12013MS		DataFile: BM050013.D							
Naphthalene	1100	0	990	ug/Kg	90				72	110	
Fluorene	1100	0	1000	ug/Kg	91				68	116	
Phenanthrene	1100	0	1100	ug/Kg	100				52	128	
Anthracene	1100	0	1100	ug/Kg	100				62	124	
Pyrene	1100	0	1000	ug/Kg	91				26	142	
Benzo(a)anthracene	1100	0	1100	ug/Kg	100				71	114	
Chrysene	1100	0	1100	ug/Kg	100				57	121	
Benzo(b)fluoranthene	1100	0	1100	ug/Kg	100				67	121	
Benzo(a)pyrene	1100	0	1200	ug/Kg	109				70	142	
Indeno(1,2,3-cd)pyrene	1100	0	1100	ug/Kg	100				40	129	
Benzo(g,h,i)perylene	1100	0	1000	ug/Kg	91				24	125	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1858

Client: Kleinfelder

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	Q1852-07MSD	Client Sample ID:	72-12013MSD					DataFile:	BM050014.D		
Naphthalene	1100	0	970	ug/Kg	88		2		72	110	20
Fluorene	1100	0	990	ug/Kg	90		1		68	116	20
Phenanthrene	1100	0	1000	ug/Kg	91		9		52	128	20
Anthracene	1100	0	1100	ug/Kg	100		0		62	124	20
Pyrene	1100	0	990	ug/Kg	90		1		26	142	20
Benzo(a)anthracene	1100	0	1100	ug/Kg	100		0		71	114	20
Chrysene	1100	0	1000	ug/Kg	91		9		57	121	20
Benzo(b)fluoranthene	1100	0	1000	ug/Kg	91		9		67	121	20
Benzo(a)pyrene	1100	0	1100	ug/Kg	100		9		70	142	20
Indeno(1,2,3-cd)pyrene	1100	0	1100	ug/Kg	100		0		40	129	20
Benzo(g,h,i)perylene	1100	0	970	ug/Kg	88		3		24	125	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1858

Client: Kleinfelder

Analytical Method: 8270E DataFile: BP024409.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		
								Qual	Low	High	RPD
PB167711BS	Naphthalene	1700	1500	ug/Kg	88				62	100	
	Fluorene	1700	1500	ug/Kg	88				61	101	
	Phenanthrene	1700	1500	ug/Kg	88				59	103	
	Anthracene	1700	1600	ug/Kg	94				61	105	
	Pyrene	1700	1700	ug/Kg	100				59	103	
	Benzo(a)anthracene	1700	1600	ug/Kg	94				60	102	
	Chrysene	1700	1500	ug/Kg	88				59	101	
	Benzo(b)fluoranthene	1700	1600	ug/Kg	94				62	109	
	Benzo(a)pyrene	1700	1700	ug/Kg	100				63	103	
	Indeno(1,2,3-cd)pyrene	1700	1600	ug/Kg	94				63	101	
	Benzo(g,h,i)perylene	1700	1500	ug/Kg	88				70	108	

4B

SEMIVOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB167711BL

Lab Name: CHEMTECH

Contract: POWE02

Lab Code: CHEM Case No.: Q1858

SAS No.: Q1858 SDG NO.: Q1858

Lab File ID: BP024408.D

Lab Sample ID: PB167711BL

Instrument ID: BNA_P

Date Extracted: 04/23/2025

Matrix: (soil/water) SOIL

Date Analyzed: 04/24/2025

Level: (low/med) LOW

Time Analyzed: 12:31

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
PB167711BS	PB167711BS	BP024409.D	04/24/2025
72-12013MS	Q1852-07MS	BM050013.D	04/23/2025
72-12013MSD	Q1852-07MSD	BM050014.D	04/23/2025
COMP-1	Q1858-01	BM050015.D	04/23/2025
COMP-2	Q1858-02	BM050016.D	04/23/2025
COMP-3	Q1858-03	BM050017.D	04/23/2025

COMMENTS: _____

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: POWE02

Lab Code: CHEM

SAS No.: Q1858 SDG NO.: Q1858

Lab File ID: BM049847.D

DFTPP Injection Date: 04/08/2025

Instrument ID: BNA_M

DFTPP Injection Time: 12:55

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	22.1
68	Less than 2.0% of mass 69	0.3 (1.2) 1
69	Mass 69 relative abundance	26.5
70	Less than 2.0% of mass 69	0.1 (0.3) 1
127	10.0 - 80.0% of mass 198	34.9
197	Less than 2.0% of mass 198	0.4
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	7
275	10.0 - 60.0% of mass 198	26.5
365	Greater than 1% of mass 198	3.7
441	Present, but less than mass 443	12.4
442	Greater than 50% of mass 198	81.2
443	15.0 - 24.0% of mass 442	15.3 (18.8) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BM049848.D	04/08/2025	13:35
SSTDICC005	SSTDICC005	BM049849.D	04/08/2025	14:14
SSTDICC010	SSTDICC010	BM049850.D	04/08/2025	14:53
SSTDICC020	SSTDICC020	BM049851.D	04/08/2025	15:32
SSTDICCC040	SSTDICCC040	BM049852.D	04/08/2025	16:12
SSTDICC050	SSTDICC050	BM049853.D	04/08/2025	17:30
SSTDICC060	SSTDICC060	BM049854.D	04/08/2025	19:28
SSTDICC080	SSTDICC080	BM049855.D	04/08/2025	20:07

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: POWE02

Lab Code: CHEM

SAS No.: Q1858 SDG NO.: Q1858

Lab File ID: BM050006.D

DFTPP Injection Date: 04/23/2025

Instrument ID: BNA_M

DFTPP Injection Time: 09:18

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	21
68	Less than 2.0% of mass 69	0.4 (1.6) 1
69	Mass 69 relative abundance	25.4
70	Less than 2.0% of mass 69	0.1 (0.6) 1
127	10.0 - 80.0% of mass 198	32.7
197	Less than 2.0% of mass 198	0.4
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.9
275	10.0 - 60.0% of mass 198	26.3
365	Greater than 1% of mass 198	3.4
441	Present, but less than mass 443	11.7
442	Greater than 50% of mass 198	75.4
443	15.0 - 24.0% of mass 442	14.7 (19.5) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BM050007.D	04/23/2025	10:38
72-12013MS	Q1852-07MS	BM050013.D	04/23/2025	14:37
72-12013MSD	Q1852-07MSD	BM050014.D	04/23/2025	15:16
COMP-1	Q1858-01	BM050015.D	04/23/2025	15:55
COMP-2	Q1858-02	BM050016.D	04/23/2025	16:35
COMP-3	Q1858-03	BM050017.D	04/23/2025	17:14

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: POWE02

Lab Code: CHEM

SAS No.: Q1858

SDG NO.: Q1858

Lab File ID: BP024274.D

DFTPP Injection Date: 04/14/2025

Instrument ID: BNA_P

DFTPP Injection Time: 10:25

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	30.4
68	Less than 2.0% of mass 69	0.7 (1.8) 1
69	Mass 69 relative abundance	36
70	Less than 2.0% of mass 69	0.2 (0.5) 1
127	10.0 - 80.0% of mass 198	48.5
197	Less than 2.0% of mass 198	0.5
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.9
275	10.0 - 60.0% of mass 198	29.1
365	Greater than 1% of mass 198	4.1
441	Present, but less than mass 443	15.2
442	Greater than 50% of mass 198	99.1
443	15.0 - 24.0% of mass 442	19.2 (19.4) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BP024275.D	04/14/2025	11:06
SSTDICC005	SSTDICC005	BP024276.D	04/14/2025	11:47
SSTDICC010	SSTDICC010	BP024277.D	04/14/2025	12:27
SSTDICC020	SSTDICC020	BP024278.D	04/14/2025	13:08
SSTDICCC040	SSTDICCC040	BP024279.D	04/14/2025	13:49
SSTDICC050	SSTDICC050	BP024280.D	04/14/2025	15:10
SSTDICC060	SSTDICC060	BP024281.D	04/14/2025	16:32
SSTDICC080	SSTDICC080	BP024282.D	04/14/2025	17:13

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: POWE02

Lab Code: CHEM

SAS No.: Q1858

SDG NO.: Q1858

Lab File ID: BP024406.D

DFTPP Injection Date: 04/24/2025

Instrument ID: BNA_P

DFTPP Injection Time: 11:10

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	26.2
68	Less than 2.0% of mass 69	0.5 (1.7) 1
69	Mass 69 relative abundance	30.1
70	Less than 2.0% of mass 69	0.1 (0.5) 1
127	10.0 - 80.0% of mass 198	42.2
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6
275	10.0 - 60.0% of mass 198	27.3
365	Greater than 1% of mass 198	4.1
441	Present, but less than mass 443	15.3
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.3 (19.3) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BP024407.D	04/24/2025	11:50
PB167711BL	PB167711BL	BP024408.D	04/24/2025	12:31
PB167711BS	PB167711BS	BP024409.D	04/24/2025	13:12

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1858 SAS No.: Q1858 SDG NO.: Q1858

EPA Sample No.: SSTDCCC040 Date Analyzed: 04/23/2025

Lab File ID: BM050007.D Time Analyzed: 10:38

Instrument ID: BNA_M GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	307209	7.763	1029600	10.56	636547	14.41
UPPER LIMIT	614418	8.263	2059200	11.057	1273090	14.91
LOWER LIMIT	153605	7.263	514800	10.057	318274	13.91
EPA SAMPLE NO.						
01 72-12013MS	287884	7.76	952796	10.56	566920	14.41
02 72-12013MSD	270803	7.76	919060	10.56	576233	14.41
03 COMP-1	268192	7.76	920312	10.56	612241	14.41
04 COMP-2	281365	7.76	963050	10.56	635987	14.41
05 COMP-3	296499	7.76	1032150	10.56	691467	14.41

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1858 SAS No.: Q1858 SDG NO.: Q1858

EPA Sample No.: SSTDCCC040 Date Analyzed: 04/23/2025

Lab File ID: BM050007.D Time Analyzed: 10:38

Instrument ID: BNA_M GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	1225580	17.157	1143350	21.398	1143260	24.403
UPPER LIMIT	2451160	17.657	2286700	21.898	2286520	24.903
LOWER LIMIT	612790	16.657	571675	20.898	571630	23.903
EPA SAMPLE NO.						
01 72-12013MS	1079720	17.16	1168530	21.40	1272700	24.40
02 72-12013MSD	1111720	17.16	1219360	21.40	1284690	24.40
03 COMP-1	1242940	17.16	1324640	21.40	1398710	24.40
04 COMP-2	1306190	17.16	1368420	21.40	1421970	24.40
05 COMP-3	1437930	17.16	1425720	21.40	1479400	24.40

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1858 SAS No.: Q1858 SDG NO.: Q1858

EPA Sample No.: SSTDCCC040 Date Analyzed: 04/24/2025

Lab File ID: BP024407.D Time Analyzed: 11:50

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	264641	7.716	1170080	10.49	779721	14.35
UPPER LIMIT	529282	8.216	2340160	10.987	1559440	14.845
LOWER LIMIT	132321	7.216	585040	9.987	389861	13.845
EPA SAMPLE NO.						
01 PB167711BL	288156	7.72	1116360	10.49	654539	14.34
02 PB167711BS	253522	7.72	1075220	10.49	697166	14.35

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1858 SAS No.: Q1858 SDG NO.: Q1858

EPA Sample No.: SSTDCCC040 Date Analyzed: 04/24/2025

Lab File ID: BP024407.D Time Analyzed: 11:50

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	1570090	17.145	1711310	21.592	1789710	24.91
UPPER LIMIT	3140180	17.645	3422620	22.092	3579420	25.41
LOWER LIMIT	785045	16.645	855655	21.092	894855	24.41
EPA SAMPLE NO.						
01 PB167711BL	1181170	17.13	1074960	21.57	1210800	24.92
02 PB167711BS	1340060	17.15	1256050	21.59	1241820	24.90

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	Kleinfelder	Date Collected:	
Project:	Henry Lea School	Date Received:	
Client Sample ID:	PB167711BL	SDG No.:	Q1858
Lab Sample ID:	PB167711BL	Matrix:	SOIL
Analytical Method:	SW8270	% Solid:	100
Sample Wt/Vol:	30.01 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BP024408.D	1	04/23/25 09:20	04/24/25 12:31	PB167711

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
91-20-3	Naphthalene	22.7	U	22.7	170	ug/Kg
86-73-7	Fluorene	25.3	U	25.3	170	ug/Kg
85-01-8	Phenanthrene	20.9	U	20.9	170	ug/Kg
120-12-7	Anthracene	33.3	U	33.3	170	ug/Kg
129-00-0	Pyrene	36.0	U	36.0	170	ug/Kg
56-55-3	Benzo(a)anthracene	23.0	U	23.0	170	ug/Kg
218-01-9	Chrysene	19.9	U	19.9	170	ug/Kg
205-99-2	Benzo(b)fluoranthene	19.0	U	19.0	170	ug/Kg
50-32-8	Benzo(a)pyrene	29.5	U	29.5	170	ug/Kg
193-39-5	Indeno(1,2,3-cd)pyrene	29.1	U	29.1	170	ug/Kg
191-24-2	Benzo(g,h,i)perylene	25.7	U	25.7	170	ug/Kg
SURROGATES						
4165-60-0	Nitrobenzene-d5	91.5		18 - 107	92%	SPK: 100
321-60-8	2-Fluorobiphenyl	92.2		20 - 109	92%	SPK: 100
1718-51-0	Terphenyl-d14	98.5		10 - 105	98%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	288000	7.722			
1146-65-2	Naphthalene-d8	1120000	10.486			
15067-26-2	Acenaphthene-d10	655000	14.339			
1517-22-2	Phenanthrene-d10	1180000	17.133			
1719-03-5	Chrysene-d12	1070000	21.574			
1520-96-3	Perylene-d12	1210000	24.915			

Report of Analysis

Client:	Kleinfelder	Date Collected:	
Project:	Henry Lea School	Date Received:	
Client Sample ID:	PB167711BL	SDG No.:	Q1858
Lab Sample ID:	PB167711BL	Matrix:	SOIL
Analytical Method:	SW8270	% Solid:	100
Sample Wt/Vol:	30.01 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BP024408.D	1	04/23/25 09:20	04/24/25 12:31	PB167711

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Kleinfelder	Date Collected:	
Project:	Henry Lea School	Date Received:	
Client Sample ID:	PB167711BS	SDG No.:	Q1858
Lab Sample ID:	PB167711BS	Matrix:	SOIL
Analytical Method:	SW8270	% Solid:	100
Sample Wt/Vol:	30.03 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BP024409.D	1	04/23/25 09:20	04/24/25 13:12	PB167711

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
------------	-----------	-------	-----------	-----	------------	-------------------

TARGETS

91-20-3	Naphthalene	1500		22.7	170	ug/Kg
86-73-7	Fluorene	1500		25.3	170	ug/Kg
85-01-8	Phenanthrene	1500		20.9	170	ug/Kg
120-12-7	Anthracene	1600		33.3	170	ug/Kg
129-00-0	Pyrene	1700		36.0	170	ug/Kg
56-55-3	Benzo(a)anthracene	1600		23.0	170	ug/Kg
218-01-9	Chrysene	1500		19.9	170	ug/Kg
205-99-2	Benzo(b)fluoranthene	1600		19.0	170	ug/Kg
50-32-8	Benzo(a)pyrene	1700		29.5	170	ug/Kg
193-39-5	Indeno(1,2,3-cd)pyrene	1600		29.1	170	ug/Kg
191-24-2	Benzo(g,h,i)perylene	1500		25.7	170	ug/Kg

SURROGATES

4165-60-0	Nitrobenzene-d5	90.4		18 - 107	90%	SPK: 100
321-60-8	2-Fluorobiphenyl	89.9		20 - 109	90%	SPK: 100
1718-51-0	Terphenyl-d14	101		10 - 105	101%	SPK: 100

INTERNAL STANDARDS

3855-82-1	1,4-Dichlorobenzene-d4	254000	7.722
1146-65-2	Naphthalene-d8	1080000	10.493
15067-26-2	Acenaphthene-d10	697000	14.351
1517-22-2	Phenanthrene-d10	1340000	17.145
1719-03-5	Chrysene-d12	1260000	21.592
1520-96-3	Perylene-d12	1240000	24.898

Report of Analysis

Client:	Kleinfelder	Date Collected:	
Project:	Henry Lea School	Date Received:	
Client Sample ID:	PB167711BS	SDG No.:	Q1858
Lab Sample ID:	PB167711BS	Matrix:	SOIL
Analytical Method:	SW8270	% Solid:	100
Sample Wt/Vol:	30.03 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BP024409.D	1	04/23/25 09:20	04/24/25 13:12	PB167711

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Kleinfelder	Date Collected:	04/22/25
Project:	Henry Lea School	Date Received:	04/22/25
Client Sample ID:	72-12013MS	SDG No.:	Q1858
Lab Sample ID:	Q1852-07MS	Matrix:	SOIL
Analytical Method:	SW8270	% Solid:	89.2
Sample Wt/Vol:	50.07 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM050013.D	1	04/23/25 09:20	04/23/25 14:37	PB167711

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
------------	-----------	-------	-----------	-----	------------	-------------------

TARGETS

91-20-3	Naphthalene	990		15.2	110	ug/Kg
86-73-7	Fluorene	1000		17.0	110	ug/Kg
85-01-8	Phenanthrene	1100		14.0	110	ug/Kg
120-12-7	Anthracene	1100		22.4	110	ug/Kg
129-00-0	Pyrene	1000		24.2	110	ug/Kg
56-55-3	Benzo(a)anthracene	1100		15.4	110	ug/Kg
218-01-9	Chrysene	1100		13.4	110	ug/Kg
205-99-2	Benzo(b)fluoranthene	1100		12.8	110	ug/Kg
50-32-8	Benzo(a)pyrene	1200		19.8	110	ug/Kg
193-39-5	Indeno(1,2,3-cd)pyrene	1100		19.5	110	ug/Kg
191-24-2	Benzo(g,h,i)perylene	1000		17.3	110	ug/Kg

SURROGATES

4165-60-0	Nitrobenzene-d5	69.4		18 - 107	69%	SPK: 100
321-60-8	2-Fluorobiphenyl	67.1		20 - 109	67%	SPK: 100
1718-51-0	Terphenyl-d14	65.1		10 - 105	65%	SPK: 100

INTERNAL STANDARDS

3855-82-1	1,4-Dichlorobenzene-d4	288000	7.763
1146-65-2	Naphthalene-d8	953000	10.557
15067-26-2	Acenaphthene-d10	567000	14.41
1517-22-2	Phenanthrene-d10	1080000	17.157
1719-03-5	Chrysene-d12	1170000	21.398
1520-96-3	Perylene-d12	1270000	24.397

Report of Analysis

Client:	Kleinfelder	Date Collected:	04/22/25
Project:	Henry Lea School	Date Received:	04/22/25
Client Sample ID:	72-12013MS	SDG No.:	Q1858
Lab Sample ID:	Q1852-07MS	Matrix:	SOIL
Analytical Method:	SW8270	% Solid:	89.2
Sample Wt/Vol:	50.07 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM050013.D	1	04/23/25 09:20	04/23/25 14:37	PB167711

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Kleinfelder	Date Collected:	04/22/25
Project:	Henry Lea School	Date Received:	04/22/25
Client Sample ID:	72-12013MSD	SDG No.:	Q1858
Lab Sample ID:	Q1852-07MSD	Matrix:	SOIL
Analytical Method:	SW8270	% Solid:	89.2
Sample Wt/Vol:	50.06 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM050014.D	1	04/23/25 09:20	04/23/25 15:16	PB167711

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
------------	-----------	-------	-----------	-----	------------	-------------------

TARGETS

91-20-3	Naphthalene	970		15.3	110	ug/Kg
86-73-7	Fluorene	990		17.0	110	ug/Kg
85-01-8	Phenanthrene	1000		14.0	110	ug/Kg
120-12-7	Anthracene	1100		22.4	110	ug/Kg
129-00-0	Pyrene	990		24.2	110	ug/Kg
56-55-3	Benzo(a)anthracene	1100		15.5	110	ug/Kg
218-01-9	Chrysene	1000		13.4	110	ug/Kg
205-99-2	Benzo(b)fluoranthene	1000		12.8	110	ug/Kg
50-32-8	Benzo(a)pyrene	1100		19.8	110	ug/Kg
193-39-5	Indeno(1,2,3-cd)pyrene	1100		19.6	110	ug/Kg
191-24-2	Benzo(g,h,i)perylene	970		17.3	110	ug/Kg

SURROGATES

4165-60-0	Nitrobenzene-d5	68.6		18 - 107	69%	SPK: 100
321-60-8	2-Fluorobiphenyl	63.7		20 - 109	64%	SPK: 100
1718-51-0	Terphenyl-d14	63.1		10 - 105	63%	SPK: 100

INTERNAL STANDARDS

3855-82-1	1,4-Dichlorobenzene-d4	271000	7.763
1146-65-2	Naphthalene-d8	919000	10.557
15067-26-2	Acenaphthene-d10	576000	14.41
1517-22-2	Phenanthrene-d10	1110000	17.157
1719-03-5	Chrysene-d12	1220000	21.398
1520-96-3	Perylene-d12	1280000	24.403

Report of Analysis

Client:	Kleinfelder	Date Collected:	04/22/25
Project:	Henry Lea School	Date Received:	04/22/25
Client Sample ID:	72-12013MSD	SDG No.:	Q1858
Lab Sample ID:	Q1852-07MSD	Matrix:	SOIL
Analytical Method:	SW8270	% Solid:	89.2
Sample Wt/Vol:	50.06 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BM050014.D	1	04/23/25 09:20	04/23/25 15:16	PB167711

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



CALIBRATION SUMMARY

Method Path : Z:\svoasrv\HPCHEM1\BNA_P\Methods\
 Method File : 8270E-BP041425.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Apr 15 04:48:42 2025
 Response Via : Initial Calibration

Calibration Files

2.5 =BP024275.D 5 =BP024276.D 10 =BP024277.D 20 =BP024278.D 40 =BP024279.D 50 =BP024280.D 60 =BP024281.D 80 =BP024282.D

	Compound	2.5	5	10	20	40	50	60	80	Avg	%RSD
1) I	1,4-Dichlorobenzen...	-----ISTD-----									
2)	1,4-Dioxane		0.551	0.517	0.552	0.512	0.481	0.490	0.477	0.512	6.10
3)	Pyridine		1.265	1.280	1.422	1.396	1.333	1.383	1.378	1.351	4.43
4)	n-Nitrosodimet...		0.424	0.443	0.469	0.457	0.442	0.460	0.444	0.448	3.29
5) S	2-Fluorophenol		1.123	1.145	1.262	1.254	1.210	1.277	1.196	1.210	4.93
6)	Aniline		1.439	1.487	1.636	1.587	1.403	1.428	1.380	1.480	6.52
7) S	Phenol-d6		1.463	1.552	1.718	1.748	1.697	1.769	1.647	1.656	6.75
8)	2-Chlorophenol		1.260	1.297	1.412	1.399	1.359	1.400	1.328	1.350	4.30
9)	Benzaldehyde		0.943	0.934	0.942	0.836	0.732	0.726	0.621	0.819	15.70
10) C	Phenol		1.475	1.565	1.720	1.750	1.698	1.779	1.643	1.661	6.54
11)	bis(2-Chloroet...		1.278	1.317	1.399	1.398	1.321	1.358	1.300	1.339	3.56
12)	1,3-Dichlorobe...		1.464	1.486	1.534	1.486	1.402	1.433	1.373	1.454	3.79
13) C	1,4-Dichlorobe...		1.496	1.489	1.550	1.494	1.436	1.463	1.399	1.475	3.28
14)	1,2-Dichlorobe...		1.472	1.441	1.502	1.445	1.362	1.411	1.339	1.424	4.10
15)	Benzyl Alcohol		0.857	0.935	1.106	1.158	1.142	1.178	1.122	1.071	11.58
16)	2,2'-oxybis(1-...		1.452	1.496	1.517	1.514	1.405	1.397	1.348	1.447	4.54
17)	2-Methylphenol		0.915	1.002	1.123	1.140	1.105	1.150	1.087	1.075	7.97
18)	Hexachloroethane		0.532	0.537	0.558	0.542	0.521	0.533	0.510	0.533	2.88
19) P	n-Nitroso-di-n...	0.941	1.008	1.008	1.088	1.084	1.041	1.038	0.989	1.025	4.76
20)	3+4-Methylphenols		1.224	1.365	1.560	1.592	1.567	1.607	1.525	1.491	9.57
21) I	Naphthalene-d8	-----ISTD-----									
22)	Acetophenone		0.470	0.487	0.522	0.519	0.490	0.495	0.483	0.495	3.85
23) S	Nitrobenzene-d5		0.322	0.337	0.367	0.368	0.357	0.356	0.347	0.351	4.76
24)	Nitrobenzene		0.328	0.334	0.362	0.364	0.349	0.353	0.340	0.347	3.96
25)	Isophorone		0.575	0.601	0.653	0.673	0.650	0.654	0.639	0.635	5.41
26) C	2-Nitrophenol		0.130	0.143	0.172	0.183	0.184	0.191	0.186	0.170	13.97
27)	2,4-Dimethylph...		0.180	0.194	0.219	0.226	0.225	0.228	0.221	0.213	8.81
28)	bis(2-Chloroet...		0.412	0.426	0.444	0.452	0.430	0.425	0.412	0.429	3.51
29) C	2,4-Dichloroph...		0.242	0.268	0.297	0.308	0.307	0.311	0.302	0.291	8.94
30)	1,2,4-Trichlor...		0.308	0.305	0.318	0.321	0.309	0.315	0.306	0.311	2.00
31)	Naphthalene		1.044	1.054	1.091	1.079	1.037	1.048	1.023	1.054	2.27
32)	Benzoic acid			0.185	0.226	0.242	0.270	0.285	0.282	0.248	15.56
33)	4-Chloroaniline		0.328	0.349	0.389	0.397	0.378	0.379	0.378	0.371	6.46
34) C	Hexachlorobuta...		0.181	0.181	0.186	0.189	0.181	0.182	0.182	0.183	1.62
35)	Caprolactam		0.085	0.092	0.111	0.112	0.115	0.121	0.115	0.107	12.53
36) C	4-Chloro-3-met...		0.284	0.310	0.349	0.357	0.356	0.363	0.351	0.339	8.82
37)	2-Methylnaphth...		0.700	0.720	0.755	0.755	0.733	0.740	0.712	0.731	2.89
38)	1-Methylnaphth...		0.700	0.714	0.737	0.731	0.712	0.718	0.691	0.715	2.25

Method Path : Z:\svoasrv\HPCHEM1\BNA_P\Methods\
 Method File : 8270E-BP041425.M

39) I	Acenaphthene-d10	-----ISTD-----									
40)	1,2,4,5-Tetrac...	0.522	0.538	0.568	0.576	0.544	0.549	0.552	0.550	3.33	
41) P	Hexachlorocycl...	0.162	0.185	0.197	0.218	0.199	0.197	0.204	0.195	8.92	
42) S	2,4,6-Tribromo...	0.235	0.256	0.278	0.289	0.289	0.296	0.294	0.277	8.29	
43) C	2,4,6-Trichlor...	0.301	0.335	0.369	0.389	0.386	0.390	0.390	0.366	9.51	
44)	2,4,5-Trichlor...	0.332	0.371	0.409	0.427	0.429	0.436	0.430	0.405	9.62	
45) S	2-Fluorobiphenyl	1.332	1.334	1.359	1.366	1.263	1.251	1.240	1.306	4.08	
46)	1,1'-Biphenyl	1.454	1.489	1.528	1.529	1.434	1.428	1.429	1.470	3.07	
47)	2-Chloronaphth...	1.063	1.100	1.135	1.139	1.085	1.090	1.078	1.099	2.60	
48)	2-Nitroaniline	0.250	0.270	0.317	0.330	0.324	0.335	0.331	0.308	10.98	
49)	Acenaphthylene	1.604	1.680	1.793	1.814	1.726	1.763	1.730	1.730	4.11	
50)	Dimethylphthalate	1.429	1.416	1.495	1.495	1.460	1.460	1.426	1.454	2.22	
51)	2,6-Dinitrotol...	0.264	0.290	0.319	0.325	0.319	0.328	0.317	0.309	7.58	
52) C	Acenaphthene	1.087	1.087	1.141	1.126	1.081	1.079	1.076	1.097	2.35	
53)	3-Nitroaniline	0.250	0.289	0.338	0.354	0.342	0.340	0.358	0.325	12.25	
54) P	2,4-Dinitrophenol		0.125	0.163	0.186	0.196	0.210	0.211	0.182	18.22	
55)	Dibenzofuran	1.801	1.807	1.871	1.833	1.746	1.768	1.723	1.793	2.84	
56) P	4-Nitrophenol	0.190	0.231	0.289	0.305	0.306	0.323	0.317	0.280	17.90	
57)	2,4-Dinitrotol...	0.334	0.368	0.439	0.448	0.442	0.463	0.456	0.421	11.82	
58)	Fluorene	1.376	1.397	1.454	1.398	1.368	1.366	1.356	1.388	2.40	
59)	2,3,4,6-Tetrac...	0.327	0.351	0.380	0.389	0.387	0.390	0.389	0.373	6.57	
60)	Diethylphthalate	1.451	1.484	1.528	1.550	1.499	1.499	1.481	1.499	2.16	
61)	4-Chlorophenyl...	0.667	0.675	0.699	0.680	0.667	0.670	0.659	0.674	1.95	
62)	4-Nitroaniline	0.263	0.300	0.353	0.369	0.364	0.380	0.376	0.344	12.99	
63)	Azobenzene	1.269	1.383	1.446	1.431	1.385	1.385	1.344	1.378	4.26	
64) I	Phenanthrene-d10	-----ISTD-----									
65)	4,6-Dinitro-2-...		0.095	0.118	0.131	0.134	0.140	0.138	0.126	13.50	
66) c	n-Nitrosodiphe...	0.567	0.585	0.616	0.630	0.591	0.608	0.580	0.597	3.71	
67)	4-Bromophenyl-...	0.203	0.212	0.220	0.230	0.220	0.225	0.222	0.219	4.11	
68)	Hexachlorobenzene	0.247	0.247	0.260	0.271	0.261	0.269	0.263	0.260	3.67	
69)	Atrazine	0.176	0.167	0.134	0.121	0.162			0.152	15.57	
70) C	Pentachlorophenol	0.143	0.155	0.175	0.196	0.197	0.203	0.201	0.181	13.35	
71)	Phenanthrene	1.098	1.086	1.117	1.127	1.070	1.082	1.058	1.091	2.27	
72)	Anthracene	0.995	1.017	1.084	1.111	1.053	1.071	1.030	1.052	3.85	
73)	Carbazole	0.968	0.998	1.060	1.084	1.014	1.037	1.031	1.027	3.78	
74)	Di-n-butylphth...	1.147	1.230	1.253	1.398	1.310	1.338	1.284	1.280	6.30	
75) C	Fluoranthene	1.289	1.274	1.316	1.347	1.280	1.304	1.274	1.298	2.07	
76) I	Chrysene-d12	-----ISTD-----									
77)	Benzidine	0.144	0.141	0.109	0.436	0.333	0.302	0.307	0.254	48.39	
78)	Pyrene	1.192	1.236	1.345	1.305	1.294	1.305	1.219	1.271	4.38	
79) S	Terphenyl-d14	0.967	0.998	1.055	1.032	1.005	0.964	0.926	0.992	4.42	
80)	Butylbenzylphth...	0.434	0.488	0.546	0.596	0.580	0.593	0.574	0.544	11.25	
81)	Benzo(a)anthra...	1.195	1.214	1.292	1.293	1.234	1.253	1.221	1.243	3.08	
82)	3,3'-Dichlorob...	0.345	0.383	0.431	0.486	0.463	0.485	0.460	0.436	12.30	
83)	Chrysene	1.180	1.181	1.229	1.213	1.173	1.193	1.167	1.191	1.88	
84)	Bis(2-ethylhex...	0.639	0.735	0.779	0.893	0.853	0.867	0.819	0.798	11.09	
85) c	Di-n-octyl pht...	0.945	1.093	1.222	1.461	1.430	1.486	1.446	1.297	16.46	

Method Path : Z:\svoasrv\HPCHEM1\BNA_P\Methods\
Method File : 8270E-BP041425.M

86) I	Perylene-d12	-----ISTD-----									
87)	Indeno(1,2,3-c...	1.252	1.304	1.407	1.444	1.401	1.445	1.466	1.388	5.77	
88)	Benzo(b)fluora...	1.130	1.174	1.217	1.258	1.196	1.205	1.212	1.199	3.30	
89)	Benzo(k)fluora...	1.088	1.131	1.217	1.212	1.154	1.166	1.138	1.158	3.94	
90) C	Benzo(a)pyrene	0.939	0.971	1.057	1.091	1.053	1.070	1.058	1.034	5.44	
91)	Dibenzo(a,h)an...	1.046	1.089	1.172	1.203	1.169	1.186	1.202	1.152	5.28	
92)	Benzo(g,h,i)pe...	1.073	1.110	1.198	1.218	1.180	1.209	1.224	1.173	4.99	

(#) = Out of Range

Method Path : Z:\svoasrv\HPCHEM1\BNA_M\Methods\
 Method File : 8270-BM040825.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Apr 09 04:00:55 2025
 Response Via : Initial Calibration

Calibration Files

2.5 =BM049848.D 5 =BM049849.D 10 =BM049850.D 20 =BM049851.D 40 =BM049852.D 50 =BM049853.D 60 =BM049854.D 80 =BM049855.D

Compound	2.5	5	10	20	40	50	60	80	Avg	%RSD

1) I 1,4-Dichlorobenzen...	-----ISTD-----									
2) 1,4-Dioxane		0.497	0.470	0.501	0.488	0.486	0.471	0.483	0.485	2.47
3) Pyridine		1.262	1.220	1.322	1.287	1.276	1.278	1.276	1.274	2.40
4) n-Nitrosodimet...		0.478	0.461	0.496	0.489	0.484	0.500	0.488	0.485	2.67
5) S 2-Fluorophenol		1.145	1.144	1.213	1.191	1.165	1.210	1.177	1.178	2.40
6) Aniline		1.304	1.315	1.396	1.325	1.153	1.142	1.154	1.256	8.22
7) S Phenol-d6		1.391	1.392	1.523	1.491	1.433	1.543	1.485	1.465	4.15
8) 2-Chlorophenol		1.180	1.191	1.250	1.218	1.174	1.239	1.199	1.207	2.42
9) Benzaldehyde		0.873	0.830	0.838	0.750	0.712	0.701	0.561	0.752	14.25
10) C Phenol		1.376	1.370	1.482	1.441	1.394	1.496	1.451	1.430	3.56
11) bis(2-Chloroet...		1.113	1.084	1.147	1.135	1.096	1.160	1.114	1.121	2.45
12) 1,3-Dichlorobe...		1.417	1.391	1.475	1.438	1.418	1.441	1.409	1.427	1.90
13) C 1,4-Dichlorobe...		1.418	1.399	1.488	1.453	1.412	1.455	1.426	1.436	2.15
14) 1,2-Dichlorobe...		1.348	1.325	1.402	1.374	1.324	1.359	1.334	1.352	2.12
15) Benzyl Alcohol		0.852	0.857	0.950	0.953	0.904	0.993	0.951	0.923	5.79
16) 2,2'-oxybis(1-...		1.268	1.235	1.312	1.267	1.192	1.261	1.214	1.250	3.18
17) 2-Methylphenol		0.858	0.856	0.916	0.891	0.850	0.920	0.877	0.881	3.29
18) Hexachloroethane		0.505	0.494	0.520	0.504	0.486	0.501	0.488	0.500	2.32
19) P n-Nitroso-di-n...	0.760	0.786	0.783	0.860	0.836	0.793	0.859	0.816	0.812	4.60
20) 3+4-Methylphenols		1.140	1.142	1.246	1.224	1.170	1.272	1.215	1.201	4.31

21) I Naphthalene-d8	-----ISTD-----									
22) Acetophenone		0.463	0.477	0.496	0.498	0.482	0.494	0.493	0.486	2.64
23) S Nitrobenzene-d5		0.335	0.344	0.366	0.368	0.361	0.370	0.371	0.359	3.96
24) Nitrobenzene		0.329	0.329	0.354	0.356	0.345	0.353	0.354	0.346	3.49
25) Isophorone		0.569	0.568	0.613	0.616	0.598	0.628	0.620	0.602	4.05
26) C 2-Nitrophenol		0.162	0.165	0.183	0.189	0.190	0.197	0.197	0.183	7.91
27) 2,4-Dimethylph...		0.185	0.194	0.208	0.213	0.212	0.221	0.221	0.208	6.48
28) bis(2-Chloroet...		0.386	0.386	0.411	0.410	0.399	0.415	0.412	0.403	3.08
29) C 2,4-Dichloroph...		0.319	0.320	0.348	0.353	0.349	0.364	0.364	0.345	5.45
30) 1,2,4-Trichlor...		0.386	0.380	0.401	0.404	0.399	0.412	0.413	0.399	3.11
31) Naphthalene		1.004	0.990	1.045	1.042	1.019	1.050	1.045	1.028	2.32
32) Benzoic acid			0.109	0.175	0.230	0.240	0.283	0.283	0.220	30.63
33) 4-Chloroaniline		0.348	0.351	0.376	0.380	0.359	0.359	0.366	0.363	3.34
34) C Hexachlorobuta...		0.238	0.236	0.245	0.249	0.249	0.255	0.258	0.247	3.33
35) Caprolactam		0.093	0.089	0.100	0.102	0.099	0.107	0.104	0.099	6.29
36) C 4-Chloro-3-met...		0.277	0.276	0.305	0.309	0.303	0.326	0.321	0.302	6.41
37) 2-Methylnaphth...		0.683	0.685	0.729	0.734	0.721	0.762	0.755	0.724	4.23
38) 1-Methylnaphth...		0.680	0.659	0.711	0.713	0.703	0.739	0.732	0.705	3.97

Method Path : Z:\svoasrv\HPCHEM1\BNA_M\Methods\
 Method File : 8270-BM040825.M

39) I	Acenaphthene-d10	-----ISTD-----									
40)	1,2,4,5-Tetrac...	0.664	0.662	0.697	0.706	0.711	0.721	0.737	0.700	4.02	
41) P	Hexachlorocycl...	0.214	0.223	0.247	0.264	0.260	0.258	0.251	0.245	7.86	
42) S	2,4,6-Tribromo...	0.258	0.255	0.280	0.294	0.298	0.323	0.328	0.291	9.94	
43) C	2,4,6-Trichlor...	0.417	0.420	0.444	0.459	0.449	0.459	0.468	0.445	4.46	
44)	2,4,5-Trichlor...	0.465	0.452	0.472	0.500	0.483	0.504	0.510	0.484	4.52	
45) S	2-Fluorobiphenyl	1.425	1.414	1.499	1.515	1.486	1.482	1.504	1.475	2.68	
46)	1,1'-Biphenyl	1.451	1.429	1.514	1.528	1.480	1.490	1.515	1.487	2.45	
47)	2-Chloronaphth...	1.140	1.144	1.188	1.194	1.163	1.162	1.189	1.169	1.88	
48)	2-Nitroaniline	0.232	0.242	0.272	0.283	0.281	0.286	0.295	0.270	8.83	
49)	Acenaphthylene	1.627	1.613	1.718	1.742	1.711	1.735	1.768	1.702	3.46	
50)	Dimethylphthalate	1.366	1.331	1.430	1.438	1.411	1.446	1.455	1.411	3.27	
51)	2,6-Dinitrotol...	0.249	0.265	0.303	0.310	0.305	0.317	0.320	0.296	9.38	
52) C	Acenaphthene	1.027	1.023	1.089	1.113	1.089	1.110	1.127	1.083	3.83	
53)	3-Nitroaniline	0.236	0.253	0.290	0.310	0.295	0.295	0.321	0.286	10.64	
54) P	2,4-Dinitrophenol		0.097	0.148	0.184	0.194	0.212	0.216	0.175	26.04	
55)	Dibenzofuran	1.737	1.709	1.821	1.845	1.815	1.862	1.879	1.810	3.52	
56) P	4-Nitrophenol	0.195	0.213	0.257	0.270	0.273	0.286	0.288	0.255	14.34	
57)	2,4-Dinitrotol...	0.309	0.333	0.395	0.421	0.420	0.440	0.443	0.394	13.45	
58)	Fluorene	1.354	1.335	1.450	1.481	1.456	1.489	1.505	1.439	4.67	
59)	2,3,4,6-Tetrac...	0.418	0.397	0.416	0.427	0.420	0.440	0.452	0.424	4.21	
60)	Diethylphthalate	1.311	1.270	1.378	1.390	1.347	1.378	1.397	1.353	3.46	
61)	4-Chlorophenyl...	0.699	0.694	0.761	0.789	0.791	0.830	0.836	0.771	7.43	
62)	4-Nitroaniline	0.227	0.247	0.301	0.320	0.316	0.327	0.331	0.295	14.02	
63)	Azobenzene	1.077	1.085	1.185	1.202	1.167	1.192	1.200	1.158	4.67	
64) I	Phenanthrene-d10	-----ISTD-----									
65)	4,6-Dinitro-2-...		0.094	0.119	0.132	0.133	0.138	0.140	0.126	13.60	
66) c	n-Nitrosodiphe...	0.561	0.578	0.606	0.619	0.602	0.610	0.614	0.599	3.54	
67)	4-Bromophenyl-...	0.212	0.212	0.227	0.239	0.237	0.247	0.252	0.232	6.83	
68)	Hexachlorobenzene	0.249	0.250	0.262	0.273	0.268	0.279	0.284	0.266	5.08	
69)	Atrazine	0.186	0.177	0.142	0.130	0.142			0.155	15.89	
70) C	Pentachlorophenol	0.181	0.186	0.191	0.204	0.197	0.204	0.210	0.196	5.46	
71)	Phenanthrene	1.031	1.030	1.095	1.133	1.104	1.135	1.147	1.096	4.41	
72)	Anthracene	0.997	1.004	1.074	1.125	1.099	1.125	1.140	1.081	5.41	
73)	Carbazole	0.922	0.952	1.024	1.059	1.038	1.060	1.071	1.018	5.69	
74)	Di-n-butylphth...	1.119	1.120	1.189	1.216	1.176	1.191	1.196	1.172	3.22	
75) C	Fluoranthene	1.187	1.206	1.310	1.387	1.395	1.463	1.488	1.348	8.78	
76) I	Chrysene-d12	-----ISTD-----									
77)	Benzidine	0.319	0.260	0.205	0.403	0.163	0.150	0.357	0.265	36.93	
78)	Pyrene	1.310	1.280	1.395	1.408	1.393	1.419	1.443	1.378	4.35	
79) S	Terphenyl-d14	0.988	1.004	1.139	1.187	1.142	1.007	1.004	1.067	7.91	
80)	Butylbenzylpht...	0.502	0.500	0.537	0.536	0.511	0.516	0.523	0.518	2.88	
81)	Benzo(a)anthra...	1.237	1.226	1.335	1.362	1.345	1.391	1.408	1.329	5.38	
82)	3,3'-Dichlorob...	0.428	0.450	0.492	0.526	0.516	0.538	0.565	0.502	9.74	
83)	Chrysene	1.179	1.173	1.261	1.292	1.277	1.318	1.346	1.264	5.21	
84)	Bis(2-ethylhex...	0.758	0.747	0.790	0.783	0.740	0.727	0.737	0.755	3.16	
85) c	Di-n-octyl pht...	1.291	1.287	1.366	1.362	1.307	1.311	1.316	1.320	2.42	

Method Path : Z:\svoasrv\HPCHEM1\BNA_M\Methods\
Method File : 8270-BM040825.M

86) I	Perylene-d12	-----ISTD-----								
87)	Indeno(1,2,3-c...	1.375	1.368	1.480	1.507	1.523	1.577	1.606	1.491	6.16
88)	Benzo(b)fluora...	1.141	1.157	1.223	1.308	1.313	1.379	1.420	1.277	8.40
89)	Benzo(k)fluora...	1.132	1.122	1.209	1.242	1.263	1.336	1.362	1.238	7.46
90) C	Benzo(a)pyrene	1.011	1.026	1.096	1.130	1.154	1.209	1.230	1.122	7.50
91)	Dibenzo(a,h)an...	1.128	1.128	1.205	1.244	1.258	1.301	1.331	1.228	6.43
92)	Benzo(g,h,i)pe...	1.178	1.179	1.255	1.265	1.269	1.310	1.328	1.255	4.65

(#) = Out of Range

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: POWE02

Lab Code: CHEM Case No.: Q1858 SAS No.: Q1858 SDG No.: Q1858

Instrument ID: BNA_M Calibration Date/Time: 04/23/2025 10:38

Lab File ID: BM050007.D Init. Calib. Date(s): 04/08/2025 04/08/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 13:35 20:07

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.178	1.114		-5.4	
Phenol-d6	1.465	1.377		-6.0	
Nitrobenzene-d5	0.359	0.354		-1.4	
Naphthalene	1.028	0.987		-4.0	
2-Fluorobiphenyl	1.475	1.485		0.7	
Fluorene	1.439	1.386		-3.7	
2,4,6-Tribromophenol	0.291	0.283		-2.7	
Phenanthrene	1.096	1.052		-4.0	
Anthracene	1.081	1.051		-2.8	
Pyrene	1.378	1.424		3.3	
Terphenyl-d14	1.067	1.201		12.5	
Benzo (a) anthracene	1.329	1.296		-2.5	
Chrysene	1.264	1.209		-4.4	
Benzo (b) fluoranthene	1.277	1.250		-2.1	
Benzo (a) pyrene	1.122	1.069		-4.7	20.0
Indeno (1,2,3-cd) pyrene	1.491	1.378		-7.6	
Benzo (g,h,i) perylene	1.255	1.140		-9.2	

All other compounds must meet a minimum RRF of 0.010.

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: POWE02

Lab Code: CHEM Case No.: Q1858 SAS No.: Q1858 SDG No.: Q1858

Instrument ID: BNA_P Calibration Date/Time: 04/24/2025 11:50

Lab File ID: BP024407.D Init. Calib. Date(s): 04/14/2025 04/14/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 11:06 17:13

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.210	1.166		-3.6	
Phenol-d6	1.656	1.634		-1.3	
Nitrobenzene-d5	0.351	0.345		-1.7	
Naphthalene	1.054	1.023		-2.9	
2-Fluorobiphenyl	1.306	1.284		-1.7	
Fluorene	1.388	1.375		-0.9	
2,4,6-Tribromophenol	0.277	0.310		11.9	
Phenanthrene	1.091	1.052		-3.6	
Anthracene	1.052	1.054		0.2	
Pyrene	1.271	1.226		-3.5	
Terphenyl-d14	0.992	0.979		-1.3	
Benzo (a) anthracene	1.243	1.201		-3.4	
Chrysene	1.191	1.138		-4.4	
Benzo (b) fluoranthene	1.199	1.201		0.2	
Benzo (a) pyrene	1.034	1.017		-1.6	20.0
Indeno (1,2,3-cd) pyrene	1.388	1.197		-13.8	
Benzo (g,h,i) perylene	1.173	0.988		-15.8	

All other compounds must meet a minimum RRF of 0.010.